



ILMATIETEEN LAITOS  
METEOROLOGISKA INSTITUTET  
FINNISH METEOROLOGICAL INSTITUTE

RAPORTTEJA  
RAPPORTER  
REPORTS  
2017:1

# DEMONSTRATION OF THE EQUIVALENCE OF PM<sub>2.5</sub> AND PM<sub>10</sub> MEASUREMENT METHODS IN KUOPIO 2014-2015

JARI WALDÉN  
TUOMAS WALDÉN  
SISKO LAURILA  
HANNELE HAKOLA



ISBN 978-952-336-010-5 (pdf)  
ISSN 0782-6079

Helsinki 2017



Published by Finnish Meteorological Institute  
(Erik Palménin aukio 1), P.O. Box 503  
FIN-00101 Helsinki, Finland

Date 2017

Authors	Name of project
Jari Waldén, Tuomas Waldén, Sisko Laurila and Hannele Hakola	PM <sub>2.5</sub> and PM <sub>10</sub> tests for the demonstration of equivalence
	Commissioned by Ministry of the Environment

## Title

**Demonstration of the equivalence of PM<sub>2.5</sub> and PM<sub>10</sub> measurement methods in Kuopio 2014-2015**

## Abstract

The Air Quality Directive, AQD, (2008/50/EC), set up the rules concerning the reference methods (RM) for the measurements of e.g. mass concentration of particulate matter in air. A member state can use any other method which it can demonstrate to display a consistent relationship with the reference method. In case of particulate matter the reference method is the gravimetric measurement method for the determination of the PM<sub>2.5</sub> and PM<sub>10</sub>, mass concentration of suspended particulate matter. Demonstration whether a non-reference method is equivalent with the reference method for measurements of the concentration of particulate matter (PM) for the size classes of PM<sub>2.5</sub> and PM<sub>10</sub>, i.e., particle sizes of less than 2.5 µm and 10 µm in aerodynamic diameter, respectively, was conducted in the city of Kuopio, Finland.

Altogether eight candidate methods (CM) i.e. measurement method for which the equivalence need to be demonstrated, participated in the equivalence studies: FH 62 I-R, Grimm model 180, MP101 CPM, Osiris, and SHARP model 5030, TEOM 1405, BAM 1020 and DustTrak 8535. The test program was constructed in accordance with the Guide for Demonstration of Equivalence of Ambient Air Monitoring Methods (GDE) by the European Commission. However, suitability test, which presumes that more than 20 % of the daily concentration values during the whole comparison study shall exceeds the upper assessment threshold (UAT) for annual limit value defined by the AQD, was not fulfilled for PM<sub>2.5</sub> measurements. This is a very typical situation in Finland. In case of PM<sub>10</sub> comparison, the daily mass concentrations exceeded the UAT for more than 20 % of the measurement period to fulfill the requirement by the GDE. The calibration function against the RM has been presented for each CM as well as the results if non-compliance with criteria on any of the CMs has occurred. The results showed that DustTrak 8535 failed to meet the criteria for equivalence for fixed measurements of PM<sub>2.5</sub> and PM<sub>10</sub>. Osiris fulfilled the criteria for PM<sub>10</sub> measurements, but failed to meet the criteria for PM<sub>2.5</sub> measurements. All other CMs, FH 62 I-R, Grimm model 180, MP101 CPM, SHARP model 5030, TEOM 1405 and BAM 1020 fulfilled the criteria for PM<sub>2.5</sub> and for PM<sub>10</sub> measurements.

---

Publishing unit  
Air Quality

---

Classification (UDK)  
502.3:613.15, 681.2

Keywords  
Air Quality, Instrumentation

---

ISSN and series title  
0782-6079 Raportteja-Raporter-Reports

---

ISBN  
978-952-336-010-5 (pdf)

Language  
English

---

Sold by  
Finnish Meteorological Institute  
P.O. Box 503, FIN-00101 Helsinki  
Finland

Pages 134

Price

Notes



## ILMATIETEEN LAITOS

Julkaisun sarja, numero ja raporttikoodi  
Raportteja 2017:1

Julkaisija	Ilmatieteen laitos, (Erik Palménin aukio 1) PL 503, 00101 Helsinki	Julkaisuaika 2017
Tekijä(t)	Projektin nimi	
Jari Waldén, Tuomas Waldén, Sisko Laurila ja Hannele Hakola	PM <sub>2.5</sub> - ja PM <sub>10</sub> -vertailumittaukset ekvivalenttiuden osoittamiseksi	
Toimeksiantaja Ympäristöministeriö		

Nimike

### **PM<sub>2.5</sub>- ja PM<sub>10</sub>-mittausmenetelmien yhdenmukaisuuden osoittaminen vertailumenetelmää vastaan Kuopiossa 2014–2015**

Tiivistelmä

Euroopan Yhteisön CAFÉ-direktiivin (2008/50/EY) määrittää mm. hiukkasmittausten vertailumenetelmän, mikä perustuu hiukkasmassan gravimetriseen määritykseen. Jäsenmaa voi käyttää muuta menetelmää vertailumenetelmän sijasta, mikäli se voi osoittaa, että tulokset ovat yhteneviä vertailumenetelmän antamien tulosten kanssa. Tällaisella menetelmällä saatuja tuloksia on tarvittaessa korjattava, jotta saataisiin vertailumenetelmää käytäväällä saatavia tuloksia vastaavat tulokset. Hiukkasmittausten osalta vertailumenetelmä PM<sub>2.5</sub> ja PM<sub>10</sub> massapitoisuuden määrittämiseksi ulkoilmassa perustuu gravimetriseen määritysmenetelmään. Tässä tutkimuksessa testattiin eräiden jatkuvatoimisten hiukkasmittalaitteiden yhdenvertaisuutta vertailumenetelmää vastaan noudattaen komission ohjetta. Vertailumittaukset tehtiin Kuopiossa ja ne käsittivät hiukkasten aerodynaamiselta halkaisijaltaan olevat alle 2.5 µm:n ja alle 10 µm:n kokoiset hiukkasmittaukset.

Vertailuun osallistuvat kaikkiaan kahdeksan eri jatkuvatoimista hiukkasanalysaattorimallia: BAM 1020, DustTrak 8535, FH 62 I-R, Grimm 180, MP101 CPM, Osiris, SHARP 5030 ja TEOM 1405. Vertailumittausohjelma toteutettiin komission ohjeen mukaisesti. Ohje määrittelee mm. pitoisuustasoista niin, että 20 % havainnoista tulee ylittää alimman arviontikynnykset (UAT) sekä PM<sub>2.5</sub> että PM<sub>10</sub> massapitoisuksille. Havaintoaineistosta osoitautui, että PM<sub>2.5</sub> massapitoisuudet eivät ylittäneet sille asetettua UAT arvoa, 17 µg/m<sup>3</sup>, yli asetetun kriteerin, vaan tästä jäätii selvästi alle. Verrattaessa Suomessa mitattuja PM<sub>2.5</sub> pitoisuksia, voidaan todeta, että hyvin harvoin mittauksissa ylittiin UAT-arvo PM<sub>2.5</sub> pitoisuksissa. PM<sub>10</sub> pitoisuksissa sen sijaan ylempi arviontikynnys, 30 µg/m<sup>3</sup>, ylittiin useammin kuin vaadittava määrä. Vertailumittaukset analysoitiin komission ohjeen mukaan ja sen perusteella esitetään jokaiselle vertailtavalle hiukkasanalysaattorille kalibrointikertoimet sekä PM<sub>2.5</sub>- että PM<sub>10</sub>-mittauksille. Raportissa osoitetaan myös kaikki poikkeamat, mitä esiintyi sallituista kriteereistä. Tulosten perusteella voidaan todeta, että yksi vertailtava analysaattori, DustTrak 8535, ei täytänyt vaadittavia kriteerejä sekä PM<sub>2.5</sub>- että PM<sub>10</sub>-mittauksille. Osiris täytti vaatimukset PM<sub>10</sub>-mittauksille, mutta ei PM<sub>2.5</sub>-mittauksille. Sen sijaan kaikki muut testatut hiukkasanalysaattorit, FH 62 I-R, Grimm 180, MP101 CPM, SHARP 5030, TEOM 1405 ja BAM 1020 täyttivät vertailumenetelmälle asetetut vaatimukset sekä PM<sub>2.5</sub>- että PM<sub>10</sub>-mittauksille.

---

Julkaisijayksikkö

Ilmanlatu

Luokitus (UDC)	Asiasanat
502.3:613.15, 681.2	Ilmanlaatu, mittalaitteet

ISSN ja avainnimike

1796–1203 Raportteja-Raporter-Reports

ISBN	Kieli	
	Englanti	
978-952-336-010-5 (pdf)		
Myynti	Sivumäärä 134	Hinta
Ilmatieteen laitos		
PL 503, 00101 Helsinki	Lisätietoja	

## Contents

1. Introduction .....	7
2. PM equivalence procedure .....	8
2.1 Requirements of Directive 2008/50/EC and 2015/1480/EU .....	8
2.2 The reference method .....	11
2.3 Candidate methods.....	12
2.4. Equivalence procedure.....	25
3. Experimental set up.....	32
3.1 Laboratory tests.....	32
3.2 Field campaigns and the measurement site.....	33
3.3 Filter weighing procedure .....	38
3.4 Data acquisition system at the measurement site.....	40
3.5 Operation of the instruments.....	41
3.5.1. Reference sampler.....	41
3.5.2 CM operating conditions.....	41
3.5.3 QA/QC procedures.....	41
3.5.4 Field audit .....	43
4. Data analysis .....	44
4.1 Description of the database .....	44
4.2 Data processing.....	45
5. Results .....	46
5.1 Meteorological conditions during the equivalence comparisons .....	46
5.2 Results of the PM <sub>10</sub> comparison .....	50
5.1.3. Results of the candidate methods for PM <sub>10</sub> .....	52
5.1.4 Summary of the PM <sub>10</sub> comparisons .....	62
5.3. Results of the PM <sub>2.5</sub> comparisons .....	64
5.3.3. Results of the candidate methods for PM <sub>2.5</sub> .....	66
5.3.4 Summary of the PM <sub>2.5</sub> comparisons .....	75
5.3. Result of the field audit.....	77
5.3. Calibration equations for CMs .....	77
5.4 Results of the automated PM analyzer at the Tasavallankatu.....	79
6. Conclusions .....	81
Acknowledgements .....	83
7. References .....	84
Annexes.....	87
A1: QA/QC procedures conducted for the RM and CM during the comparison experiments.....	87
A2 PM <sub>10</sub> equivalence tests for BAM .....	90
A3. PM <sub>10</sub> equivalence tests for DustTrak analyzers .....	92
A4. PM <sub>10</sub> equivalence tests for FH 62 I-R analyzers.....	94
A5. PM <sub>10</sub> equivalence tests for Grimm analyzers.....	96
A6. PM <sub>10</sub> equivalence tests for Environnement MP101 analyzers.....	98
A7. PM <sub>10</sub> equivalence tests for Osiris analyzers.....	100
A8. PM <sub>10</sub> equivalence tests for SHARP beta signal .....	102
A9. PM <sub>10</sub> equivalence tests for SHARP c-dust analyzers.....	104

A10. PM <sub>10</sub> equivalence tests for TEOM analyzers .....	106
A11: Calibration factors for CM for each PM <sub>10</sub> campaign .....	108
A12. PM <sub>2.5</sub> equivalence tests for BAM analyzers .....	109
A13. PM <sub>2.5</sub> equivalence tests for DustTrak analyzers .....	111
A14. PM <sub>2.5</sub> equivalence tests for FH 62 I-R analyzers .....	113
A15. PM <sub>2.5</sub> equivalence tests for Grimm analyzers .....	115
A16. PM <sub>2.5</sub> equivalence tests for Environnement MP101 analyzers .....	117
A17. PM <sub>2.5</sub> equivalence tests for Osiris analyzers .....	119
A18. PM <sub>2.5</sub> equivalence tests for SHARP beta analyzers .....	121
A19. PM <sub>2.5</sub> equivalence tests for SHARP c-dust analyzers .....	123
A20. PM <sub>2.5</sub> equivalence tests for TEOM analyzers .....	125
A21: Calibration factors for CM for each PM <sub>2.5</sub> campaign .....	127
A22 Daily values of PM 10: Reference method, BAM, FH 62IR, Osiris .....	128
A23 Daily values of PM 10: Reference method, TEOM, SHARP, MP 101 .....	129
A24 Daily values of PM 10: Reference method, Grimm 180.....	130
A25 Daily values of PM <sub>2.5</sub> : Reference method, TEOM, SHARP c and MP 101 .....	131
A26 Daily values of PM <sub>2.5</sub> : Reference method, Grimm 180 .....	132
A27 Daily values of PM <sub>2.5</sub> : Reference method, BAM 1020 and FH-62IR .....	133
A28 Findings from the audit surveillance study .....	134

## **List of abbreviations**

AQD	Air Quality Directive
CM	Candidate Method
DQO	Data Quality Objectives
EC	European Commission
EU	European Union
EMEP	Programme for Monitoring and Evaluation of long range transmission of air Pollutants in Europe
FMI	Finnish Meteorological Institute
GDE	Guidance for Demonstration of Equivalence
HSY	Helsinki Region Environmental Services Authority
LV	Limit Value
MS	Member States
PM	Particulate Matter
PTFE	Polytetrafluoroethylene
RM	Reference Method
TSP	Total Suspended Particles
UAT	Upper Assessment Threshold
WG	Working Group

## 1. Introduction

The European Commission (EC) has the air quality on high priority in European policy. Mostly the pollutants have health effects on human beings and can damage vegetation. To prevent the harmful effects of pollutants in humans and vegetation the European Commission has prepared directives to define the limit values for the maximum concentrations of certain pollutants in ambient air. The new Commission Directive 2015/1480/EC (CD) amend some of the annexes in the Air Quality Directive 2008/50/EC (AQD), but also defined the responsible authorities and their tasks, the limit values and data quality objectives (DQO) for specific pollutants as well as updated the reference methods for the measurements. The national authorities also appoint a National Reference Laboratory for specific tasks amongst, i.e. the coordination of the appropriate use of the reference methods and the demonstration of equivalence of non-reference methods. The member states (MS) can, however, use another measurement method if it can be shown to give results equivalent with the reference method. To harmonize the process of demonstrating the equivalency of a candidate method (CM) with the reference method, the EC working group on Guidance for the Demonstration of Equivalence prepared a guide. As a result of the working group, the test criteria set out in the EC Guidance for the Demonstration of Equivalence of Ambient Air Monitoring Methods (GDE) was published in 2005 and reviewed in 2009. Throughout the report, the latest version of the GDE i.e. from the year 2010 is used. The meaning of the term ‘equivalent method’ was defined in the GDE as *‘An equivalent method to the reference method for the measurement of a specified air pollutant, is a method meeting the data quality objectives for fixed measurements specified in the relevant air quality directive’*. This definition is used also in this report.

Member states have conducted the equivalence tests especially with the analyzers of automated particulate matter (Harrison et al. 2006, Beijk et al. 2007, de Jonge 2008, Bertrand, et al. 2009) over the past years. Additionally testing laboratories, like TÜV Rheinland in Germany, have completed a great number of tests for demonstrating

equivalency of the automated PM analyzers ([www.qal1.de](http://www.qal1.de)). Similarly, the UK testing scheme for air quality measurements (MCERT) have conducted and published equivalency tests for the PM analyzers ([www.sira.uk](http://www.sira.uk)).

This report provides test results of the comparisons of the automated candidate methods against the reference methods for PM<sub>2.5</sub> and for PM<sub>10</sub> measurements. This is the second trial for demonstration of equivalency for automated PM analyzers in Finland. The first was conducted in Helsinki in 2007 - 2008 (Walden et al., 2010). The comparisons as well as the analysis of the comparison results were conducted according to the GDE. The candidate methods tested in this comparison were those used or planned to be used in the local air quality networks in Finland.

## 2. PM equivalence procedure

### 2.1 Requirements of Directive 2008/50/EC and 2015/1480/EU

The Air Quality Directive 2008/50/EC (AQD) states that member states shall apply the reference measurement methods for measurements of the atmospheric pollutants for which limit or target values are defined in the Directive. Other measurement methods may be used, subject to the conditions also set out in the Directive. The reference measurement method for the sampling and measurement of mass concentration of PM<sub>10</sub> as well as PM<sub>2.5</sub> is gravimetric and is prescribed in European standard (European Norm, EN) EN 12341 (EN 12341, 2014) prepared by the European Committee for Standardization (CEN). The Directive defines that PM<sub>10</sub> and PM<sub>2.5</sub> means particulate matter that passes through a size-selective inlet with 50 % efficiency at cut-off size of 10 µm and 2.5 µm as aerodynamic diameter, respectively. In Figure 2.1 the sampling efficiency for the PM<sub>10</sub> sampling inlet is shown (Kaminski and Kuhlbusch, 2010). From the figure one can see that the sampling efficiency for 10 µm particles is 50 % reaching zero at a size of 20 µm. The standard prescribes the measurement method including the inlet design criteria for both of the size

classes as well as the storage and weighing procedures for the filters. The approved filter materials for the collection of the PM fraction are also defined. In addition to the reference methods, the AQD lays down that the mass concentration for PM results shall be referred to the prevailing ambient conditions in terms of temperature and atmospheric pressure. The DQO defined for the PM measurements in the AQD are listed in Table 2.1.




Figure 2.1: The sampling efficiency of the  $PM_{10}$  sampling inlet.

*Table 2.1. The Data Quality Objective for the PM<sub>2.5</sub> and PM<sub>10</sub> measurements for fixed and indicative measurements according to the AQD.*

		PM <sub>2.5</sub> and PM <sub>10</sub>
Fixed measurements		
uncertainty <sup>(1)</sup>		25 %
minimum data capture		90 %
Indicative measurements		
uncertainty <sup>(1)</sup>		50 %
minimum data capture		90 %
minimum time coverage		14 % <sup>(2)</sup>

<sup>(1)</sup> Describes the relative expanded measurement uncertainty at the 95 % confidence level.

<sup>(2)</sup> One measurement a week at random, evenly distributed over the year, or eight weeks evenly distributed over the year.

Beside the reference method for measurements of mass concentration of PM<sub>10</sub> and PM<sub>2.5</sub>, automated measurement methods have been manufactured and used for the PM measurements. The automated methods can provide data at much shorter intervals, e.g. from tens of seconds up to hourly values, than the reference method and can be used to trace sudden changes in the PM concentrations in the air and in the calculation of air quality indexes.

In 2002, EC launched a working group on Guidance for the Demonstration of Equivalence. After couple of years of work, the WG completed its work in 2005 with a document *Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods, GDE*. To facilitate the use of the GDE for the demonstration of equivalence of the candidate methods (CM) against the reference method (RM) for PM monitoring, an Excel macro was made available on the Commission web page (<http://ec.europa.eu/environment/air/quality/legislation/assessment.htm>). The macro (R. Beijk, 2011) allows the user to test of the equivalency for input pairs of data values of the CM and the RM. The GDE document was revised in 2008 and was implemented into the Directive 2008/50/EC. Some editorial corrections were made to the version 2010, which is the present version and is referred to here.

## 2.2 The reference method

The reference sampler used during the PM equivalence tests in Kuopio ( $PM_{10}$  and  $PM_{2.5}$ ) was a sequential type sampler SEQ47/50 by Sven Leckel, Ingenieurbüro GmbH, Germany. Two identical units of the samplers were used both for the  $PM_{10}$  and the  $PM_{2.5}$  tests. The layout of the reference sampler is shown in Figure 2.2. The size classification inlet followed the designing criteria of the EN 12341 and is shown in Figure 2.3. By switching the jets inside, the inlet can be used for  $PM_{10}$  and  $PM_{2.5}$  measurements. The sampler is equipped with a heater to keep the temperature above the dew point to prevent sample filters from freezing at winter conditions. In addition, cooled sample tube by sheath flow and enclosed sample filters within the filter cartridge prevent the volatilization of semi-volatilized compounds such as sulphate, ammonia and nitrate.

The filter types used in the tests were polytetrafluoroethylene (PTFE) filters, Millipore Fluoropore FSLW047 # 3  $\mu m$ , by Millipore.



*Figure 2.2: The layout of the reference method, the flow controlling units by MCZ and the PM inlets by Digitel, used for the  $PM_{10}$  and the  $PM_{2.5}$  equivalence tests.*



*Figure 2.3: The sampling inlet of the sequential sampler SEQ47/50.*

### **2.3 Candidate methods**

The following candidate instruments, manufacturer and the acronym of the instrument used in this report (in parenthesis) took part in the PM equivalence tests:

1. FH 62 I-R by Thermo Fisher Scientific, USA (FH 62 I-R);
2. Grimm Environmental Dust Monitor, model 180, by GRIMM Aerosol Technik GmbH & Co. Member of Durag Group, Germany (Grimm 180);
3. MP101 CPM, by Environnement SA, France (MP101\_CPM)
4. Osiris, by Turnkey Instruments Ltd, England (Osiris);

5. Synchronized Hybrid Ambient Real-time Particulate Monitor, model 5030 by Thermo Fisher Scientific, USA (SHARP);
6. Tapered element oscillating microbalance, TEOM 1405 by Thermo Fisher Scientific, USA (TEOM 1405);
7. BAM-1020 Continuous Particle Monitor, Met One Instruments, Inc. (BAM);
8. DustTrak Aerosol monitor, model 8535, TSI Incorporated, USA (DustTrak).

Short descriptions of the CMs are given below. The set-up and the sample equipment of each of the CMs are described. The type of sampling inlets, sampling flow rates, sampling period and the condition of the sample tube (heated/not heated) are reported. As a general rule, correction coefficients or any other factors installed in the operational software, were checked and recorded. Also the version of software were checked and recorded. For parallel CMs the correction coefficients and factors were set to be equal. This means that the CMs were tested against the RM with the same basic measurement signal, sampling equipment and sampling tube conditions during the test campaigns. If deviations in the set-up of the CMs from the equivalence tests occur in routine use, the test results may not be valid. Examples of such deviations are deviations from the sampling tube temperature, a different sample flow rate, or a different sample inlet. In these circumstances, more evidence is needed to demonstrate the validity of the test results. If the deviation of the CM from the test condition is associated with the measurement signal and can be transformed to the test conditions mathematically, the test results obtained in the equivalence tests are then valid. The manufacturer is responsible for informing on any changes made in the instruments, which may cause a significant change in the performance of the instrument. If the manufacturer apply the ISO EN 15267 (EN 15267 1-2, 2009) in the production the information on the changes made to instrument is mandatory. In other case it is voluntary. A change of materials can also have influence on the performance of the instrument, and the customers should be informed on this. An example of this is the change of filter material with material, which is more resistant for the humidity effect.

*FH 62 I-R*

The ESM FH 62 I-R monitor by Thermo Fisher, USA, shown in Figure 2.4, uses the technique of  $\beta$ -attenuation (Kr-85 source). The attenuation of  $\beta$ -rays by a filter is directly related to the amount of mass on the filter. The air sample is collected on the pure spot of the filter tape and remains at the measurement/sample point until it is full loaded or after 24 hour sampling after which the filter tape rotates to bring a new pure spot on the measurement/sample point. The analysis of the sample, however, takes place cumulatively over the 24 h. To avoid condensation of water on the filter, the sampling tube is heated (35 °C). This process not only leads to the loss of water, but also to the loss of certain semi-volatile compounds such as ammonium nitrate. By changing the sample inlet, the device is capable of making measurements of PM<sub>10</sub> and PM<sub>2.5</sub> at a sample flow of 1 m<sup>3</sup>/h. The sample inlet was one of the commercial types designed according to the EN-standards for PM<sub>2.5</sub> and PM<sub>10</sub>. The measurement range for normal operation is from 0 µg/m<sup>3</sup> to 5000 µg/m<sup>3</sup>. The manufacturer provides a calibration kit, i.e., a zero plate and a plate of known amount of mass concentration on a film foil to calibrate the instrument. The calibration kit was used for calibration of the instrument during the equivalence test. The manufacturer of the instrument installed a correction factor with a default value of 1.3 in the operational software of the instrument to correct the measurement signal according to the guideline by the EC (EC WG on PM, 2001). In these tests, the correction factor was set to 1.0 according to the policy mentioned in the previous chapter.



Figure 2.4. FH 62 I-R monitor.

*Grimm Environmental Dust Monitor, model 180*

The Grimm ambient dust monitor 180 is a stationary continuous fine dust measuring system for the simultaneous and continuous measurement of PM<sub>10</sub>, PM<sub>2.5</sub> and PM<sub>1</sub>. The Grimm 180, shown in Figure 2.5, does not have PM<sub>2.5</sub> or PM<sub>10</sub> sampling heads according to EN-standards. The sample inlet of the Grimm is the manufacturer's own design, but it has been tested against the PM<sub>10</sub> reference method according to EN 12341 (LUBW, 2005). The sample flow rate of the Grimm was 1.2 l/min as stated by the manual and the sampling tube was inside the shield tube at ambient temperature. The concentration range for dust particles is from 0.1 to 1500 µg/m<sup>3</sup>. The instrument uses an optical technique, based on light scattering, to divide particles into different sizes in diameter. The value of the refraction index of the particles, i.e., how much the velocity of light is reduced due to the reflection from the surface of the particles, has been programmed into the software. Specific algorithms are used to transfer the number of particles of certain size into mass.

The calculated cut-off point curves are then applied to define the mass concentration for PM<sub>1</sub>, PM<sub>2.5</sub> and PM<sub>10</sub>. The sample air passes through an isothermal air drying system-during which moisture is extracted via a Nafion tube. This reduces the possibility of nucleonic condensation and therefore artificial growth/weight. The pump of the Nafion dryer starts at relative humidity of 50 % reducing the relative humidity down to 35 %.



*Figure 2.5. Grimm Environmental Dust Monitor, model 180.*

#### *MP101 CPM*

The MP101 CPM, shown in Figure 2.6, measures particulate concentration based on the technique of  $\beta$ -attenuation by measuring the amount of radiation a sample absorbs when exposed to low-energy  $\beta$ -rays (C-14 source). For continuous and simultaneous measurement of fine dust the CM was equipped with a CPM module using the optical measurement technology. By changing the sample inlet, the device is capable of measurements of PM<sub>10</sub> and PM<sub>2.5</sub> at a sample flow of 1 m<sup>3</sup>/h. The sample inlets for PM<sub>2.5</sub> and PM<sub>10</sub> were designed according to the EN-standards by the manufacturer. The sampling tube is equipped with shielded flow to avoid condensation in the sample air when entering the analyzer. The measurement ranges of the instrument are selectable up to 10 000  $\mu\text{g}/\text{m}^3$ .

The sampling time was set to 1 h in the operational software of the instrument by the manufacturer.



*Figure 2.6: Environnement MP101 CPM monitor.*

### *Osiris*

The Osiris is one of Turnkey's families of direct-reading airborne particle monitors, which can be used as a portable instrument or deployed in a semi-permanent installation. The Osiris, shown in Figure 2.7, indicates continuously the concentration of total suspended particles (TSP), PM<sub>10</sub>, PM<sub>2.5</sub> and PM<sub>1</sub> at a range of up to 6000 µg/m<sup>3</sup> based on an optical method. The sample flow rate of the Osiris is 0.6 l/min. The sample tube was heated (35 °C) to avoid condensation. The sample inlet was designed by the manufacturer; no test report for the sample inlet against the reference method was available.



*Figure 2.7. Osiris optical dust monitor*

*SHARP, model 5030*

The SHARP monitor by Thermo Fisher combines light-scattering photometry and beta radiation attenuation (C-14 source) in one instrument, shown in Figure 2.8. The instrument combines nephelometry and the beta attenuation method to provide a continuous reading of the PM concentration. The sampling strategy is the same as in FH 62 I-R. The output signals can be selected to provide the dust signal,  $\beta$ -signal and optical signal. The C-dust signal is the optical signal corrected with the  $\beta$ -signal providing the fast response signal. The  $\beta$ -signal represents beta radiation attenuation method with a time interval of an hour. The optical signal is a direct signal from the nephelometer but is uncalibrated. For normal use, the C-dust signal is recommended. Control of relative humidity and frequent filter changes eliminate water vapor without loss of volatile organic compounds. By changing the sample inlet, the device is capable of measurements of  $PM_{10}$  and  $PM_{2.5}$  at a sample flow of  $1\text{ m}^3/\text{h}$ . The commercial sample inlet was used during the tests. The concentration ranges can be from 0 to  $1000\text{ }\mu\text{g}/\text{m}^3$  or from 0 to  $10\,000\text{ }\mu\text{g}/\text{m}^3$ . For calibration of the instrument

the same calibration kit as in the FH 62 I-R was used. To avoid condensation, the sample tube was heated (35 °C).



*Figure 2.8. Synchronized Hybrid Ambient Real-time Particulate Monitor, SHARP, model 5030*

#### *TEOM 1405*

The TEOM 1405, shown in Figure 2.9, uses the tapered element oscillating microbalance technique to measure the concentration of the particulate matter in the air. It is a direct mass measurement technique on a filter with real-time data output. The sample filter need to be changed according to the loading percentile of the filter as indicated by the instrument or at regular intervals. By changing the sample inlet, the device is capable of making measurements of PM<sub>10</sub> and PM<sub>2.5</sub> at a sample flow of 1 m<sup>3</sup>/h. The sample inlet provided by the manufacturer was used during the equivalence measurements. In the case of the PM<sub>2.5</sub> test, the sharp cut cyclone (cut to size for PM<sub>2.5</sub>) was installed in the sampling tube to remove particles larger than PM<sub>2.5</sub>. The measurement concentration range of the particles for the TEOM 1405 can be up to 5 g/m<sup>3</sup>. To avoid condensation, the sample tube was heated (50 °C). The correction equation used in the software of the device by the manufacturer was of the form:  $y = a + b \cdot C$ , where  $a = 3 \mu\text{g}/\text{m}^3$ ,  $b = 1.03$  and  $C$  is the measurement signal.



Figure 2.9: Thermo Scientific Ambient Particulate Monitor, TEOM 1405

### BAM 1020

The BAM 1020 particulate monitor shown in Figure 2.10 uses the beta attenuation method. The mass of the sample collected on the filter tape is measured by detecting the attenuation of  $\beta$ -rays from a radioactive source of C-14. The measurement cycle consist of the measurement of the  $\beta$ -count rate through a clean filter spot, sampling interval and the measurement of the  $\beta$ -count rate through sampled filter spot. The analysis of the  $\beta$ -count rate takes 8 min and the removal of the filter tape between the analysis and sampling points takes 1 – 2 min allowing a sampling time of 42 min for the whole measurement cycle of

60 min. However, with the tested CMs the Estonian Environmental Research Centre (EERC) selected the measurement cycle as 180 min, which is the practice in Estonia. The main reason for the practice by the EERC was to decrease the running costs of the instrument by extending the duration of the filter tape from the normal 60 days up to 180 days. This practice was agreed with the FMI and EERC at the beginning of the comparison. The sampling system consists of sampling tube, sampling inlet for PM<sub>10</sub> and a sharp cut cyclone for PM<sub>2.5</sub> sampling. A smart inlet heater was installed in the sampling tube to control the maximum relative humidity at the filter tape (45 % as factory setting). There are some additional instructions by the manufacturer to clean the nozzle and vane of fibers released from the filter tape.



Figure 2.10: BAM 1020 Ambient Beta Gauge Particulate Monitor

*DustTrak 8535*

DustTrak 8535 follows the principles of light scattering from dust particles by laser light. The instruments were sent to comparison tests without a sample heater and heat shield assembly. The instrument is presented in Figure 2.11. The data was stored on the mass memory of the instrument, collected every two weeks through a USB memory stick and stored on the database with the other CMs at the FMI.



Figure 2.11: *DustTrak 8535*.

In Table 2.2 the participating CMs are presented including the manufacturer or their representative organization, the measurement methods and type of sampling inlets and the

temperature of the sampling tube for both the PM<sub>2.5</sub> and for PM<sub>10</sub> comparisons. The software version used for each of the CMs is presented in Annex A1.

*Table 2.2: List of CMs, representative organizations/manufacturers, measurement methods, type of sampling inlets, and sampling tube temperature for the PM<sub>2.5</sub> and PM<sub>10</sub> comparisons. The last column indicates the field campaigns for which the PM<sub>2.5</sub> and PM<sub>10</sub> comparison took place (T = Tasavallankatu and S = Savilahdentie).*

Instruments	Manufacturer / Representative / Instrument provider	Method	PM10/PM2.5 inlet	Sample tube	Field campaign (T/S)
REFERENCE SAMPLER	Leckel SEQ 47/50	Sequential sampler	Leckel	Ambient	T/S
Environnement MP-101+CPM	ENVIRONNEMENT SA (France)	$\beta$ -attenuation + optical	Environment EN	Shield tube, ambient temp	T/S
GRIMM-180	Finnish Meteorological Institute	Optical (light scattering)	Grimm	Shield tube, nafion drier	T/S
SHARP 5030	Finnish Meteorological Institute	Light scattering + $\beta$ -attenuation	Digitel /EN	Heated 35 °C	T/S
FH-62-IR	Finnish Meteorological Institute	$\beta$ -attenuation	Digitel /EN	Heated 35 °C	T/S
BAM 1020	Estonian Environmental Research Centre (Estonia)	$\beta$ -attenuation	US-EPA	Heated 40 °C	T/S
TEOM-1405	JPP-kalibrointi ky, FMI	Tapered Element Oscillating Microbalance	US-EPA	Heated 50 °C	T/S
OSIRIS	Hnu-Nordion (Finland) / Turnkey Instruments	Optical (nephelometer)	Osiris	Heated 35 °C	T/S
DUST TRAK DRX 8533	Oy Teknocalor Ab (Finland)/ TSI Inc (USA)	Optical (light scattering)	TSI	Ambient	S

The companies of Environnement SA (France), HNU-Nordion/Turnkey Instruments (USA) and Oy Teknocalor Ab/ TSI Inc (USA) each provided two CMs for the tests. The Finnish Meteorological Institute (FMI) provided two SHARP, two FH 62 I-R and one TEOM 1405. One TEOM 1405 was provided by JPP kalibrointi Ky. The Estonian Environmental Research Centre provided two BAMs. Maintenance of the Thermo Fisher instruments (FH 62 I-R, SHARP and TEOM) as well as the sampling inlets was provided by Oy Kontram ab and Sintrol Oy provided spare parts for the BAMs.

The size selective inlets used for RM and CMs are shown in Figures 2.12 and 2.13. In case of Grimm, Osiris and DustTrak the sampling inlets were not size selective but the size classification is obtained by light scattering.



*Figure 2.12. The sampling inlet for the reference method used in Leckel: Complete (in the left), separated (in the middle) and the impactor plate (in the right).*



*Figure 2.13. Size selective inlets used for CM: EU-Digitel (FH 62 I-R, SHARP), US-EPA (BAM) and EU-Environnement (MP101).*

There are differences on the sampling efficiency of the size selective inlets with cut size of  $10 \mu\text{m}$ , shown in Figure 2.14 (Kaminski and Kuhlbusch, 2010). The blue and pink curve represents the EU inlet with different distance of impactor plate from the jets. The standard allows a distance variation of  $8 \pm 2 \text{ mm}$ . The black curve presents an inlet where the

distance is larger than allowed. The other two curves, the US-EPA inlet (red spots) and according to German standard DIN EN 481 (green curve) which is designed for working environment condition agree with each other very well, but are not so steep in cut off size than the others. The cut off size of 50 % efficiency is between 9.73 to 10.28  $\mu\text{m}$  for the inlets according to EU-design, while the US-EPA and DIN EN 481 collect larger particles than the EU-inlets but is also smoother in the smaller size.




Figure 2.14: The sampling efficiency curve for  $\text{PM}_{10}$  particles with different design criteria.

## 2.4. Equivalence procedure

The analysis of the comparison data was conducted according to the GDE. Before the analysis of the equivalence tests, a preliminary assessment needs to be made in order to ensure that the CM:

- fulfils the requirements of data capture and time coverage set for continuous/fixed measurements; and

- has the potential to meet the uncertainty criteria of the data quality objective at the limit or target value concentration for continuous or fixed measurements of the specified pollutant

After the candidate method has passed this preliminary assessment, the test and evaluation programme relevant to the candidate method can be selected. In the case of PM measurements, the tests are performed according to test programme number 3: Methods for particulate matter. Both laboratory and field tests need to be considered. Laboratory tests are needed in case if deviation from the EN-standard takes place either if:

- the automated filter changer maintains a different storage conditions of the sampling filters than prescribed or
- use of different weighing conditions of the filters in the weighing room or weighing chamber

It should also be pointed out that the limiting conditions for claiming of the PM equivalence are associated with site specific conditions, ranges of fractions of the constituent as well as the size or shape of the particles. This means that even though the generalization of equivalence claim is valid for a wide range of conditions and compounds, this is not the case for PM. The performance characteristics of the CM are influenced by environmental conditions, but also e.g. by the fraction of a semi-volatile constituent in the sample, which is site-specific and depends also on geographic location. Equivalence claims for a specific CM may thus not be applicable in general, but the CM may still be useable in specific conditions or at certain locations.

The test programme 3 of the GDE is suitable to evaluate a CM for monitoring the PM<sub>2.5</sub> and PM<sub>10</sub> fractions of total suspended particulates in ambient air. The equivalence claims can be focused on the sample inlets and/or the measurement method (e.g., β-ray attenuation, optical method, oscillating balance). Basically the procedure for claiming the equivalence of the CM against a RM is a method involving calibration. The term “correction factor” has been omitted from GDE. Field campaigns were performed with the

CMs that took part in the tests. Field tests were performed in such a way that the candidate and the reference methods were compared side-by-side. Two CMs of the same model, as well as two RMs, were included in the tests. The measurements were designed to assess:

- ‘between-sampler/instrument’ uncertainty of the candidate method through the use of two samplers or instruments
- ‘comparability’ of the candidate and reference methods

The GDE requires two identical CMs but allows one RM to be used for tests. However in this case a default value for the uncertainty of the results of the reference method  $u^2 = 0.67$  ( $\mu\text{g.m}^{-3}$ )<sup>2</sup> may be used, see Eq (2.1).

The evaluation of the collected data included the following steps:

#### A *Suitability of datasets*

According to GDE, data may only be removed from the dataset when sound technical reasons can be found for doing so. However, when suspicious data are found, as discovered by, e.g., Grubb’s test, it is permitted to remove up to 2.5 % of data pairs, as long as the number of valid data pairs per comparison is  $\geq 40$ .

Of the full dataset:

1.  $\geq 20$  % of the results obtained using the reference method  $\geq$  UAT

where UAT is the Upper Assessment Threshold for the annual limit value. For  $\text{PM}_{10}$ , the UAT is  $28 \mu\text{g/m}^3$ , while for  $\text{PM}_{2.5}$  it is  $17 \mu\text{g/m}^3$ .

### B Between-sampler/instrument uncertainty

The between-sampler/instrument uncertainty is determined for CM by calculating the differences of all 24 hour results for each pair of CM's. The standard uncertainty of the between-sampler/instrument uncertainty is then calculated from the differences as:

$$u_{bs}^2 = \frac{\sum_{i=1}^n (y_{i,1} - y_{i,2})^2}{2n} \quad (2.1)$$

where  $y_{i,1}$  and  $y_{i,2}$  are the averages of results of CM over 24 h period  $i$  and  $n$  is the number of 24 hour measurement results.

The  $u_{bs}$  is determined for the data set in the following way:

- for the complete dataset
- for splitted dataset:
  - for PM<sub>10</sub>: concentrations  $\geq 30 \mu\text{g}/\text{m}^3$
  - for PM<sub>2.5</sub>: concentrations  $\geq 18 \mu\text{g}/\text{m}^3$

A between-sampler/instrument uncertainty  $u_{bs} \leq 2.5 \mu\text{g}/\text{m}^3$  shall be met for both of the data sets for PM<sub>10</sub> and for PM<sub>2.5</sub>. If this is not achieved, it is an indication of unsuitable performance of either one or both samplers and instruments, and equivalence shall not be declared for the candidate method when this criterion is not satisfied.

For the reference method, RM, the between-sampler/instrument uncertainty,  $u_{bs,RM}$ , shall fulfill the criterion,  $u_{bs,RM} \leq 2.0 \mu\text{g}/\text{m}^3$ .

### C Comparison with the reference method

Evaluation of the uncertainty due to the ‘lack of comparability’ between candidate and reference methods is established as the average of both of the CM’s and that of each of the candidate instruments individually (GDE (09)) using a regression technique that leads to symmetrical treatment of both variables. The relationship between the measurement results of both methods can be described by a linear relation of the form:

$$y_i = a + bx_i \quad (2.2)$$

where  $a$  is the intercept and  $b$  is the slope of the linear line. The procedure of comparisons is applied to the average of RM and to each of the candidate instruments individually for the full dataset obtained and to a number of subsets:

- *full data set*
- *datasets representing PM concentrations*
  - $\geq 30 \mu\text{g}/\text{m}^3$  for  $\text{PM}_{10}$ , or
  - $\geq 18 \mu\text{g}/\text{m}^3$  for  $\text{PM}_{2.5}$ ,
- provided that the subset contains 40 or more valid data pairs*
- *datasets for each individual site*

Preconditions for acceptance of the full dataset are:

- the slope  $b$  is insignificantly different from 1:  $|b-1| \leq 2 \cdot u(b)$  and (2.3a)

- the intercept  $a$  is insignificantly different from 0:  $|a| \leq 2 \cdot u(a)$  (2.3b)

where the  $u(b)$  and  $u(a)$  are the standard uncertainties of the slope and intercept. If these preconditions are not met, the CM may be calibrated using the values obtained for the slope and/or intercept. The calibration shall only be applied to the full dataset. The algorithms for the calculation of the standard measurement uncertainties of  $u(a)$  and  $u(b)$  are related to the results of the individual CM against the results of the RM (see Annex b in GDE).

From above one can see that there is no general requirement regarding how much  $a$  can deviate from zero and  $b$  deviate from 1. Instead, the smaller  $u(a)$  and  $u(b)$  are, the smaller deviation is allowed in the significance test for the intercept ( $= 0 \pm a$ ) and for the slope ( $= 1 \pm b$ ), respectively. Therefore no rejection of the CM has been made based on the values of  $a$  and  $b$ .

The procedure is applied for each specific situation for which a separate claim is made, e.g. for specific site type (GDE). The tests of equivalence for each of the full data and for the subset data were performed with the Excel macro prepared for the data analysis by RIVM (R. Beijk, 2011). The macro includes calculation of the combined standard measurement uncertainty,  $u_{CR}$ . The combined standard uncertainty,  $u_{CR}$ , includes the uncertainty of the reference sampler, which is calculated from the between sampler value, see Eq (2.1). If only one RM is used,  $u_{bs,RM}^2 = 0.67 \text{ } (\mu\text{g}/\text{m}^3)^2$  shall be used. However,  $u_{bs,RM}^2 = 1.0 \text{ } (\mu\text{g}/\text{m}^3)^2$  is used as default value in any case. The equation of the  $u_{CR}$  can be expressed in the form:

$$u_{CR}(y_i) = \sqrt{\frac{RSS}{(n-2)} - u^2(x_i) + [a + (b-1)x_i]^2} \quad (2.4)$$

Where

RSS = the sum of (relative) residuals resulting from the orthogonal regression  
 $u(x_i)$  = uncertainty of the results of the reference method

where RSS can be expressed as:

$$RSS = \sum_{i=1}^n (y_i - a - bx_i)^2 \quad (2.5)$$

The relative combined standard uncertainty,  $w_{c,CM}$ , is then calculated as:

$$w_{c,CM}^2 = \frac{u_{CR}^2}{y_i^2} \quad (2.6)$$

where  $y_i$  is the limit value of  $\text{PM}_{2.5}$  or  $\text{PM}_{10}$  used for the calculation. The limit values used here are  $30 \text{ } \mu\text{g}/\text{m}^3$  for  $\text{PM}_{2.5}$  and  $50 \text{ } \mu\text{g}/\text{m}^3$  for  $\text{PM}_{10}$  (GDE). The calculated relative expanded measurement uncertainty,  $W_{CM}$ , can be expressed as  $W_{CM}=k \cdot w_{c,CM}$ , where the coverage factor  $k=2$ . When comparing  $W_{CM}$  with the data quality objective (DQO) of the AQD (see in Table 2.1), the following two cases apply:

1.  $W_{CM} \leq W_{DQO}$ : the CM is accepted as an equivalence method to the RM (2.7a)
2.  $W_{CM} > W_{DQO}$ : the CM is not accepted as an equivalence method to the RM (2.7b)

In case 2, the results of the CM can be corrected using the results from the regression equation, Eq. (2.2), obtained for the full dataset. The term calibration is used here for correcting the data. The form of calibration equations depends on the conditions in Eq (2.3a and b):

1.  $b$  and  $a$  significantly difference from 1 and 0, respectively:  $y_{i,cal} = \frac{y_i - a}{b}$  (2.8a)

2.  $b$  significantly different from 1:  $y_{i,cal} = y_i / b$  (2.8b)

3.  $a$  significantly different from 0:  $y_{i,cal} = y_i - a$  (2.8c)

After applying the calibrated equation, the recalculated relative expanded measurement uncertainty,  $W_{CM,cal}$  of the CM needs to satisfy requirement 1 in Eq (2.7a) for the full dataset as well as for each of the subsets. If this is not met, the CM is not a method equivalent to the reference method.

The calibration equations obtained for the candidate methods should also be used for the station PM-analyzer if applicable, to demonstrate the equivalency between the station method with the reference method.

The sampling treatment of the CMs is somewhat different with each other (see in Table 2.2). In some cases, the sample tube is heated to avoid condensation and in some case the sampling tube consist of shield flow to maintain the sample at ambient temperature to avoid condensation. In some CM the sample tube include the shield flow but is also equipped

with a drier to remove excessive water content in the sample flow, and in some CM there is no additional treatment for the sampling. The sampling treatment has a consequence also for the evaporation of semi-volatile components from the sample. Especially this is important in case of heated sample tubes since evaporation of the semi-volatile compounds occurred already at room temperature. The fraction of semi-volatile compounds in the air has been studied for inorganic ions in the size fraction of PM<sub>2.5</sub> in Helsinki in 2006–2007 (Timonen et al. 2014) and at a background station in the southern part of Finland during the period of 2003 to 2009 (Ruoho-Airola, 2012). The major semi-volatile compounds considered here are the NH<sub>4</sub><sup>+</sup>, NO<sub>3</sub><sup>-</sup>, and SO<sub>4</sub><sup>2-</sup> ions. In both studies a clear diurnal and seasonal behavior was observed. Sulphate was the most abundant ion with an average concentration of 1.7 µg/m<sup>3</sup>. The concentration of nitrate and ammonia were considerably lower i.e. 0.7 and 0.8 µg/m<sup>3</sup>, respectively. The sum of the average mass concentration of the most easily evaporated compounds i.e. ammonia and nitrate is 1.5 µg/m<sup>3</sup>. If half of the amount evaporates at 20 °C this means 10 % at the average concentration of PM<sub>2.5</sub> which was 7.3 µg/m<sup>3</sup> over the whole comparison period. On the other hand, the proportion of semi-volatile compounds to whole mass of PM<sub>2.5</sub> was estimated to be on 18 % based on the measurements by TEOM using the FDMS (Timonen et al. 2014).

### 3. Experimental set up

#### 3.1 Laboratory tests

The GDE defines the circumstances in which the laboratory tests are needed. The laboratory tests are not required for CM's but only for RM if deviations from the prescribed conditions according to EN-standard (EN 12341, 2014) occurred.

During these comparisons the reference sampler, SEQ47/50-CD, used for comparison studies was a sequential sampler which can be loaded with 15 sampling filters. The instrument is equipped with a Peltier cooler to maintain the sampled filters at temperature

$\leq 23$  °C as is required by the EN 12341:2014. In winter conditions the filter cartridge will be automatically heated.

The other requirement where the laboratory tests are needed is when the conditioning conditions of the sampling filters deviate from those prescribed by the EN-standard. The weighing system of the filters was made in house and validated according to the standard. The weighing system is also accredited according to ISO EN 17025:2005. The scope of accreditation covers the mass range of the sampled filters from 0.055 to 55 mg, which corresponds with a mass concentration of 1 to 200  $\mu\text{g}/\text{m}^3$  ([http://www.finias.fi/Scopes/K043\\_M11\\_2015.pdf](http://www.finias.fi/Scopes/K043_M11_2015.pdf)). The environmental conditions of the weighing system are controlled and recorded and actions are taken in case of deviations from prescribed conditions.

Based on the performance characteristics of the reference sampler and that the conditions inside the weighing facility of the filters followed the EN 12341 requirements, no additional laboratory tests were conducted (GDE, 2010).

### **3.2 Field campaigns and the measurement site**

The equivalence comparisons took place during March 2014 and July 2015. The site for comparison tests was selected in the city of Kuopio, some 400 km northeast from the city of Helsinki. The location represents inland conditions and is not influenced by marine climate. The city is surrounded by lakes and forest and agriculture area. There are some 110 000 inhabitants in Kuopio.



*Figure 3.1: The measurement sites in the city of Kuopio. The left map shows the local air quality networks in Finland and on the right the location of the two measurement sites in Kuopio are shown.*

The equivalence tests for the candidate measurement methods against the reference method for PM<sub>2.5</sub> and PM<sub>10</sub> were conducted at two sites in the city of Kuopio. The sites were selected based on a preliminary assessment study of the air quality in the city of Kuopio by dispersion modelling, and on the annual reports of air quality in the city of Kuopio. The exceedances of daily limit values of mass concentration of PM<sub>10</sub> occur regularly between 20 to 30 times per year. Mostly these exceedances occur in conditions, where due to the winter sanding to enable safe road traffic against the ice both for pedestrians on the sidewalks and for cars on the roads, the sand particles are resuspended into the air by the winds. This causes the elevated mass concentration particularly in the PM<sub>10</sub> size class. Not only the winter sanding but also the use of studded tires in cars causes particles that are resuspended into the air. At the Savilahdentie there is no existing air quality station but the site was selected for comparison study based on the dispersion model estimations which showed high concentrations of PM due to traffic. The test started at the air quality station in Tasavallankatu with two PM<sub>10</sub> comparison campaigns in series followed by two campaigns for PM<sub>2.5</sub>. The change of site to Savilahdentie took place right after finishing the campaigns in Tasavallankatu. Here the strategy for comparison was changed from that in Tasavallankatu. One reference sampler was installed with PM<sub>10</sub>-inlet and the other with

$\text{PM}_{2.5}$ -inlet through the rest of the campaigns. The parallel comparisons with only one RM are allowed by GDE but additional uncertainty component is required to be included in the uncertainty budget. The basic reason for the change of strategy was to provide the participation of two different models of CMs in the tests. Both of the CMs were able to measure both size classes at the same time. The change of strategy was agreed in very early stage of the comparison due to the lack of space at Tasavallankatu. However, at the beginning of the measurement campaigns at Savilahdenkatu only one of the two different CMs was submitted in the test, namely DustTrak model DRX 8535 by TSI Inc., which is an optical method. The other method was withdrawn from the comparison.

At Tasavallankatu the first  $\text{PM}_{10}$  comparison tests started in 2014 with a spring campaign on March 20 to May 10 consisting of 51 daily samples for RM. The second  $\text{PM}_{10}$  comparison, summer campaign, continued from May 11 to July 14, consisting of 52 daily samples for RM. The first  $\text{PM}_{2.5}$  campaign started with summer tests from July 17 to October 1 including 60 daily samples for RM. The second campaign for  $\text{PM}_{2.5}$  continued on October 2 and ended on December 1, including 60 daily samples for RM.

At Savilahdentie the comparisons continued with two campaigns for  $\text{PM}_{10}$  comparisons followed by two campaigns for  $\text{PM}_{2.5}$  comparisons for those CMs which have size selective inlets i.e. with FH 62 I-R, MP101\_CPM, SHARP, TEOM 1405 and BAM. At the beginning of the campaigns at Savilahdentie there was a short period when both of the RMs were installed with  $\text{PM}_{10}$  inlets to check the consistency of the samplers. The third campaign of  $\text{PM}_{10}$  comparisons for those having size selective inlets started on December 5, 2014 to February 5, 2015 including 50 daily samples for RM. The fourth  $\text{PM}_{10}$  comparisons started on February 6 to April 8 including 50 daily samples for RM. The third campaign of  $\text{PM}_{2.5}$  comparisons started on April 5 to May 21 including 41 daily samples for RM. The fourth campaign of  $\text{PM}_{2.5}$  comparisons started on May 22 to July 7 including 42 daily samples for RM. At Savilahdentie parallel measurements by reference method were conducted for  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$  measurements allowing Grimm, DustTrak and Osiris to conduct simultaneous comparison for  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$ . This meant that there were four comparisons for  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$  at Savilahdentie for the CMs following the optical methods but two

campaigns for  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$  using the size selective inlets. In Figure 3.2 the different field campaigns are shown.




Figure 3.2: The dates of the equivalence test campaigns for  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$  at Tasavallankatu and at Savilahdentie. At Savilahdentie the parallel measurements of  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$  by RM allowed simultaneous comparison of  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$  by CMs based on optical methods.

The measurement station at Tasavallankatu, see Figure 3.3, is located beside main road (6 m from the kerbside and 60 m from the closest crossing) from the motorway to the center of city of Kuopio. In the neighborhood, there is also a residential area with private houses. Most of the houses have their own heating system by wood. The main power plant, which takes care of heat production in the area, is located near the site. The energy source is peat pellet, which is transported with trucks. The traffic density at Tasavallankatu is some 22 000 vehicles/day. At the air quality station, mass concentration of  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$  and the concentration of nitrogen oxides are measured. The station was classified as a traffic station according to the AQD. The meteorological parameters, wind speed and direction, air temperature, pressure and relative humidity are also measured.



*Figure 3.3. The air quality measurement site at Tasavallankatu in the city of Kuopio. The measurement van and the cabin are shown in front. The green cabin in behind is for the air quality measurement station.*

The Savilahdentie site is a traffic station, 10 m from the kerbside and more than 200 m from the closest crossing. The station lies between the freeway from south to north splitting the city of Kuopio and the Savilahdentie which is the motorway connecting the industrial area, shopping centers, university campus area and the technopolis area to the city. The traffic density is about 21 000 vehicles/day, consisting mostly of working traffic but also some drive through traffic. The measurement cabins were installed beside a small relay pumping station.



*Figure 3.4. Savilahdentie site.*

### **3.3 Filter weighing procedure**

The EN 12341 standard describes the environmental conditions for filter conditioning during the filter weighing process. The environmental conditions prescribed by the standard are  $(20 \pm 1)^\circ\text{C}$  for the temperature and (45 to 50) % for the relative humidity. The weighing facility of the filters was made in house, consisting of the weighing chamber and the conditioning and control system. The schema of the operation is shown in Figure 3.5. The environmental condition inside the weighing facility is gained by injecting dry air obtained from the laboratory compressed air system through the water bath and the humidifying system into the weighing chamber. The dry air flow rate is controlled by mass flow controllers and injected through the Perma-pure humidifiers at temperatures controlled by water baths. The first water bath (VH1) and the humidifier (PPK1) are to reach the conditions in temperature and water content close to the requirements by the standard and the second (VH2 and PPK2) are used more for adjustments. The conditions are recorded by the dew point measurement probe (dew point and temperature transmitter, DMP 248) by Vaisala. The data was stored in the laboratory computer MICRO COMPUTER with software made in house. The filters were weighed by ultra-micro

balance (XP2U) by Mettler Toledo with repeatability of  $0.25 \mu\text{g}$  as standard deviation and a resolution of  $0.1 \mu\text{g}$ . The calibration of the laboratory balance carried out once a year by an accredited calibration laboratory through which the traceability of the weighing results is linked to the national standard. In addition, the tare of the balance was checked on a daily basis by an automated function of the balance. The balance is connected to the readout unit and to a laptop where the BalanceLink software by Mettler Toledo was installed for collecting weighing results of the filters. The filter trays inside the chamber were capable of carrying more than 30 filters inside the chamber. The filter weighing facility is shown in Figure 3.6.




Figure 3.5. Schematic layout of the control system for weighing the filters for the reference method.



*Figure 3.6: The weighing box for the filters and the control units for conditioning of temperature and humidity (on the left) with the balance (on the top middle) and the filter tray (on the right).*

The conditioning and the conditions during the weighing of the filters were kept within the requirements by the EN-standard EN 12341:2014. Once exceedance of the allowed limits occur the weighing was postpone until the conditions were acceptable.

### **3.4 Data acquisition system at the measurement site**

The data acquisition system used for the PM equivalence comparisons was a commercial EnviDas2000 for Windows (EnviDas, 2003) installed in a microcomputer. The data was collected as an average of 15 s values once a minute from every CM. The data protocols for each of the CMs were installed in the software with the help of the manufacturers. Two identical data acquisition systems were used so as to have all the CMs continuously connected. The station microcomputers were connected through a modem line with the server computer where the data management software, Enview 2000 (Enview, 2004), was installed. Enview software stored the data in the database of the FMI for further analysis.

### ***3.5 Operation of the instruments***

#### **3.5.1. Reference sampler**

Two identical sequential samplers, SEQ47/50 from Leckel (see Figure 2.2) were used as a reference sampler. Up to 15 filters were loaded in the filter cartridge. One filter was used as a field blank for each of the samplers and 14 filters were used for defining the mass concentration of PM during the 24 h sampling time. The storage condition for the filters in the reference sampler follows the conditions set up by the EN 12341 standard. The samplers were installed on the roof of the measurement cabin, shown in Figures 3.3 and 3.4. The sampling with both of the samplers started at 10 am on each day. The time interval of  $\pm 1$  h was reserved for the loading and unloading of the sampling filters for every two week period.

#### **3.5.2 CM operating conditions**

The candidate methods were installed inside the cabin or the measurement van. Exceptions to this were the Osiris and DustTrak, which were installed on the roof. The location of these two CMs was approved by both manufacturers. The sampling inlets of the CMs were installed at varying heights from the roof level, from 60 cm up to 180 cm. The free circle around the sampling inlets was 120 cm in diameter or larger. The conditions inside the cabin and the measurement van were controlled by an air-conditioning system maintaining the inside temperature between 20 to 25 °C through the whole comparison study. Power failures occurred especially in the beginning of the measurements when the site was changed. The UPS system protected the microcomputers for data collection but not the CMs.

#### **3.5.3 QA/QC procedures**

The QA/QC procedures for the CM and for the RM stated by the GDE are listed in Table 3.1. The sampling flow rates for the reference samplers and the CMs equipped with the

size selective inlets are presented in Annex A1, Table A1a. The flow rates were measured with a mass flow meter by TSI Inc, model 4043. The mass flow meter was calibrated at the calibration laboratory of the Finnish Meteorological Institute, which maintains the traceability of the primary calibration facility to the national metrology institute by regular calibration against the primary flow calibration method. The uncertainty of the flow calibration facility at the sampling flow rate of the size selective inlets is 0.7 % ([https://www.finias.fi/Documents/K043\\_M12\\_2016.pdf](https://www.finias.fi/Documents/K043_M12_2016.pdf)). The uncertainty of the mass flow meter of TSI model 4043 was estimated based on the calibration certificate as 1.5 % as expanded uncertainty. The flow measurement of SHARP (in Table A1a, SHARP A) turned out to be problematic due to the connection of the sampling tube which did not fit well with the flow meter. The connection was improved at the same time as the change of inlets. The temperature and the pressure for each of the flow measurements were also recorded and are presented in Tables A1b and A1c. The sampling flow rate of the CMs operated by optical methods were checked by the Finnish representative (Hnu Nordion Oy, for Osiris) or by the manufacturer (Grimm 180 and DustTrak). The sample flow rates of the CMs were also recorded by the data acquisition system where applicable.

The cleaning of the size selective inlets was conducted according to the recommendation by the manufacturers of the CMs. In case of RMs the cleaning of the inlets was conducted according to the EN-standard (EN-12341). The time schedule for cleaning of the filters is presented in Annex A1.d. The impactor plates were wiped by silicon vacuum grease after the cleaning of inlets in order to prevent the larger particles than the cut off size to bounce off the plate. The grease of the impactor plate needs to perform according to the EN-standard or according to the instructions by the manufacturer. There are some differences in the instructions by the manufacturers of the inlets e.g. in case of the US-EPA inlet the impactor plate is not greased, but instead the O-ring and the O-ring seat need to be greased according to the manufacturer's instruction (inlet manual BX 820-9800).

The Finnish representatives conducted the maintenance of the CMs except in case of BAM, where the maintenance was conducted by a technician from the EERC.

*Table 3.1. The QA/QC procedures conducted at the measurement sites for the CMs and RMs.*

Calibration, checks and maintenance	Frequency	GDE/FMI	Action criteria
Checks of status values of operational parameters	Daily	GDE	If error messages occurred, contacts to local network operator or representative.
Checks of sensors for temperatures, pressure and/or humidity	Every 3 months or according to manufacturer	GDE	If deviations larger than acceptable, contacts to local representative: T: $\pm 2$ °C P: $\pm 1$ kPa RH: $\pm 5$ % RH
Check of sampling velocity	Every 2 months	FMI	Adjustment if deviation > 4 %
Cleaning and greasing size selective inlets	Every 1 months (RM) Every 2 months (CMs)	FMI	
Zero check of the CM reading	Every year	FMI	2 $\mu\text{g}/\text{m}^3$
Calibration of the CM	According to manufacturer	FMI/ Representatives	
Maintenance of the CM	According to manufacturer	FMI/ Representatives	

### 3.5.4 Field audit

The field audit was conducted on the QA/QC procedures and the operation of the instruments. The audit was performed by external auditors at the measurement site as required by the GDE due to lack of accreditation of the Finnish Meteorological Institute. During the audit, the procedures for conducting the QA/QC procedures according to Table

3.1 were checked through demonstration of the activities at the site and examining the logbook. The auditors were experienced in air quality measurements, content of the relevant EN-standard, quality systems and are responsible for the air quality network in the city of Kuopio.

## 4. Data analysis

### 4.1 Description of the database

The data from the CMs was collected as minute readings by commercial EnviDas data acquisition system. The data was stored in the memory of the station microcomputer and transmitted over a GSM modem connection once a day to the data server at FMI. A few exceptions from the minute values occurred. In case of DustTrak, the data was stored in the memory of the device where it was copied to the memory stick every two weeks and stored on the server of FMI for data analysis. The frequency of the data collection from DustTrak was every 4 seconds. In case of one of the Grimm instruments, the data needed to be collected through the HyperTerminal software into the station micro computer. In case of the BAMs, the sampling time was set to three hours which means that the data values were constant during this period reaching another value after each 3 hour period. The change of period was synchronized with the time of change of the filters in the RM. In the data server of FMI, the Envview software stored the data in the database. In case of the RM, the results of the sampling, i.e. the sampling time, the sampling flow rate, and the sampling volume were collected every time when filters were changed from the memory stick of the RM and stored on the on the server computer of FMI.

The data base of the equivalence tests for PM<sub>2.5</sub> and PM<sub>10</sub> includes the minute values, hourly values, and 24-hourly values for the CMs. The 24-hourly values were calculated during the sampling time of the RM. In addition to the mass concentration of the CMs, also data of the diagnostic information of CMs was collected. Such diagnostic information

includes the sampling flow rate, the temperature, the relative humidity, and the signals from optical-, combined- or beta-methods and the cumulative mass concentration were collected when applicable.

## **4.2 Data processing**

During the process of calculating the hourly and 24 h averages, the data was inspected at every step of the calculation, starting from the minute values to see the response of each CM. The averages for the different time intervals were performed, provided that 75 % of the data was available over each calculation period. The notes and observations made in the logbooks of the RMs or CMs at both of the measurement cabins were studied. If technical reasons for malfunctioning of any of the RMs or CMs were observed, the data was flagged. In case of the RMs, no technical problems in the operation of the devices occurred. In case of the operation of CMs, several of the reasons e.g. power and pump failures, malfunctioning of measurement probe pointed out by diagnostic information, moisturizing of resulted in incorrect behavior of the CMs. These malfunctioning were checked, and the calculation of the average values was rejected if requirements for averaging the results were not met. The 24 h means for each of the CM were calculated for the same period as RMs.

The following data treatments were performed for PM<sub>10</sub> and PM<sub>2.5</sub> (see Ch. 2.4):

- Data capture, (AQD)	Pass/Fail
- Test of suitability of the data, GDE	Pass/Fail
- Test of between-sampler for each pair of CMs, GDE	Pass/Fail
- Test of between-sampler for RMs, GDE	Pass/Fail
- Test of comparability, GDE	
- uncorrected data	
- all data (averages of RMs and CMs over whole test campaigns)	Pass: OK


	Fail: Calibration
- subsets	Pass/Fail
- each of the CMs individually	Pass/Fail
- calibrated data (if ‘all data’ failed)	
- all data	Pass/Fail
- subsets	Pass/Fail
- each of the CMs individually	Pass/Fail

The tests of comparability were made with reference to the concentration value of  $30 \mu\text{g}/\text{m}^3$  for  $\text{PM}_{2.5}$  and that of  $50 \mu\text{g}/\text{m}^3$  (the daily limit value) for  $\text{PM}_{10}$ , according to GDE. If the test failed to meet the DQO criteria for fixed measurements (see Eq. 2.7b), a check was made for the CM as to whether the test value,  $W_{CM}$ , fulfilled the DQO for indicative measurements (see Table 2.1). It also occurred that some of the points did not seem to fit with the rest of the data, and therefore Grubb’s test or the judgment of an expert was used to remove some of the data points. As stated in GDE, no more than 2.5 % of the data could be removed based on this method.


## 5. Results

### 5.1 Meteorological conditions during the equivalence comparisons

The spring 2014 was rather mild compared with the climatological reference period 1981–2010 (Pirinen et al. 2012). The daily mean temperatures (T) and relative humidity (RH %) during the comparison periods in Tasavallankatu and in Savilahdentie are presented in Figures 5.1a and b together with the average of the data from the climatological reference period (1981 – 2010) from Kuopio.



**Figure 5.1a:** The 24 h mean values of temperature (scale on the left) and relative humidity (scale on the right) during the equivalence campaigns for  $PM_{10}$  and  $PM_{2.5}$  measurements at Tasavallankatu. The dash lines represents the average of air temperature and relative humidity over the climatological reference period (1981 – 2010) from Kuopio.



**Figure 5.1b:** The 24 h mean values of temperature (scale on the left) and relative humidity (scale on the right) during the equivalence campaigns for  $PM_{10}$  and  $PM_{2.5}$  measurements at Savilahtentie. The dash lines represents the average of air temperature and relative humidity over the climatological reference period (1981 – 2010) from Kuopio.

The roses for wind speed, wind direction, PM<sub>10</sub>- and PM<sub>2.5</sub> concentrations are presented from Tasavallankatu in Figure 5.2.a and from Savilahdentie in Figure 5.2.b.





Figure 5.2.a. Wind direction, wind speed, PM<sub>10</sub> concentration and PM<sub>2.5</sub> concentration as a function of wind direction during the equivalence campaigns for PM<sub>10</sub> and PM<sub>2.5</sub> measurements at Tasavallankatu.

In Tasavallankatu site the wind sector is very narrow indicating that near the ground the wind is along the road, perpendicular to Tasavallankatu (see Figure 3.1) where the strongest winds are blowing from the Tasavallankatu. The PM<sub>10</sub> concentrations are at the highest at winds from the Tasavallankatu, while the PM<sub>2.5</sub> concentrations are more evenly distributed to all wind sectors. In Savilahdentie the dominating wind sector is from the lake across the Savilahdentie and from the freeway (see in Figure 3.1). The wind speed is rather evenly distributed to all wind sectors. With the PM<sub>10</sub> concentrations, a clear peak comes from the


Savilahdentie and from the freeway. In wind sector  $30^\circ$  the particles are emitted from both of the roads. The  $\text{PM}_{2.5}$  concentration pattern is very similar to that of the  $\text{PM}_{10}$  concentration but smoother and with much lower concentration.




*Figure 5.2.b. Wind direction, wind speed,  $\text{PM}_{10}$  concentration and  $\text{PM}_{2.5}$  concentration as a function of wind direction during the equivalence campaigns for  $\text{PM}_{10}$  and  $\text{PM}_{2.5}$  measurements at Savilahdentie.*

## 5.2 Results of the PM<sub>10</sub> comparison

The equivalence tests of the candidate measurement methods against the reference method started with the field campaigns for the PM<sub>10</sub> size class. The timing and the duration of the different campaigns of both size classes are shown in Figure 3.2. The PM<sub>10</sub> comparison started with a spring campaign March 20, 2014. The timing was important for the PM<sub>10</sub> size class since resuspension of particles from the ground is the major source to mass concentration in this size class. In March, the weather conditions can typically include high-pressure situations where the temperature can be low during the night, causing an inversion. During the day, the temperature increases as the sunrises helping the mixing of the boundary layer. Dry air does not bound particles but rather accumulate them in the boundary layer causing the elevated PM concentration. The content of the sand particles are mostly silicon, which may give a different response with the CM analyzers compared to a normal case where the particles are generated and formed from other processes and sources. Later in spring the city authorities are cleaning



*Figure 5.3a: The time series of the two RMs for the PM<sub>10</sub> comparisons at Tasavallankatu. LV is the limit value for 24 h mass concentration for PM<sub>10</sub> and UAT is the upper assessment value for PM<sub>10</sub>.*



*Figure 5.3b. The time series of the two RMs for the PM<sub>10</sub> comparisons at Savilahdentie. LV is the limit value for 24 h mass concentration for PM<sub>10</sub> and UAT is the upper assessment threshold value for PM<sub>10</sub>.*

the roads and pedestrian walkways from winter sand, which can also cause short peaks in the mass concentrations of particularly PM<sub>10</sub>.

In Figure 5.3, the time series of daily concentrations of the RMs for PM<sub>10</sub> for two campaigns at Tasavallankatu (5.3.a) and at Savilahdentie (5.3.b) are shown. The annual Upper Assessment Threshold value of 28 µg/m<sup>3</sup>, and the daily limit value for PM<sub>10</sub> of 50 µg/m<sup>3</sup> are shown. Especially in Figure 5.3.b the episode from the inversion situation is clearly seen. The scatter plot of RM2 against RM1 is shown in Figure 5.4. The standard uncertainty for the between sampler test for the reference method, according to eq. (2.1), is 0.6 µg/m<sup>3</sup> for all data and for the data of > 30 µg/m<sup>3</sup>.




Figure 5.4. The scatter plot of RM2 against RM1 for  $\text{PM}_{10}$  at Tasavallankatu.

### 5.1.3. Results of the candidate methods for $\text{PM}_{10}$

In the following tables each of the CMs have been analyzed against the reference method as described in chapter 4.2. The analysis includes the data capture, suitability of data, between CM tests and comparability test. The comparability test includes the tests of significance for slope and intercept indicating if the calibration function for CM is required. The comparability tests are performed for each of the campaigns as well as to two different category of the data i.e. larger than  $30 \mu\text{g}/\text{m}^3$  and less than  $100 \mu\text{g}/\text{m}^3$ . The calibration equation according to Eq. (2.8a) is expressed in any case i.e. irrespectively if the slope  $b$  and/or  $a$  are significantly different from 1 and 0, respectively. The results that exceeds the acceptable limit, see in chapter 2.4, are printed in red color.

*Table 5.1: The test results for the BAM impactor in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.*

Test PM <sub>10</sub>	Criteria	BAM A + B	BAM A	BAM B	Pass/Fail
Data capture	≥ 90 % of RM	98,6 %	98,1 %	98,6 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	1,32			Pass
> 30 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	2,03			Pass
Test of comparability					
All data					
Number of data	Significant	211	210	211	
Slope		1,06	1,07	1,06	
Intercept	Significant	-0,5	0,1	-1,0	
Calibrated data					
All data					
Number of data		211	210	211	
Calibration		0,942y + 0,437	0,935y -0,065	0,948y + 0,922	
Expanded relative uncertainty	25 %	12,6 %	13,2 %	11,9 %	Pass
Tasavallankatu: Tasa 1					
Number of data		53	53	53	
Calibration		0,800y + 1,976	0,795y + 1,506	0,801y + 2,523	
Expanded relative uncertainty	25 %	11,5 %	13,1 %	10,6 %	Pass
Tasavallankatu: Tasa 2					
Number of data		54	54	54	
Calibration		0,935y + 0,92	0,923y + 0,436	0,946y + 1,439	
Expanded relative uncertainty	25 %	5,7 %	7,1 %	4,7 %	Pass
Savilahdentie: Savi 1					
Number of data		55	54	55	
Calibration		1,020y + 1,412	1,004y + 1,099	1,007y + 1,933	
Expanded relative uncertainty	25 %	6,0 %	6,9 %	5,2 %	Pass
Savilahdentie: Savi 2					
Number of data		49	49	49	
Calibration		0,950y -0,026	0,944y -0,538	0,956y + 0,495	
Expanded relative uncertainty	25 %	17,7 %	18,5 %	17,1 %	Pass
< 100 µg/m <sup>3</sup>					
Number of data		205	204	205	
Calibration		0,858y + 1,919	0,842y + 1,621	0,865y + 2,336	
Expanded relative uncertainty	25 %	10,3 %	10,6 %	9,6 %	Pass
> 30 µg/m <sup>3</sup>					
Number of data		46	45	46	
Calibration		0,961y -1,951	0,956y -2,779	0,967y -1,439	
Expanded relative uncertainty	25 %	22,9 %	23,5 %	22,0 %	Pass

*Table 5.2: The test results for DustTrak in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.*

Test PM <sub>10</sub>	Criteria	DustTrak A + B	DustTrak A	DustTrak B	Pass/Fail
Data capture	≥ 90 % of RM	100,0 %	100,0 %	99,4 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data > 30 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup> u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	5,94 2,07			Fail Pass
Test of comparability					
All data					
Number of data		171	171	170	
Slope	Significant	0,13	0,14	0,13	
Intercept	Significant	10,3	8,6	12,0	
Calibrated data					
All data					
Number of data		171	171	170	
Calibration		7,478y -76,819	7,182y -61,667	7,808y -93,767	
Expanded relative uncertainty	25 %	402,3 %	406,8 %	408,4 %	Fail
Savilahdentie: Savi 1-1					
Number of data		45	45	45	
Calibration		1,984y -18,143	1,956y -17,541	2,009y -18,703	
Expanded relative uncertainty	25 %	116,9 %	118,2 %	115,6 %	Fail
Savilahdentie: Savi 2-2					
Number of data		44	44	44	
Calibration		9,149y -126,854	8,579y -100,982	9,800y -156,385	
Expanded relative uncertainty	25 %	551,8 %	500,2 %	618,4 %	Fail
Savilahdentie: Savi 1-3					
Number of data		41	41	41	
Calibration		7,824y -44,783	7,725y -28,415	7,918y -61,491	
Expanded relative uncertainty	25 %	175,3 %	173,6 %	183,8 %	Fail
Savilahdentie: Savi 2-4					
Number of data		41	41	40	
Calibration		0,385y + 6,712	0,371y + 7,782	0,411y + 5,297	
Expanded relative uncertainty	25 %	106,0 %	104,4 %	102,8 %	Fail
< 100 µg/m <sup>3</sup>					
Number of data		164	164	163	
Calibration		5,761y -55,073	5,563y -43,744	5,951y -67,193	
Expanded relative uncertainty	25 %	1132,0 %	1194,7 %	1099,9 %	Fail
> 30 µg/m <sup>3</sup>					
Number of data		35	35	35	
Calibration		8,278y -90,352	7,964y -74,566	8,613y -107,353	
Expanded relative uncertainty	25 %	355,9 %	365,7 %	348,8 %	Fail

*Table 5.3: The test results for FH 62 I-R in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.*

Test PM <sub>10</sub>	Criteria	FH 62 A + B	FH 62 A	FH 62 B	Pass/Fail
Data capture	≥ 90 % of RM	91,1 %	88,8 %	91,1 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	2,7			Fail
> 30 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	5,2			Fail
Test of comparability					
All data					
Number of data	Significant	195	190	195	
Slope		0,77	0,81	0,73	
Intercept	Significant	0,7	0,3	1,0	
Calibrated data					
All data					
Number of data		195	190	195	
Calibration		1,300y -0,904	1,241y -0,321	1,363y -1,436	
Expanded relative uncertainty	25 %	16,5 %	18,3 %	15,9 %	Pass
Tasavallankatu: Tasa 1					
Number of data		51	51	51	
Calibration		1,378y -2,070	1,415y -2,924	1,328y -0,988	
Expanded relative uncertainty	25 %	14,0 %	18,7 %	12,1 %	Pass
Tasavallankatu: Tasa 2					
Number of data		50	50	50	
Calibration		1,211y + 0,001	1,222y + 0,11	1,196y -0,035	
Expanded relative uncertainty	25 %	13,6 %	16,2 %	11,7 %	Pass
Savilahdentie: Savi 1					
Number of data		47	42	47	
Calibration		1,366y -2,081	1,286y -1,547	1,442y -2,365	
Expanded relative uncertainty	25 %	21,1 %	22,6 %	19,5 %	Pass
Savilahdentie: Savi 2					
Number of data		47	47	47	
Calibration		1,294y -0,067	1,237y -0,200	1,356y + 0,084	
Expanded relative uncertainty	25 %	22,4 %	23,6 %	21,5 %	Pass
< 100 µg/m <sup>3</sup>					
Number of data		189	183	188	
Calibration		1,372y -1,850	1,333y -1,641	1,413y -2,120	
Expanded relative uncertainty	25 %	17,1 %	18,8 %	16,2 %	Pass
> 30 µg/m <sup>3</sup>					
Number of data		44	44	44	
Calibration		1,281y + 1,574	1,216y + 3,057	1,353y -0,040	
Expanded relative uncertainty	25 %	21,3 %	22,4 %	22,6 %	Pass

Table 5.4: The test results for Grimm in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.

Test PM <sub>10</sub>	Criteria	Grimm B + C	Grimm B	Grimm C	Pass/Fail
Data capture	≥ 90 % of RM	92,7 %	92,7 %	92,7 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	2,09			Pass
> 30 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	4,48			Fail
Test of comparability					
All data					
Number of data		179	179	179	
Slope	Significant	1,17	1,17	1,18	
Intercept	Significant	-2,5	-2,7	-2,4	
Calibrated data					
All data					
Number of data		179	179	179	
Calibration		0,855y + 2,139	0,788y + 3,02	0,850y + 2,075	
Expanded relative uncertainty	25 %	17,0 %	17,7 %	18,0 %	Pass
Tasavallankatu: Tasa 2					
Number of data		46	46	46	
Calibration		1,024y + 0,74	0,958y + 0,643	1,091y + 0,965	
Expanded relative uncertainty	25 %	10,6 %	12,1 %	10,7 %	Pass
Savilahdentie: Savi 1					
Number of data		49	49	49	
Calibration		1,561y -3,490	1,776y -4,433	1,381y -2,653	
Expanded relative uncertainty	25 %	21,4 %	24,7 %	20,2 %	Pass
Savilahdentie: Savi 2					
Number of data		43	43	43	
Calibration		0,778y + 2,329	0,768y + 3,053	0,784y + 1,671	
Expanded relative uncertainty	25 %	16,3 %	15,8 %	18,9 %	Pass
Savilahdentie: Savi 3					
Number of data		41	41	41	
Calibration		0,969y + 0,732	0,989y + 1,143	0,946y + 0,426	
Expanded relative uncertainty	25 %	18,7 %	18,6 %	20,4 %	Pass
< 100 µg/m <sup>3</sup>					
Number of data		178	178	178	
Calibration		0,871y + 1,927	0,883y + 1,915	0,854y + 2,02	
Expanded relative uncertainty	25 %	17,0 %	17,1 %	18,2 %	Pass
> 30 µg/m <sup>3</sup>					
Number of data		27	27	27	
Calibration		0,701y + 13,052	0,689y + 13,974	0,700y + 12,88	
Expanded relative uncertainty	25 %	42,4 %	43,8 %	44,9 %	Fail

*Table 5.5: The test results for MP101 in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.*

Test PM <sub>10</sub>	Criteria	MP101 A + B	MP101 A	MP101 B	Pass/Fail
Data capture	≥ 90 % of RM	98,0 %	90,7 %	97,1 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	2,31			Pass
> 30 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	4,32			Fail
Test of comparability					
All data					
Number of data	Significant	201	186	199	
Slope		1,23	1,19	1,25	
Intercept	Significant	-2,9	-2,5	-2,8	
Calibrated data					
All data					
Number of data		201	186	199	
Calibration		0,811y + 2,311	0,838y + 2,102	0,800y + 2,242	
Expanded relative uncertainty	25 %	11,0 %	10,3 %	11,6 %	Pass
Tasavallankatu: Tasa 1					
Number of data		50	48	49	
Calibration		0,897y + 0,486	0,918y -0,055	0,886y + 0,538	
Expanded relative uncertainty	25 %	10,7 %	11,0 %	10,5 %	Pass
Tasavallankatu: Tasa 2					
Number of data		52	52	52	
Calibration		1,037y -0,610	1,014y -0,395	1,059y -0,806	
Expanded relative uncertainty	25 %	7,9 %	8,0 %	8,4 %	Pass
Savilahdentie: Savi 1					
Number of data		49	40	49	
Calibration		0,974y -0,687	0,975y -0,663	0,985y -0,794	
Expanded relative uncertainty	25 %	8,0 %	8,1 %	9,8 %	Pass
Savilahdentie: Savi 2					
Number of data		50	46	49	
Calibration		0,803y + 2,371	0,828y + 2,869	0,794y + 1,457	
Expanded relative uncertainty	25 %	12,4 %	12,3 %	11,5 %	Pass
< 100 µg/m <sup>3</sup>					
Number of data		194	180	193	
Calibration		0,887y + 0,826	0,910y + 0,64	0,860y + 1,086	
Expanded relative uncertainty	25 %	9,4 %	8,9 %	11,1 %	Pass
> 30 µg/m <sup>3</sup>					
Number of data		47	46	46	
Calibration		0,788y + 6,223	0,815y + 5,547	0,776y + 6,13	
Expanded relative uncertainty	25 %	13,5 %	12,8 %	15,2 %	Pass

Table 5.6: The test results for Osiris in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.

Test PM <sub>10</sub>	Criteria	Osiris A + B	Osiris A	Osiris B	Pass/Fail
Data capture	≥ 90 % of RM	95,3 %	95,0 %	95,3 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	2,63			Fail
> 30 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	5,92			Fail
Test of comparability					
All data					
Number of data	Significant	284 0,71	283 0,74	284 0,69	
Slope	Significant	0,1	-0,3	0,5	
Intercept					
<b>Calibrated data</b>					
<i>All data</i>					
Number of data		284	283	284	
Calibration		1,401y -0,153	1,346y + 0,462	1,456y -0,795	
Expanded relative uncertainty	25 %	15,7 %	17,3 %	18,8 %	Pass
<i>Tasavallankatu: Tasa 1</i>					
Number of data		51	52	51	
Calibration		1,144y + 1,862	1,219y + 2,359	1,067y + 1,624	
Expanded relative uncertainty	25 %	12,1 %	11,4 %	14,1 %	Pass
<i>Tasavallankatu: Tasa 2</i>					
Number of data		50	50	50	
Calibration		1,228y + 0,528	1,276y + 1,142	1,182y -0,023	
Expanded relative uncertainty	25 %	9,1 %	8,6 %	9,9 %	Pass
<i>Savilahdentie: Savi 1</i>					
Number of data		49	48	49	
Calibration		1,948y -0,394	2,046y -0,805	1,855y -0,005	
Expanded relative uncertainty	25 %	14,4 %	14,5 %	15,0 %	Pass
<i>Savilahdentie: Savi 2</i>					
Number of data		51	51	51	
Calibration		1,396y + 2,573	1,340y + 3,468	1,456y + 1,624	
Expanded relative uncertainty	25 %	18,1 %	20,7 %	18,6 %	Pass
<i>Savilahdentie: Savi 3</i>					
Number of data		40	40	41	
Calibration		1,327y + 1,042	1,184y + 0,991	1,488y + 0,882	
Expanded relative uncertainty	25 %	9,6 %	9,2 %	9,4 %	Pass
<i>Savilahdentie: Savi 4</i>					
Number of data		43	42	42	
Calibration		0,886y + 2,559	0,699y + 3,249	1,197y + 1,494	
Expanded relative uncertainty	25 %	16,9 %	21,3 %	14,8 %	Pass
< 100 µg/m <sup>3</sup>					
Number of data		277	276	277	
Calibration		1,338y + 0,57	1,294y + 1,058	1,360y + 0,308	
Expanded relative uncertainty	25 %	15,3 %	17,1 %	17,7 %	Pass
> 30 µg/m <sup>3</sup>					
Number of data		48	48	48	
Calibration		1,415y -1,367	1,342y + 1	1,494y -4,109	
Expanded relative uncertainty	25 %	24,7 %	26,9 %	31,3 %	Fail

*Table 5.7: The test results for Sharp beta in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.*

Test PM <sub>10</sub>	Criteria	Sharp beta A + B	Sharp beta A	Sharp beta B	Pass/Fail
Data capture	≥ 90 % of RM	99,0 %	96,6 %	98,5 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	1,28			Pass
> 30 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	2,26			Pass
Test of comparability					
All data					
Number of data		203	198	202	
Slope	Significant	0,71	0,72	0,69	
Intercept	Significant	1,6	1,5	1,9	
Calibrated data					
All data					
Number of data		203	198	202	
Calibration		1,415y -2,233	1,394y -2,063	1,439y -2,607	
Expanded relative uncertainty	25 %	12,8 %	11,2 %	13,7 %	Pass
Tasavallankatu: Tasa 1					
Number of data		51	50	50	
Calibration		1,471y -2,871	1,488y -3,288	1,451y -2,375	
Expanded relative uncertainty	25 %	6,2 %	6,2 %	7,6 %	Pass
Tasavallankatu: Tasa 2					
Number of data		52	52	52	
Calibration		1,194y -0,642	1,224y -0,568	1,163y -0,681	
Expanded relative uncertainty	25 %	7,2 %	7,7 %	7,6 %	Pass
Savilahdentie: Savi 1					
Number of data		49	47	49	
Calibration		1,803y -5,429	1,706y -4,788	1,734y -4,849	
Expanded relative uncertainty	25 %	11,6 %	10,7 %	11,0 %	Pass
Savilahdentie: Savi 2					
Number of data		51	49	51	
Calibration		1,405y -1,105	1,387y -1,554	1,430y -1,067	
Expanded relative uncertainty	25 %	16,3 %	15,6 %	16,9 %	Pass
< 100 µg/m <sup>3</sup>					
Number of data		197	193	196	
Calibration		1,489y -3,301	1,450y -2,879	1,498y -3,456	
Expanded relative uncertainty	25 %	12,5 %	10,6 %	13,5 %	Pass
> 30 µg/m <sup>3</sup>					
Number of data		47	44	47	
Calibration		1,396y + 0,002	1,376y -0,081	1,427y -1,084	
Expanded relative uncertainty	25 %	23,5 %	20,1 %	24,8 %	Pass

Table 5.8: The test results for Sharp c-dust in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.

Test PM <sub>10</sub>	Criteria	Sharp c-dust A + B	Sharp c-dust A	Sharp c-dust B	Pass/Fail
Data capture	≥ 90 % of RM	96,1 %	91,2 %	95,6 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data	U <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	1,60			Pass
> 30 µg/m <sup>3</sup>	U <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	3,10			Fail
Test of comparability					
All data					
Number of data	Significant	197	187	196	
Slope	Significant	0,71	0,74	0,70	
Intercept	Significant	2,0	1,6	2,2	
Calibrated data					
All data					
Number of data		197	187	196	
Calibration		1,404y -2,750	1,344y -2,104	1,431y -3,130	
Expanded relative uncertainty	25 %	17,2 %	14,7 %	17,5 %	Pass
Tasavallankatu: Tasa 1					
Number of data		49	45	48	
Calibration		1,553y -4,890	1,558y -5,233	1,510y -3,780	
Expanded relative uncertainty	25 %	17,8 %	22,3 %	17,3 %	Pass
Tasavallankatu: Tasa 2					
Number of data		52	52	52	
Calibration		1,175y -0,678	1,208y -0,745	1,142y -0,585	
Expanded relative uncertainty	25 %	9,9 %	10,6 %	9,8 %	Pass
Savilahdentie: Savi 1					
Number of data		47	46	47	
Calibration		1,648y -4,554	1,477y -3,395	1,635y -4,364	
Expanded relative uncertainty	25 %	12,8 %	12,5 %	14,3 %	Pass
Savilahdentie: Savi 2					
Number of data		49	44	49	
Calibration		1,401y -2,273	1,340y -2,082	1,431y -2,535	
Expanded relative uncertainty	25 %	26,0 %	18,6 %	25,0 %	Pass
< 100 µg/m <sup>3</sup>					
Number of data		191	184	190	
Calibration		1,486y -3,904	1,400y -2,891	1,479y -3,770	
Expanded relative uncertainty	25 %	16,5 %	15,1 %	17,1 %	Pass
> 30 µg/m <sup>3</sup>					
Number of data		41	34	41	
Calibration		1,385y -0,174	1,319y + 1,074	1,423y -1,638	
Expanded relative uncertainty	25 %	30,3 %	23,4 %	30,3 %	Fail

Table 5.9: The test results for TEOM in all the tests for demonstration of equivalence for PM<sub>10</sub> measurements.

Test PM <sub>10</sub>	Criteria	TEOM A + B	TEOM A	TEOM B	Pass/Fail
Data capture	≥ 90 % of RM	94,6 %	89,8 %	94,6 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass	Pass
Between sampler test					
All data	U <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	1,83			Pass
> 30 µg/m <sup>3</sup>	U <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	3,38			Fail
Test of comparability					
All data					
Number of data	Significant	194	184	194	
Slope	Significant	1,15	1,17	1,13	
Intercept	Significant	2,4	2,3	2,5	
Calibrated data					
All data					
Number of data		194	184	194	
Calibration		0,868y -2,068	0,852y -1,981	0,884y -2,215	
Expanded relative uncertainty	25 %	14,4 %	13,9 %	15,2 %	Pass
Tasavallankatu: Tasa 1					
Number of data		48	48	48	
Calibration		0,730y + 0,161	0,715y + 0,223	0,745y + 0,135	
Expanded relative uncertainty	25 %	14,3 %	13,9 %	14,9 %	Pass
Tasavallankatu: Tasa 2					
Number of data		48	48	48	
Calibration		0,881y -1,222	0,875y -1,363	0,888y -1,071	
Expanded relative uncertainty	25 %	9,2 %	9,7 %	8,7 %	Pass
Savilahdentie: Savi 1					
Number of data		48	46	48	
Calibration		1,064y -3,287	1,031y -3,006	1,099y -3,617	
Expanded relative uncertainty	25 %	8,7 %	9,9 %	7,9 %	Pass
Savilahdentie: Savi 2					
Number of data		50	42	50	
Calibration		0,873y -2,912	0,857y -2,621	0,891y -3,414	
Expanded relative uncertainty	25 %	18,1 %	16,8 %	20,2 %	Pass
< 100 µg/m <sup>3</sup>					
Number of data		187	177	187	
Calibration		0,804y -0,623	0,792y -0,600	0,813y -0,604	
Expanded relative uncertainty	25 %	13,6 %	13,4 %	14,0 %	Fail
> 30 µg/m <sup>3</sup>					
Number of data		44	41	44	
Calibration		0,878y -3,541	0,861y -3,236	0,896y -3,853	
Expanded relative uncertainty	25 %	26,8 %	26,3 %	28,2 %	Fail

### 5.1.4 Summary of the PM<sub>10</sub> comparisons

The number of sites and the number of campaigns cover the requirements of the GDE for PM<sub>10</sub> comparison tests. The meteorological conditions over the campaigns periods represents the annual variation met in southern Finland. The concentration ranges at both sites in Kuopio were satisfactory, fulfilling the suitability concentration criterion defined by the GDE. During the episode situation in late winter, the PM<sub>10</sub> concentrations exceeded the limit values at Savilahdentie where the daily concentrations reached 300 µg/m<sup>3</sup>, which was an exceptionally high concentration. The episode, mainly caused by the resuspended sand particles, lasted from the 9<sup>th</sup> to 18<sup>th</sup> in March 2015. The chemical content of the sand particles during the episode was rather homogeneous including mostly silica. This has an influence on some of the CM e.g. those using the optical method, since the refraction index and density can be different from the particles that are used for the calibration of the instrument. The exceptional episode was also the reason to provide an additional range from 0 to 100 µg/m<sup>3</sup> for the calibration equations for the CMs. As seen from Figure 5.3 the range up to 100 µg/m<sup>3</sup> excludes the data only from the episode period leaving most of the data still in the data analysis. In most of the cases the uncertainty of the measurement results is the same or slightly decreased and only minor change for the slope occurs if any. One should also keep in mind that the scope of the reference method according to EN 12341 is from 0 to 150 µg/m<sup>3</sup> for PM<sub>10</sub>.

During the PM<sub>10</sub> comparison the CMs were functioning reasonably well. The one Grimm instrument was delayed from the beginning of the campaign due to the calibration of the instrument at the manufacturer, which was the cause to discard the first campaign of Tasa 1 for Grimm (see table 5.4). Also one of the meteorological sensors of Grimm went out of order causing a short period of lack of data.

To have a better view over the for PM<sub>10</sub> concentration ranges the monthly mean PM<sub>10</sub> concentrations from different parts in Finland together with the values obtained in this comparison are presented in Figure 5.5.




Figure 5.5: The monthly mean concentration of PM<sub>10</sub> in 2014-15 in Helsinki (blue dot), in Oulu (orange dot), in Vaasa (grey dot), background station in Virolahti (brown dot), and in the PM<sub>10</sub> equivalence tests (red square) in Kuopio.

Figure 5.5 shows that the concentration ranges from the comparison joins very well with the concentration ranges at different sites across Finland in spite of lack of data during the months in autumn.

Regarding the results for demonstration of equivalence, Tables 5.1 to 5.9, the data capture should be 90% of the reference method. This requirement is fulfilled with every CM except for FH 62 A which has a data capture of 89 % of the RM. Suitability requirement is fulfilled by all CMs. Between the CM, a test is performed to all data and to data with the concentration above 30 µg/m<sup>3</sup>. The BAM and SHARP β-signals completely fulfilled this requirement. The Grimm, MP101, SHARP C-dust signal and the TEOM passed the requirement for all data, but failed with the data where the concentration was above 30

$\mu\text{g}/\text{m}^3$ . The DustTrak, FH 62 and Osiris failed with both of the requirements. The additional range between 0 to 100  $\mu\text{g}/\text{m}^3$  was analyzed in order to represent better the concentration ranges met in Finland. Test of comparability (see in Chapter 2.4 and 4.2) indicates that the correction of the data with the calibration equation needs to apply for all CMs.

### 5.3. Results of the $\text{PM}_{2.5}$ comparisons

The comparison tests for the CMs of  $\text{PM}_{2.5}$  measurements continued with summer and autumn comparisons at Tasavallankatu. At Savilahdentie the comparison started at the beginning of December. As mentioned earlier the measurement strategy changed at Savilahdentie with respect to the use of the reference method. One of the reference samplers was sampling  $\text{PM}_{10}$  and the other  $\text{PM}_{2.5}$  continuously. There was a short period for parallel measurement of  $\text{PM}_{10}$  at the beginning of the campaigns at Savilahdentie to demonstrate that the samplers are equal.






Figure 5.8a: The time series of the two RMs for  $\text{PM}_{2.5}$  comparisons from Tasavallankatu. UAT is the annual upper assessment threshold value for  $\text{PM}_{2.5}$ .



*Figure 5.8b. The time series of the two RMs for PM<sub>2.5</sub> comparisons from Savilahdentie. UAT is the annual upper assessment threshold value for PM<sub>2.5</sub>.*

In Figure 5.8 the time series of daily concentrations of RMs for PM<sub>2.5</sub> for campaigns at Tasavallankatu (5.8a) and at Savilahdentie (5.8b) are presented. In the figure the annual Upper Assessment Threshold value is 17 µg/m<sup>3</sup> showing that all the concentration values for PM<sub>2.5</sub> are less than that. At Savilahdentie the concentrations are even lower than at Tasavallankatu while for PM<sub>10</sub> the situation was opposite.

The scatterplot of RM2 against the RM1 is presented in Figure 5.9 from the campaigns in Tasavallankatu. The standard uncertainty for the between sampler test for the reference method, according to eq. (2.1), is 0.6 µg/m<sup>3</sup> for all data.



*Figure 5.9: The scatter plot of the RM2 against the RM1 for PM<sub>2.5</sub> from Tasavallankatu.*

### 5.3.3. Results of the candidate methods for PM<sub>2.5</sub>

In the following tables all the results for the analysis of the CMs against the RM are presented. The analysis is the same as in PM<sub>10</sub> including the data capture, suitability of data, between CM tests and comparability test. The comparability test includes the tests of significance for slope and intercept indicating if the calibration function for CM is required. The comparability tests are performed for each of the campaigns. The calibration equation according to Eq. (2.8a) is expressed in any case i.e. irrespectively if the slope  $b$  and/or  $a$  are significantly different from 1 and 0, respectively. The analysis of data for larger than 18 µg/m<sup>3</sup> was not performed as required by the GDE due to the fact there were no many days fulfilling this requirement as seen in the figure 5.8. The results that exceeds the acceptable limit, see in chapter 2.4, are printed in red color.

Table 5.10: The test results of BAM for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.

Test PM <sub>2.5</sub>	Criteria	BAM A + B	BAM A	BAM B	Pass/Fail
Data capture	≥ 90 % of RM	95,1 %	84,4 %	95,1 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	Fail	Fail	Fail	Fail
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	0,64			Pass
> 18 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>				
Test of comparability					
All data					
Number of data	Significant	195	173	195	
Slope		0,91	0,95	0,98	
Intercept	Significant	-0,7	-0,8	-1,9	
<b>Calibrated data</b>					
<i>All data</i>					
Number of data		195	173	195	
Calibration		1,100y + 0,733	1,058y + 0,883	1,016y + 1,934	
Expanded relative uncertainty	25 %	7,4 %	7,3 %	6,8 %	Pass
<i>Tasavallankatu: Tasa 1</i>					
Number of data		59	52	59	
Calibration		1,124y + 1,31	1,120y + 0,868	1,112y + 1,762	
Expanded relative uncertainty	25 %	0,3 %	3,3 %	1,8 %	Pass
<i>Tasavallankatu: Tasa 2</i>					
Number of data		56	50	56	
Calibration		0,919y + 1,521	0,930y + 1,176	0,915y + 1,813	
Expanded relative uncertainty	25 %	7,9 %	13,2 %	8,2 %	Pass
<i>Savilahdentie: Savi 1</i>					
Number of data		40	40	40	
Calibration		1,320y - 0,597	1,339y + 0,051	1,291y + 1,41	
Expanded relative uncertainty	25 %	21,4 %	23,0 %	20,6 %	Pass
<i>Savilahdentie: Savi 2</i>					
Number of data		40	31	40	
Calibration		0,823y + 1,672	0,929y + 1,374	0,820y + 2,74	
Expanded relative uncertainty	25 %	22,2 %	14,3 %	21,6 %	Pass

*Table 5.11. The test results of DustTrak for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.*

Test PM <sub>2.5</sub>	Criteria	DustTrak A + B	DustTrak A	DustTrak B	Pass/Fail
Data capture	≥ 90 % of RM	98,2 %	98,2 %	97,6 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	Fail	Fail	Fail	Fail
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	2,88			Fail
> 18 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>				
Test of comparability					
All data					
Number of data		164	164	163	
Slope	Significant	1,66	2,03	1,56	
Intercept	Significant	1,7	-2,4	4,2	
<b>Calibrated data</b>					
<i>All data</i>					
Number of data		164	164	163	
Calibration		0,602y -1,002	0,492y + 1,19	0,641y -2,685	
Expanded relative uncertainty	25 %	37,9 %	48,8 %	35,0 %	Fail
<i>Savilahdentie: Savi 1-1</i>					
Number of data		43	43	43	
Calibration		0,340y + 0,469	0,330y + 0,758	0,348y + 0,196	
Expanded relative uncertainty	25 %	36,3 %	38,1 %	35,4 %	Fail
<i>Savilahdentie: Savi 2-2</i>					
Number of data		40	40	40	
Calibration		0,957y -6,944	0,923y -4,444	0,991y -9,588	
Expanded relative uncertainty	25 %	53,0 %	53,0 %	53,1 %	Fail
<i>Savilahdentie: Savi 1-3</i>					
Number of data		40	40	40	
Calibration		0,551y + 0,491	0,547y + 1,726	0,535y -0,520	
Expanded relative uncertainty	25 %	48,3 %	50,0 %	50,8 %	Fail
<i>Savilahdentie: Savi 2-4</i>					
Number of data		41	41	40	
Calibration		0,303y + 2,708	0,289y + 3,577	0,320y + 1,678	
Expanded relative uncertainty	25 %	61,7 %	64,1 %	58,6 %	Fail

Table 5.12. The test results of FH 62 I-R for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.

Test PM <sub>2.5</sub>	Criteria	FH 62 A + B	FH 62 A	FH 62 B	Pass/Fail
Data capture	≥ 90 % of RM	95,1 %	92,2 %	95,1 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>
Between sampler test					
All data > 18 µg/m <sup>3</sup>	u <sub>bs</sub> < 2.5 µg/m <sup>3</sup> u <sub>bs</sub> < 2.5 µg/m <sup>3</sup>	1,07			Pass
Test of comparability					
All data					
Number of data	Significant	195	189	195	
Slope		1,18	1,38	1,06	
Intercept	Significant	-2,0	-3,1	-1,6	
Calibrated data					
All data					
Number of data		195	189	195	
Calibration		0,850y + 1,709	0,722y + 2,227	0,940y + 1,488	
Expanded relative uncertainty	25 %	17,3 %	24,6 %	15,0 %	Pass
Tasavallankatu: Tasa 1					
Number of data		57	54	59	
Calibration		0,919y + 1,801	0,737y + 2,811	1,041y + 1,359	
Expanded relative uncertainty	25 %	16,7 %	23,9 %	15,2 %	Pass
Tasavallankatu: Tasa 2					
Number of data		56	54	56	
Calibration		0,881y + 1,796	0,790y + 2,229	0,929y + 1,49	
Expanded relative uncertainty	25 %	16,4 %	21,0 %	11,9 %	Pass
Savilahdentie: Savi 1					
Number of data		41	41	40	
Calibration		0,814y + 1,503	0,647y + 1,902	0,988y + 1,125	
Expanded relative uncertainty	25 %	42,5 %	58,4 %	34,6 %	Fail
Savilahdentie: Savi 2					
Number of data		41	40	40	
Calibration		0,415y + 3,173	0,401y + 3,2	0,487y + 3,121	
Expanded relative uncertainty	25 %	78,9 %	81,1 %	70,6 %	Fail

*Table 5.13. The test results of Grimm for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.*

Test PM <sub>2.5</sub>	Criteria	Grimm B + C	Grimm B	Grimm C	Pass/Fail
Data capture	≥ 90 % of RM	96,1 %	94,4 %	93,9 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	0,60			Pass
> 18 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>				
Test of comparability					
All data					
Number of data	Significant	173	170	169	
Slope	Significant	1,34	1,39	1,36	
Intercept	Significant	-0,7	-1,0	-0,7	
<b>Calibrated data</b>					
<i>All data</i>					
Number of data		173	170	169	
Calibration		0,747y + 0,532	0,718y + 0,735	0,737y + 0,536	
Expanded relative uncertainty	25 %	12,6 %	11,9 %	13,8 %	Pass
Tasavallankatu: Tasa 1					
Number of data		50	47	46	
Calibration		1,019y -0,854	0,983y -0,926	1,087y -0,985	
Expanded relative uncertainty	25 %	10,4 %	11,5 %	10,2 %	Pass
Savilahdentie: Savi 1-1					
Number of data		43	43	43	
Calibration		0,730y + 0,25	0,703y + 0,549	0,747y + 0,023	
Expanded relative uncertainty	25 %	11,5 %	10,0 %	14,9 %	Pass
Savilahdentie: Savi 2-2					
Number of data		40	40	40	
Calibration		0,745y -0,159	0,717y + 0,122	0,739y -0,345	
Expanded relative uncertainty	25 %	24,2 %	23,4 %	24,4 %	Pass
Savilahdentie: Savi 1-3					
Number of data		40	40	40	
Calibration		0,718y + 0,571	0,735y + 0,652	0,701y + 0,497	
Expanded relative uncertainty	25 %	16,3 %	16,6 %	16,1 %	Pass

Table 5.14. The test results of MP101 for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.

Test PM <sub>2.5</sub>	Criteria	MP101 A + B	MP101 A	MP101 B	Pass/Fail
Data capture	≥ 90 % of RM	95,6 %	93,7 %	95,1 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	1,25			Pass
> 18 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>				
Test of comparability					
All data					
Number of data	Significant	196	192	195	
Slope	Significant	1,23	1,12	1,31	
Intercept	Significant	0,4	0,6	0,3	
<b>Calibrated data</b>					
<i>All data</i>					
Number of data		196	192	195	
Calibration		0,812y -0,306	0,894y -0,517	0,763y -0,266	
Expanded relative uncertainty	25 %	8,9 %	9,7 %	9,4 %	Pass
<i>Tasavallankatu: Tasa 1</i>					
Number of data		57	56	57	
Calibration		0,876y -0,901	0,950y -0,623	0,824y -1,196	
Expanded relative uncertainty	25 %	10,8 %	12,4 %	14,6 %	Pass
<i>Tasavallankatu: Tasa 2</i>					
Number of data		57	56	57	
Calibration		0,873y -0,924	0,867y -0,728	0,836y -0,802	
Expanded relative uncertainty	25 %	15,6 %	18,4 %	16,7 %	Pass
<i>Savilahdentie: Savi 1</i>					
Number of data		40	40	40	
Calibration		0,701y + 0,251	0,847y -0,260	0,735y -0,123	
Expanded relative uncertainty	25 %	22,6 %	18,5 %	18,6 %	Pass
<i>Savilahdentie: Savi 2</i>					
Number of data		42	40	41	
Calibration		0,762y + 0,291	0,856y -0,142	0,672y + 0,624	
Expanded relative uncertainty	25 %	22,0 %	21,5 %	21,5 %	Pass

Table 5.15: The test results of Osiris for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.

Test PM <sub>2.5</sub>	Criteria	Osiris A + B	Osiris A	Osiris B	Pass/Fail
Data capture	≥ 90 % of RM	95,1 %	94,6 %	95,1 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	0,31			Pass
> 18 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>				
Test of comparability					
All data					
Number of data	Significant	195	194	195	
Slope		0,30	0,31	0,30	
Intercept	Significant	0,3	0,1	0,5	
Calibrated data					
All data					
Number of data		195	194	195	
Calibration		3,324y -1,073	3,278y -0,429	3,331y -1,662	
Expanded relative uncertainty	25 %	124,2 %	115,0 %	134,9 %	Fail
Tasavallankatu: Tasa 1					
Number of data		61	61	61	
Calibration		5,498y -5,562	4,854y -3,187	6,318y -8,607	
Expanded relative uncertainty	25 %	187,0 %	132,2 %	296,0 %	Fail
Tasavallankatu: Tasa 2					
Number of data		51	51	51	
Calibration		3,641y -4,824	3,654y -4,450	3,622y -5,172	
Expanded relative uncertainty	25 %	112,7 %	101,0 %	127,3 %	Fail
Savilahdentie: Savi 1					
Number of data		41	40	41	
Calibration		1,881y + 1,36	1,928y + 1,467	1,738y + 1,357	
Expanded relative uncertainty	25 %	24,7 %	20,3 %	22,6 %	Pass
Savilahdentie: Savi 2					
Number of data		42	42	42	
Calibration		6,505y -2,357	8,562y -2,529	5,178y -2,155	
Expanded relative uncertainty	25 %	411,9 %	519,6 %	343,6 %	Fail

Table 5.16. The test results of SHARP beta for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.

Test PM <sub>2.5</sub>	Criteria	Sharp beta A + B	Sharp beta A	Sharp beta B	Pass/Fail
Data capture	≥ 90 % of RM	95,6 %	96,1 %	93,2 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>
Between sampler test					
All data	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	<b>0,58</b>			<b>Pass</b>
> 18 µg/m <sup>3</sup>	u <sub>bs</sub> < 2,5 µg/m <sup>3</sup>				
Test of comparability					
All data					
Number of data		196	197	191	
Slope		1,03	1,04	1,04	
Intercept		0,0	0,0	-0,1	
<b>Calibrated data</b>					
<i>All data</i>					
Number of data		196	197	191	
Calibration		0,971y -0,003	0,966y + 0,002	0,959y + 0,127	
Expanded relative uncertainty	25 %	0,2 %	0,2 %	3,9 %	Pass
<i>Tasavallankatu: Tasa 1</i>					
Number of data		57	58	54	
Calibration		0,970y + 0,047	0,957y + 0,181	0,942y + 0,199	
Expanded relative uncertainty	25 %	2,3 %	2,7 %	7,4 %	Pass
<i>Tasavallankatu: Tasa 2</i>					
Number of data		57	57	57	
Calibration		0,975y -0,103	0,947y -0,097	0,988y + 0,008	
Expanded relative uncertainty	25 %	4,3 %	6,3 %	5,8 %	Pass
<i>Savilahdentie: Savi 1</i>					
Number of data		40	40	40	
Calibration		1,132y -0,564	1,165y -0,688	1,073y -0,304	
Expanded relative uncertainty	25 %	11,0 %	9,7 %	14,5 %	Pass
<i>Savilahdentie: Savi 2</i>					
Number of data		42	42	40	
Calibration		0,828y + 0,635	0,877y + 0,398	0,848y + 0,593	
Expanded relative uncertainty	25 %	13,4 %	9,6 %	16,9 %	Pass

*Table 5.17. The test results of SHARP c-dust for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.*

Test PM <sub>2.5</sub>	Criteria	Sharp c-dust A + B	Sharp c-dust A	Sharp c-dust B	Pass/Fail
Data capture	≥ 90 % of RM	95,1 %	94,6 %	92,7 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>
Between sampler test					
All data	U <sub>bs</sub> < 2.5 µg/m <sup>3</sup>	1,68			Pass
> 18 µg/m <sup>3</sup>	U <sub>bs</sub> < 2.5 µg/m <sup>3</sup>				
Test of comparability					
All data					
Number of data		195	194	190	
Slope	Significant	1,17	1,38	1,11	
Intercept	Significant	-1,4	-3,2	-0,6	
Calibrated data					
All data					
Number of data		195	194	190	
Calibration		0,854y + 1,187	0,726y + 2,328	0,900y + 0,559	
Expanded relative uncertainty	25 %	7,3 %	16,2 %	5,9 %	Pass
Tasavallankatu: Tasa 1					
Number of data		58	58	53	
Calibration		0,978y + 0,127	0,927y + 0,726	0,986y -0,085	
Expanded relative uncertainty	25 %	7,2 %	12,2 %	5,7 %	Pass
Tasavallankatu: Tasa 2					
Number of data		57	53	57	
Calibration		0,908y + 0,397	0,855y + 0,62	0,943y + 0,317	
Expanded relative uncertainty	25 %	9,9 %	13,8 %	9,2 %	Pass
Savilahdentie: Savi 1					
Number of data		40	41	40	
Calibration		0,953y + 0,709	1,046y + 0,498	0,872y + 0,984	
Expanded relative uncertainty	25 %	14,5 %	17,2 %	18,7 %	Pass
Savilahdentie: Savi 2					
Number of data		40	42	40	
Calibration		0,644y + 2,763	0,469y + 4,314	0,681y + 1,394	
Expanded relative uncertainty	25 %	27,7 %	65,8 %	18,7 %	<b>Fail</b>

*Table 5.18. The test results of TEOM for all the tests for demonstration of equivalence for PM<sub>2.5</sub> measurements.*

Test PM <sub>2.5</sub>	Criteria	TEOM A + B	TEOM A	TEOM B	Pass/Fail
Data capture	≥ 90 % of RM	95,6 %	95,1 %	95,6 %	Pass
Suitability of data	20 % > 17 µg/m <sup>3</sup>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>	<b>Fail</b>
Between sampler test					
All data	U <sub>bs</sub> < 2,5 µg/m <sup>3</sup>	0,42			Pass
> 18 µg/m <sup>3</sup>	U <sub>bs</sub> < 2,5 µg/m <sup>3</sup>				
Test of comparability					
All data					
Number of data		196	195	196	
Slope		0,99	1,03	0,96	
Intercept	Significant	1,7	1,5	1,8	
Calibrated data					
All data					
Number of data		196	195	196	
Calibration		1,009y -1,681	0,975y -1,487	1,045y -1,865	
Expanded relative uncertainty	25 %	8,8 %	9,7 %	8,2 %	Pass
Tasavallankatu: Tasa 1					
Number of data		58	58	58	
Calibration		1,276y -3,881	1,268y -3,930	1,280y -3,806	
Expanded relative uncertainty	25 %	10,0 %	10,9 %	9,2 %	Pass
Tasavallankatu: Tasa 2					
Number of data		57	57	57	
Calibration		0,974y -2,365	0,926y -2,154	1,019y -2,519	
Expanded relative uncertainty	25 %	8,4 %	9,9 %	7,8 %	Pass
Savilahdentie: Savi 1					
Number of data		40	40	40	
Calibration		1,223y -2,986	1,243y -3,054	1,203y -2,911	
Expanded relative uncertainty	25 %	14,6 %	14,9 %	14,4 %	Pass
Savilahdentie: Savi 2					
Number of data		41	40	41	
Calibration		0,865y + 0,142	0,905y -0,042	0,913y -0,104	
Expanded relative uncertainty	25 %	20,8 %	21,0 %	19,7 %	Pass

### 5.3.4 Summary of the PM<sub>2.5</sub> comparisons

The number of sites and the number of campaigns for the PM<sub>2.5</sub>-comparison cover the requirements of the GDE. The meteorological conditions over the campaigns periods represent the annual variation met in Southern Finland. The concentration ranges at both sites in Kuopio were, however, quite low. Regarding to the suitability criterion, i.e. 20 % of the concentration of PM<sub>2.5</sub> during the whole measurement period should be higher than 17 µg/m<sup>3</sup>, was not fulfilled. Figure 5.10 presents the monthly mean PM<sub>2.5</sub> concentrations

from different parts in Finland together with the values obtained from this comparison. Considering the mass concentration of PM<sub>2.5</sub> in different locations in Finland, it is very unlikely to fulfill the suitability criterion in any part of Finland. At this PM<sub>2.5</sub> comparison tests, the suitability criterion was not considered as discarding factor for CMs.




Figure 5.10: The monthly mean concentration of PM<sub>2.5</sub> in 2014-15 in Helsinki (blue dot), in Oulu (orange dot), in Vaasa (grey dot), a background station in Virolahti (brown dot), and in the PM<sub>2.5</sub> equivalence tests (red square).

Figure 5.10 shows that the concentration ranges from the comparison covers reasonably well the concentration ranges of PM<sub>2.5</sub> at different sites across Finland in spite of lack of data during the months in winter. The concentration range at the site of Helsinki-Mannerheim is however, higher compared with the concentrations in other stations.

Regarding the results for demonstration of equivalence, Tables 5.10 to 5.18, the data capture is fulfilled with the CMs. One of the BAM instrument had a technical problem which was not able to repair in due time causing a data capture of less than required. Suitability requirement for the concentration range measured during the comparison period, is not fulfilled by any of the CMs. The between test for CMs, is performed only to

all data but not to data with the concentration above 18  $\mu\text{g}/\text{m}^3$  since there were no concentrations above this limit. The DustTrak failed the requirement for the between CM test. Test of comparability (see in Chapter 2.4 and 4.2) indicates that the correction of the data with the calibration equation needs to apply for all CMs except for  $\beta$ -signal of SHARP and slope of TEOM. Also with regard of the results of SHARP both of the signals are analyzed showing better performances for the  $\beta$ -signal than for the c-dust signal, which is the signal for normal use of the CM. The results of BAM failed with the number of measurements at one campaign, Savi 2, due to the jam of the filter roll. Due to the bad timing of the malfunction, the repair was not possible to organize as quickly as desirable.

#### **5.4. Result of the field audit**

The audit surveillance visit was conducted on June 4<sup>th</sup>, 2015 at the Savilandentie. The audit report is in Finnish, but the findings and the response is reported in Annex 28. Non-compliances were observed in some of the QA/QC procedures according to table 3.1, which were corrected or an influence could not be demonstrated in the results.

#### **5.5. Calibration equations for CMs**

The calibration equations from the comparability tests, tables 5.1 – 5.9, are collected for all data in the tables 5.19 and 5.20 for  $\text{PM}_{10}$  and for  $\text{PM}_{2.5}$  candidate methods.

Table 5.19. The calibration range and the equations (see Eq. (2.2),  $b$ =slope,  $a$ =intercept and  $y$  = is the CM) against the reference method for  $PM_{10}$  together with the relative combined standard uncertainty.

PM10 Candidate method	< 325 $\mu\text{g}/\text{m}^3$		< 100 $\mu\text{g}/\text{m}^3$		< 100 $\mu\text{g}/\text{m}^3$	
	Calibration equation $PM_{10}$	Relative expanded uncertainty	Calibration equation $PM_{10}$	Relative expanded uncertainty	Calibration equation through origin $PM_{10}$	Relative expanded uncertainty
BAM 1020	$0,942y + 0,437$	12,6%	$0,858y + 1,919$	10,3%	$0,913y$	11,7%
GRIMM 180	$0,855y + 2,139$	17,0 %	$0,871y + 1,927$	17,0 %	$0,922y$	17,9 %
SHARP 5030 C-dust	$1,404y - 2,750$	17,2%	$1,486y - 3,904$	16,5%	$1,319y$	16,3%
SHARP 5030 (beta)	$1,415y - 2,233$	12,8%	$1,489y - 3,301$	12,5%	$1,351y$	12,5%
FH 62 IR	$1,300y - 0,904$	16,5%	$1,372y - 1,850$	17,1%	$1,297y$	12,6%
TEOM 1405	$0,868y - 2,068$	14,4%	$0,804y - 0,623$	13,6%	$0,788y$	13,0%
MP101M	$0,811y + 2,311$	11,0%	$0,887y + 0,826$	9,4%	$0,910y$	9,6%
OSIRIS	$1,401y - 0,153$	15,7%	$1,338y + 0,57$	15,3%	$1,363y$	15,7%
DustTrak	$7,478y - 76,819$	402,3%	$5,761y - 55,073$	1132,0%	$2,07y$	94,0%

Table 5.20: The calibration range and the equations (see Eq. (2.1),  $b$ =slope,  $a$ =intercept and  $y$  = the CM) against the reference method for  $PM_{2.5}$  together with the relative combined standard uncertainty.

PM2.5 Candidate method	< 25 $\mu\text{g}/\text{m}^3$		< 25 $\mu\text{g}/\text{m}^3$	
	Calibration equation $PM_{10}$	Relative expanded uncertainty	Calibration equation through the origin $PM_{2.5}$	Relative expanded uncertainty
BAM 1020	$1,100y + 0,733$	7,4%	$1,215y$	19,9%
GRIMM 180 (*)	$0,747y + 0,532$	12,6 %	$0,780y$	12,3 %
SHARP 5030 C-dust	$0,854y + 1,187$	7,3%	$1,009y$	27,7%
SHARP 5030 (beta)	$0,971y - 0,003$	0,2%	$0,971y$	0,2%
FH 62 IR	$0,850y + 1,709$	17,3%	$1,097y$	51,8%
TEOM 1405	$1,009y - 1,681$	8,8%	$0,821y$	31,4%
MP101M	$0,812y - 0,306$	8,9%	$0,780y$	31,4%
OSIRIS (*)	$3,324y - 1,073$	124,2%	$2,020y$	76,1%
Dusttrak(*)	$0,602y - 1,002$	37,9%	$0,550y$	143,9%
(* Range < 75 $\mu\text{g}/\text{m}^3$ )				

The calibration equations in table 5.19 and table 5.20 are from the whole comparison data for both of the PM size classes. The concentration range met in  $PM_{10}$  comparison tests was from 0 to 325  $\mu\text{g}/\text{m}^3$ , while the more practical concentration range in Finland is from 0 to 100  $\mu\text{g}/\text{m}^3$ . In case of  $PM_{2.5}$  the concentrations were relatively low and the estimated range for the applicability of the calibration equations in table 20 is from 0 to 25  $\mu\text{g}/\text{m}^3$  except for Grimm 180, Osiris and Dustrak

where the range is from 0 to 75  $\mu\text{g}/\text{m}^3$ . At some of the CMs the calibration equation can be replaced with the slope (the right column at both of the tables). However, the calibration slope through the origin cannot be used with Dusttrak for PM<sub>10</sub> measurements, or with SHARP (C-dust signal), FH 62 IR, TEOM 1405, MP101, Osiris and Dusttrak for PM<sub>2.5</sub> measurements.

The correction based on the tables 5.19 and 5.20, or by using the tables 5.1 to 5.18 is made according to:

$$y_{i,cal} = b \cdot y_i + a \quad (5.1)$$

Where

$y_{i,cal}$  = corrected values

$b$  = slope of the calibration equation

$y_i$  = measurement value (raw) of the CM

$a$  = intercept of the calibration equation

The additional factors installed at the CM were removed when possible but in any case the factors if existing were the same for the same model of CM. The EU-factor in FH 62IR and SHARP were set to EU-factor = 1.00. The factory setting of this factor has been 1.3 by default. The TEOM 1405 corrects the signal by a correction factor of  $TEOM = 1.03 \cdot y + 3 \mu\text{g}/\text{m}^3$  where  $y$  is the raw signal of the TEOM. During the campaigns, the correction factor was not changed i.e. all results gained for TEOM includes the factory correction as default. At any other CMs there were no additional correction factors. The operator needs to be aware of the possible factors and their values in using the calibration equations to correct the results to be equivalent with the reference values.

## **5.4 Results of the automated PM analyzer at the Tasavallankatu**

The final result of the comparison was to test the applicability of the gained calibration factor to the station analyzer, which was MP 101 for PM<sub>10</sub>. The results of the comparability tests shows that the raw data needs to be corrected (see table 5.21 third column, MPM 101 raw). The correction is made according to table 5.19 calibration equation where the corrected results are presented in the 4<sup>th</sup> column at table 5.21. The corrected results give slightly

overestimated results than the reference method with lower uncertainty than by using the raw data.

*Table 5.21: The test results for the station analyzer, MP 101, at Tasavallankatu for PM<sub>10</sub> measurement.*

Test PM <sub>10</sub>	Criteria	MP101 raw	MP101 correct	Pass/Fail
Data capture	≥ 90 % of RM	98,0 %	98,0 %	Pass
Suitability of data	20 % > 28 µg/m <sup>3</sup>	Pass	Pass	Pass
Test of comparability All data				
Number of data		99	99	
Slope	Significant	1,08	0,95	
Intercept	Significant	-0,8	0,2	
<b>Calibrated data</b>				
<i>All data</i>				
Number of data		99	99	
Calibration		1,080y - 0,84	0,954y + 0,16	
Expanded relative uncertainty	25 %	17,7 %	13,7 %	Pass
<i>Tasavallankatu: Tasa 1</i>				
Number of data		48	48	
Calibration		1,1296y - 1,403	0,9969y - 0,30	
Expanded relative uncertainty	25 %	24,3 %	11,8 %	Pass
<i>Tasavallankatu: Tasa 2</i>				
Number of data		51	51	
Calibration		0,9889y + 0,046	0,874y + 0,92	
Expanded relative uncertainty	25 %	10,1 %	23,2 %	Pass
> 30 µg/m <sup>3</sup>				
Number of data		22	22	
Calibration		1,147y - 3,89	0,9264y - 0,89	
Expanded relative uncertainty	25 %	20,0 %	18,3 %	Pass

## 6. Conclusions

The Air Quality Directive 2008/50/EC (AQD) defines the reference measurement methods for the atmospheric pollutants where limit or target values have been defined. Other measurement methods than the reference methods can be used for e.g. particulate matter if the Member State can demonstrate that the method displays a consistent relationship with the reference method. The comparison tests to study the equivalency between the automated PM analyzers and the reference method for measurements of mass concentration of PM<sub>2.5</sub> and PM<sub>10</sub> were performed in accordance with the guidance document (GDE 2010). The automated PM analyzers of BAM 1020, DusTrak 8535, FH 62 I-R, Grimm model 180, MP101 CPM, Osiris, SHARP model 5030 and TEOM 1405 took part in the comparison to test the equivalency against the PM reference method both for the PM<sub>10</sub> and PM<sub>2.5</sub> measurements in ambient air. The number of sites, campaigns, and the number of samples/campaign fulfilled the criteria of the GDE. The suitability criteria for the proportion to exceed the UAT values for more than 20 % were satisfied with the PM<sub>10</sub> measurements, but not for PM<sub>2.5</sub> measurements. Consequently, the concentration range for applying the results for the CMs in PM<sub>2.5</sub> measurements was defined.

Care shall be taken when comparing the results of this study with others in case of the environmental conditions, but also with the concentration levels, origin and composition of particles. The composition of particles can be formed from a fraction of semi volatile compounds that can be quite different from various parts in Finland but more likely different across the Europe. The fraction of semi-volatile compounds was not measured and not taken into account in the calibration equations due to the lack of proper instrumentation. Estimation of the semi-volatile compounds have been performed in Finland by different techniques (Timonen et al. 2012, Ruoho-Airola 2014) indicating proportion of semi-volatile compounds between 10 to 18 % of concentration in PM<sub>2.5</sub>. At Tasavallankatu in Kuopio the source area of PM<sub>2.5</sub> is evenly distributed into the wind sectors showing the various sources of particles while in Savilahdentie the source area is clearly from the traffic side. In case of PM<sub>10</sub> the source area is even more concentrated on

the traffic side at both sites. The content of the particles from the traffic side are mainly silica-based particles and combustion based particles. Considering that the source area, especially for PM<sub>10</sub> was more concentrated in the sector where solid particles gives the major contribution on mass, this may have more influence on the optical methods than in  $\beta$  attenuation or micro balance method. This can be the origin for the variation of the calibration factors from different campaigns, especially with optical methods. Also the calibration equations differs more for the winter campaign, indicating some problems in sampling due to the collection of snow on the sampling inlets. The configuration, software (versions of firmware), materials and equipment can vary in the CMs, raising a question whether the results are applicable to be used elsewhere. There were no modifications on the operation of the CMs except the sampling time for BAM 1020, which was 3 hour instead of 1 hour. The maintenance of the CMs were performed according to the manual and by Finnish representatives of the manufacturer. The QA/QC procedures were performed according to the GDE and the manufacturer instructions.

The results showed that only one CM of the tested analyzers, DustTrak 8535, failed to meet the criteria for equivalence for fixed measurements of PM<sub>2.5</sub> and PM<sub>10</sub>. Osiris fulfilled the criteria for PM<sub>10</sub> measurements, but failed to meet the criteria for PM<sub>2.5</sub> measurements. All the other CMs, FH 62 I-R, Grimm model 180, MP101 CPM, SHARP model 5030, TEOM 1405 and BAM 1020 fulfilled the criteria for the reference method for PM<sub>2.5</sub> and for PM<sub>10</sub> measurements in ambient air. The calibration equations calculated for the CMs can be used for correcting the measured concentration values. It is, however, the task of the responsible authority in the member state to approve the tested CMs with the calibration equation as an equivalent method to be used for the measurement of mass concentration of PM<sub>2.5</sub> and PM<sub>10</sub> in ambient air according to the air quality directive.

## Acknowledgements

The project was funded by the Ministry of Environment, Ministry of Traffic and Communication and the Finnish Meteorological Institute. The authors like to express their special thanks to Mr Erkki Pärjälä and Mikko Sokura from the City of Kuopio for excellent cooperation and help during the whole course of the comparison experiments. The Manufacturers, Finnish representatives, JPP-kalibrointi KY, Estonian Environmental Research Centre, the Expert Service Group and the Observation Group at the Finnish Meteorological Institute are thanked for their help and providing the candidate methods for testing. Ministerial Adviser, Tarja Lahtinen from the Ministry of Environment is thanked for keen interest and valuable comments during the project and Dr Virpi Tarvainen for revising the language.

## 7. References

Beijk R, Hoogerbrugge R, Hafkenscheid TL, Arkel FT van, Stefess GC, Meulen A van der, Wesseling JP, Sauter FJ, Albers RAW, RIVM rapport 680708001, ([www.rivm.nl](http://www.rivm.nl))

Bertrand, L., Gérard G., and Schroyen G., (2009). PM<sub>10</sub> and PM<sub>2,5</sub> Equivalence test Lodelinsart, Progress report (winter 2008-2009) Rapport no 09-960, Institut scientifique de service public, (<http://airquality.issep.be>).

De Jonge D. (2008). Field experiment on 11 automated PM monitors. Technical report. Department of Air Quality Research, Municipal Health Service Amsterdam, (GGD Amsterdam).

AQD 2008/50/EC: Directive of the European Parliament and of the Council of 21 May 2008 on ambient air quality and cleaner air for Europe

EC WG on PM (2001). Report on guidance to member states on PM<sub>10</sub> monitoring and intercomparisons with the reference method.

Matos, J. et al. 2008. Determinação do factor equivalência entre método referência e automático PM10 Estação de Referência do Ambiente Alfragide 2008 Amadora 2008 ([http://www.apambiente.pt/servicos/LaboratorioReferencia/EstacaoReferenciaQualidadeAr/Documents/Factor equivalencia metodo referencia e automatico PM10 APA 2008.pdf](http://www.apambiente.pt/servicos/LaboratorioReferencia/EstacaoReferenciaQualidadeAr/Documents/Factor%20equivalencia%20metodo%20referencia%20e%20automatico%20PM10%20APA%202008.pdf))

EN12341 Air Quality – Determination of the PM10 fraction of suspended particulate matter – reference method and field test procedure to demonstrate reference equivalence of measurement methods. November 1998.

EN12341: 2014. Air Quality – Standard gravimetric measurement method for the determination of the PM<sub>10</sub> or PM<sub>2,5</sub> mass concentration of suspended particulate matter.

EN14907 Ambient air quality Standard gravimetric measurement method for the determination of the PM<sub>2.5</sub> mass fraction of suspended particulate matter. September 2005.

EN 15267-1 Air quality. Certification of automated measuring systems. Part 1: General principles

EN 15267-2 Air quality. Certification of automated measuring systems. Part 2: Initial assessment of the AMS manufacturer's quality management system and post certification surveillance for the manufacturing process.

EnviDas for Windows, 2003. Environmental Data Acquisition System, Envitech Ltd.

Envview 2000 Software manager 2004. Environmental data Management Systems, Envitech Ltd.

Demonstration of Equivalence of Ambient Air Monitoring Methods; 2005, EC-working group on guidance for the demonstration on equivalence) ([http://ec.europa.eu/environment/air/pdf/equivalence\\_report.pdf](http://ec.europa.eu/environment/air/pdf/equivalence_report.pdf))

Harrison et al. (2006). UK Equivalence Programme for Monitoring of Particulate Matter, Final report BV/AQ/AD202209/DH/2396, June 5, 2006.

Järvi, L., Hannuniemi, H., Hussein, T., Junninen, H., Aalto, P., Hillamo, R., Mäkelä, T., Keronen, P., Siivola, E., Vesala, T. and Kulmala, M.: The urban measurement station SMEAR III: Continuous monitoring of air pollution and surface-atmosphere interactions in Helsinki, Finland. *Boreal. Environ. Research.*, 14 suppl. A, 86-109, 2009

Kaminski H. and Kuhlbusch T., 2010. Überprüfung der Trenngrade zweier PM<sub>10</sub> – Vorabscheider für 2,3 m<sup>3</sup>/h mittels Messung mit polydispersem Aerosol. Bericht Nr LP. 59/2010. Institute für Energie- und Umwelttechnik e. V. (IUTA).

LUBW 2 21-18, 2005. Prüfung des schwestaubmessgerätes environmental dust monitor – Model 180 der firma Grimm Aerosoltechnik auf Gleichwertige ergebnisse wie mit der referenzmethode der Europäischen norm Din EN 12341 bei der messung von schwebstaub der fraktion PM<sub>10</sub>.

Pirinen, P., Simola, H., Aalto, J., Kaukoranta, J-P., Karlsson, P., Ruuhela, R., 2012. Tilastoja Suomen ilmastosta 1981–2010. 83 s.

Walden, J., Hillamo, R., Aurela, M., Makela, T., Laurila, S., 2010. Demonstration of the equivalence of PM<sub>2.5</sub> and PM<sub>10</sub> measurement methods in Helsinki 2007-2008. 103 s. Finnish Meteorological Institute, Studies 3, Helsinki.

WHO (2006) *Health risk of particulate matter from long-range transboundary air pollution*. Joint WHO/Convention Task Force on the health aspects of air pollution. Denmark: World Health Organization Europe, Publication E88189.

## Annexes

#### **A1: QA/QC procedures conducted for the RM and CM during the comparison experiments.**

*Table A1a. The flow rates of the RMs and CMs equipped with the size selective inlets. Measurements are expressed as [l/min] at prevailing ambient air conditions.*

Date	Leckel RM1	Leckel RM2	Sharp A	Sharp B	TEOM B	TEOM A	MP-101 A	MP-101 B	FH62 B	FH62 A	BAM A	BAM B
16.4.2014	38,3	38,2		19,3	17,4	18,1	15,3	17,5	18,1	18,4	18,1	17,9
24.6.2014	39,2	39,9		18,4	17,2	17,3	14,9	16,9	17,4	17,5	17,2	17,1
9.7.2014												
15.7.2014	37,9	38,1		17,1	16,9	16,5			16,1	16,4	16,2	16,1
29.7.2014	38,0	37,6	15,9	16,7	16,8	16,4	16,7	15,8	16,0	15,7	15,8	16,0
9.9.2014			16,3	17,4	16,9	16,9	17,1	16,3	16,5	16,8	16,8	16,8
7.10.2014	38,5	38,5	16,5	16,5	15,6	16,8	16,6	16,3	16,4	16,5	16,5	16,4
9.1.2015			16,0	16,0	16,1	16,3	16,1	15,6	16,2	16,1	16,0	16,2
18.2.2015	38,5	38,2	16,4	16,3	17,0	16,4	16,4	16,4	16,4	16,6	16,5	16,6
5.3.2015	38,5	38,3										
1.4.2015	37,5	37,6	16,1	16,0	16,6	16,4	16,2	16,4	16,2	16,4	16,2	16,2
9.4.2015	38,2	38,2	16,7	16,3	17,0	16,7	16,5	15,3	16,6	16,6	16,6	16,6
6.5.2015	38,5	38,6	16,5	16,5	17,3	17,4	16,8	17,0	16,9	16,9	16,8	16,8
1.7.2015	39,3	39,2	16,6	16,6	17,7	17,7	17,0	16,7	16,9	17,0	16,9	17,0

*Table A1.b. Ambient pressure (hPa) measurements during the check of flow rates of each RMs and CMs.*

*Table A1.c. Ambient temperature (°C) measurements during the check of flow rates of each RM and CMs.*

*Table A1.d. Cleaning and greasing of the inlets of the RMs and CMs.*

Leckel (11/0070)	Leckel (09/0081)	SHARP A / B	TEOM A / B	MP-101 A / B	FH-62-IR A / B	BAM A / B	Dust track	Grimm
2.4.2014	3.4.2014	29.4.2014	29.4.2014	29.4.2014	29.4.2014	29.4.2014	9.4.2015	1.7.2014
29.4.2014	30.4.2014	10.6.2014	10.6.2014	10.6.2014	10.6.2014	10.6.2014		9.4.2015
27.5.2014	27.5.2014	16.7.2014	16.7.2014	16.7.2014	16.7.2014	16.7.2014		
24.6.2014	25.6.2014	7.10.2014	7.10.2014	7.10.2014	7.10.2014	7.10.2014		
16.7.2014	16.7.2014	5.12.2014	5.12.2014	5.12.2014	5.12.2014	5.12.2014		
12.8.2014	13.8.2014	21.1.2015	21.1.2015	21.1.2015	21.1.2015	21.1.2015		
9.9.2014	9.9.2014	18.3.2015	18.3.2015	18.3.2015	18.3.2015	18.3.2015		
21.10.2014	22.10.2014	9.4.2015	9.4.2015	9.4.2015	9.4.2015	9.4.2015		
18.11.2014	19.11.2014	6.5.2015	6.5.2015	6.5.2015	6.5.2015	6.5.2015		
5.12.2014	5.12.2014	3.6.2015	3.6.2015	3.6.2015	3.6.2015	3.6.2015		
8.1.2015	8.1.2015	1.7.2015	1.7.2015	1.7.2015	1.7.2015	1.7.2015		
4.2.2015		5.2.2015						
18.2.2015		19.2.2015						
18.3.2015		19.3.2015						
9.4.2015		9.4.2015						
22.4.2015		22.4.2015						
20.5.2015		20.5.2015						
3.6.2015		3.6.2015						
15.6.2015		15.6.2015						
1.7.2015		1.7.2015						

*Table A1. e. Software versions of the instruments*

INSTRUMENTS	Serial number	SOFTWARE version	Changes during the maintenance	Changes on the performance of the instrument	Date	Additional factors installed on signal
Leckel SEQ 47/50			No	No		No
Leckel SEQ 47/50			No	No		No
Environnement MP-101+CPM A	3298	v 3.6 e	No	No		No
Environnement MP-101+CPM B	3008	v 3.6 i	No	No		No
GRIMM-180 B	18A07019	Firmware Version,12.30	No	No		No
GRIMM-180 C	18A06016	Firmware Version, 12.30	No	No		No
SHARP 5030 A	E-204	Firmware version V1.16	V1.21	No changes on the performance of the instrument	During maintenance of the instrument. Performed by the Finnish representative 15.7.2014	No
SHARP 5030 B	E-2154	Firmware version V1.16	V1.21	No changes on the performance of the instrument	During maintenance of the instrument. Performed by the Finnish representative 15.7.2014	No
FH-62-IR A	1163	V 1.10	No	No		No
FH-62-IR B	1159	V 1.10	No	No		No
BAM 1020 A	H1207	3236-02 3.1.0	No	No		No
BAM 1020 B	H1208	3236-02 3.1.0	No	No		No
TEOM-1405 A	1405A2 2686 1312	Version 1.57.00	No	No		1.03*Sign + 3
TEOM-1405 B	1405A2 2638 1310	Version 1.57.00	No	No		1.03*Sign + 3
OSIRIS A	TNO 3384	Rev. 04.34	No	No		No
OSIRIS B	TNO 3383	Rev. 04.34	No	No		No
DUST TRAK DRX 8533 EP A	8533144308	Firmware Version,3.3	No	No		No
DUST TRAK DRX 8533 EP B	8533144421	Firmware Version,3.3	No	No		No

## A2 PM<sub>10</sub> equivalence tests for BAM




Figure A2.1. Time series of 24 h concentration values of BAM and the RMs during the  $\text{PM}_{10}$  equivalence campaigns in Kuopio.




Figure A2.2: Scatter plot of BAM impactor versus the RMs: Calibrated data.

Table A2.1. Equivalence test results for BAM against the RMs for  $\text{PM}_{10}$ , all data

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>0,942y + 0,437</b>		N (Spring)	93	n
Regression (i=0)	<b>0,947y</b>	n	N (Summer)	43	n
N	211	n	N (Fall)	0	n
Outliers	13	n	N (Winter)	75	n
Outliers	6,2%	%	Outliers	10	n
Mean CM	25,1	µg/m³	Outliers	4,7%	%
Mean RM	24,1	µg/m³	Mean CM	24,071	µg/m³
Number of RM > 0,6 * LV	46	n	Mean RM	24,071	µg/m³
Number of RM > LV	16	n	Number of CM > 0,6 * LV	49	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>1,06178</b>	significant	Slope b	1,000	
Uncertainty of b	0,006		Uncertainty of b	0,006	
Intercept a	<b>-0,46385</b>		Intercept a	0,005	
Uncertainty of a	0,281		Uncertainty of a	0,264	
r^2	0,993		r^2	0,993	
Slope b forced trough origin	<b>1,056</b>	significant			
Uncertainty of b (forced)	0,0052				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,42	µg/m³	Calibration	<b>0,942y + 0,437</b>	µg/m³
Uncertainty of calibration (forced)	0,26	µg/m³	u(calibration)	<b>0,415</b>	µg/m³
Random term	3,33	µg/m³	Random term	<b>3,14</b>	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	<b>0,00</b>	µg/m³
Bias at LV	2,63	µg/m³	Bias at LV	<b>-0,01</b>	µg/m³
Combined uncertainty	4,24	µg/m³	Combined uncertainty	<b>3,14</b>	µg/m³
Expanded relative uncertainty	<b>16,9%</b>	pass	Expanded relative uncertainty	<b>12,6%</b>	pass
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³

### A3. $PM_{10}$ equivalence tests for DustTrak analyzers




Figure A3.1: Time series of 24 h concentration values of DustTrak and the RMs during the  $PM_{10}$  equivalence campaigns in Kuopio.




Figure A3.2. Scatter plot of DustTrak versus the RMs: Calibrated data.

Table A3.1. Equivalence test results for DustTrak against the RMs for  $PM_{10}$ , all data.

RAW DATA		RESULTS AFTER CALIBRATING	
Regression	<b>7,478y + -76,819</b>	N (Spring)	82
Regression (i=0)	<b>4,16y</b>	N (Summer)	31
N	171	N (Fall)	0
Outliers	5	N (Winter)	58
Outliers	2,9%	Outliers	112
Mean CM	13,8	Outliers	65,5%
Mean RM	26,6	Mean CM	26,571
Number of RM > 0,6 * LV	35	Mean RM	26,571
Number of RM > LV	21	Number of CM > 0,6 * LV	57
		Number of CM > LV	48
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)	
Slope b	<b>0,13373</b>	Slope b	2,446
Uncertainty of b	0,015	Uncertainty of b	0,110
Intercept a	<b>10,27330</b>	Intercept a	-38,413
Uncertainty of a	0,777	Uncertainty of a	5,807
r^2	0,314	r^2	0,314
Slope b forced trough origin	<b>0,240</b>		
Uncertainty of b (forced)	0,0182		
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)	
Uncertainty of calibration	<b>1,07</b>	Calibration	<b>7,478y -76,819</b>
Uncertainty of calibration (forced)	0,91	u(calibration)	<b>1,068</b>
Random term	8,73	Random term	94,70
Additional uncertainty (optional)	0,00	Additional uncertainty (optional)	0,00
Bias at LV	-33,04	Bias at LV	33,87
Combined uncertainty	34,17	Combined uncertainty	100,57
Expanded relative uncertainty	<b>136,7%</b>	Expanded relative uncertainty	<b>402,3%</b>
Ref sampler uncertainty	1,00	Ref sampler uncertainty	1,00
Limit value	50	Limit value	50

#### A4. $PM_{10}$ equivalence tests for FH 62 I-R analyzers




Figure A4.1: Time series of 24 h concentration values of FH 62 I-R and the RMs during the  $PM_{10}$  equivalence campaigns in Kuopio.




Figure A4.2. Scatter plot of FH 62 I-R versus the RMs: Calibrated data

Table A4.1. Equivalence test results for FH 62 I-R against the RMs for  $PM_{10}$ , all data.

RAW DATA		RESULTS AFTER CALIBRATING		
Regression	<b>1,3y + -0,904</b>		N (Spring)	92
Regression (i=0)	<b>1,288y</b>	n	N (Summer)	40
N	195		N (Fall)	0
Outliers	7	n	N (Winter)	63
Outliers	3,6%	%	Outliers	20
Mean CM	20,7	µg/m³	Outliers	10,3%
Mean RM	26,1	µg/m³	Mean CM	26,068
Number of RM > 0,6 * LV	44	n	Mean RM	26,068
Number of RM > LV	18	n	Number of CM > 0,6 * LV	42
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>0,76908</b>	significant	Slope b	1,001
Uncertainty of b	0,006		Uncertainty of b	0,007
Intercept a	<b>0,69496</b>	significant	Intercept a	-0,033
Uncertainty of a	0,274		Uncertainty of a	0,356
r^2	0,990		r^2	0,990
Slope b forced trough origin	<b>0,777</b>	significant		
Uncertainty of b (forced)	0,0048			
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,39	µg/m³	Calibration	<b>1,300y -0,904</b>
Uncertainty of calibration (forced)	0,24	µg/m³	u(calibration)	<b>0,388</b>
Random term	3,10	µg/m³	Random term	4,13
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00
Bias at LV	-10,85	µg/m³	Bias at LV	0,03
Combined uncertainty	11,28	µg/m³	Combined uncertainty	4,13
Expanded relative uncertainty	<b>45,1%</b>	fail	Expanded relative uncertainty	<b>16,5%</b>
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00
Limit value	50	µg/m³	Limit value	50

## A5. $PM_{10}$ equivalence tests for Grimm analyzers




Figure A5.1: Time series of 24 h concentration values of Grimm and the RMs during the  $PM_{10}$  equivalence campaigns in Kuopio.




Figure A5.2. Scatter plot of Grimm versus the RMs: Calibrated data

Table A5.1. Equivalence test results for Grimm against the RMs for  $PM_{10}$ , all data.

RAW DATA		RESULTS AFTER CALIBRATING	
Regression	<b>0,855y + 2,139</b>	N (Spring)	45
Regression (i=0)	<b>0,908y</b>	N (Summer)	53
N	179	N (Fall)	14
Outliers	8	N (Winter)	67
Outliers	4,5%	Outliers	5
Mean CM	17,7	Outliers	2,8%
Mean RM	17,3	Mean CM	17,312
Number of RM > 0,6 * LV	27	Mean RM	17,312
Number of RM > LV	11	Number of CM > 0,6 * LV	26
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)	
Slope b	<b>1,16941</b>	Slope b	0,996
Uncertainty of b	0,020	Uncertainty of b	0,017
Intercept a	<b>-2,50173</b>	Intercept a	0,075
Uncertainty of a	0,508	Uncertainty of a	0,435
r^2	0,946	r^2	0,946
Slope b forced trough origin	<b>1,102</b>		
Uncertainty of b (forced)	0,0152		
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)	
Uncertainty of calibration	1,14	Calibration	<b>0,855y + 2,139</b>
Uncertainty of calibration (forced)	0,76	u(calibration)	<b>1,136</b>
Random term	4,85	Random term	4,26
Additional uncertainty (optional)	0,00	Additional uncertainty (optional)	0,00
Bias at LV	5,97	Bias at LV	-0,14
Combined uncertainty	7,69	Combined uncertainty	4,26
Expanded relative uncertainty	<b>30,8%</b>	Expanded relative uncertainty	<b>17,0%</b>
Ref sampler uncertainty	1,00	Ref sampler uncertainty	1,00
Limit value	50	Limit value	50

## A6. $PM_{10}$ equivalence tests for Environnement MP101 analyzers




Figure A6.1: Time series of 24 h concentration values of MP101 and the RMs during the  $PM_{10}$  equivalence campaigns in Kuopio




Figure A6.2: Scatter plot of MP101 versus the RMs: Calibrated data.

Table A6.1. Equivalence test results for MP101 against the RMs for  $PM_{10}$ , all data

RAW DATA		RESULTS AFTER CALIBRATING		
Regression	<b>0,811y + 2,311</b>		N (Spring)	70
Regression (i=0)	<b>0,83y</b>	n	N (Summer)	50
N	201		N (Fall)	12
Outliers	8	n	N (Winter)	69
Outliers	4,0%	%	Outliers	3
Mean CM	29,3	µg/m³	Mean CM	26,090
Mean RM	26,1	µg/m³	Mean RM	26,090
Number of RM > 0,6 * LV	47	n	Number of CM > 0,6 * LV	44
Number of RM > LV	17	n	Number of CM > LV	15
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>1,23377</b>	significant	Slope b	1,000
Uncertainty of b	0,006		Uncertainty of b	0,005
Intercept a	<b>-2,85155</b>	significant	Intercept a	0,013
Uncertainty of a	0,296		Uncertainty of a	0,240
r^2	0,995		r^2	0,995
Slope b forced trough origin	<b>1,204</b>	significant		
Uncertainty of b (forced)	0,0061			
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,42	µg/m³	Calibration	<b>0,811y + 2,311</b>
Uncertainty of calibration (forced)	0,30	µg/m³	u(calibration)	<b>0,422</b>
Random term	3,42	µg/m³	Random term	2,74
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00
Bias at LV	8,84	µg/m³	Bias at LV	-0,01
Combined uncertainty	9,47	µg/m³	Combined uncertainty	2,74
Expanded relative uncertainty	<b>37,9%</b>	fail	Expanded relative uncertainty	<b>11,0%</b>
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00
Limit value	50	µg/m³	Limit value	50

## A7. $PM_{10}$ equivalence tests for Osiris analyzers




Figure A7.1: Time series of 24 h concentration values of Osiris and the RMs during the  $PM_{10}$  equivalence campaigns in Kuopio.




Figure A7.2. Scatter plot of Osiris versus the RMs: Calibrated data

Table A7.1. Equivalence test results for Osiris against the RMs for  $PM_{10}$ , all data.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>1,401y + -0,153</b>		N (Spring)	142	n
Regression (i=0)	<b>1,398y</b>	n	N (Summer)	73	n
N	284	n	N (Fall)	0	n
Outliers	15	n	N (Winter)	69	n
Outliers	5,3%	%	Outliers	31	n
Mean CM	15,9	µg/m³	Mean CM	22,132	µg/m³
Mean RM	22,1	µg/m³	Mean RM	22,132	µg/m³
Number of RM > 0,6 * LV	48	n	Number of CM > 0,6 * LV	48	n
Number of RM > LV	19	n	Number of CM > LV	21	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>0,71389</b>	significant	Slope b	1,002	
Uncertainty of b	0,005		Uncertainty of b	0,007	
Intercept a	<b>0,10954</b>		Intercept a	-0,045	
Uncertainty of a	0,200		Uncertainty of a	0,281	
r^2	0,988		r^2	0,988	
Slope b forced trough origin	<b>0,715</b>	significant			
Uncertainty of b (forced)	0,0040				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,31	µg/m³	Calibration	<b>1,401y -0,153</b>	
Uncertainty of calibration (forced)	0,20	µg/m³	u(calibration)	<b>0,310</b>	µg/m³
Random term	2,71	µg/m³	Random term	3,93	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	-14,20	µg/m³	Bias at LV	0,06	µg/m³
Combined uncertainty	14,45	µg/m³	Combined uncertainty	3,93	µg/m³
Expanded relative uncertainty	<b>57,8%</b>	fail	Expanded relative uncertainty	<b>15,7%</b>	pass
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³

### A8. $PM_{10}$ equivalence tests for SHARP beta signal




Figure A8.1: Time series of 24 h concentration values of SHARP beta signal and the RMs during the  $PM_{10}$  equivalence campaigns in Kuopio.




Figure A8.2. Scatter plot of SHARP beta signal versus the RMs: Calibrated data

Table A8.1. Equivalence test results for SHARP beta signal against the RMs for PM<sub>10</sub>, all data

RAW DATA		RESULTS AFTER CALIBRATING	
Regression	<b>1,415y + -2,233</b>	N (Spring)	72
Regression (i=0)	<b>1,38y</b>	N (Summer)	50
N	203	N (Fall)	12
Outliers	9	N (Winter)	69
Outliers	4,4%	Outliers	18
Mean CM	19,7	Outliers	8,9%
Mean RM	25,6	Mean CM	25,615
Number of RM > 0,6 * LV	47	Mean RM	25,615
Number of RM > LV	18	Number of CM > 0,6 * LV	46
		Number of CM > LV	17
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)	
Slope b	<b>0,70662</b>	Slope b	1,001
Uncertainty of b	0,004	Uncertainty of b	0,006
Intercept a	<b>1,57753</b>	Intercept a	-0,029
Uncertainty of a	0,196	Uncertainty of a	0,277
r^2	0,993	r^2	0,993
Slope b forced trough origin	<b>0,724</b>		
Uncertainty of b (forced)	0,0040		
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)	
Uncertainty of calibration	0,28	Calibration	<b>1,415y -2,233</b>
Uncertainty of calibration (forced)	0,20	u(calibration)	<b>0,282</b>
Random term	2,14	Random term	3,20
Additional uncertainty (optional)	0,00	Additional uncertainty (optional)	0,00
Bias at LV	-13,09	Bias at LV	0,03
Combined uncertainty	13,27	Combined uncertainty	3,20
Expanded relative uncertainty	<b>53,1%</b>	Expanded relative uncertainty	<b>12,8%</b>
Ref sampler uncertainty	1,00	Ref sampler uncertainty	1,00
Limit value	50	Limit value	50

### A9. $PM_{10}$ equivalence tests for SHARP c-dust analyzers




Figure A9.1: Time series of 24 h concentration values of SHARP c-dust and the RMs during the  $PM_{10}$  equivalence campaigns in Kuopio.




Figure A9.2. Scatter plot of SHARP c-dust versus the RMs: Calibrated data  
Table A9.1. Equivalence test results for SHARP c-dust against the RMs for  $PM_{10}$ , all data

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>1,404y + -2,75</b>		N (Spring)	69	n
Regression (i=0)	<b>1,362y</b>	n	N (Summer)	50	n
N	197	n	N (Fall)	12	n
Outliers	9	n	N (Winter)	66	n
Outliers	4,6%	%	Outliers	19	n
Mean CM	19,8	µg/m³	Outliers	9,6%	%
Mean RM	25,0	µg/m³	Mean CM	24,980	µg/m³
Number of RM > 0,6 * LV	41	n	Mean RM	24,980	µg/m³
Number of RM > LV	15	n	Number of CM > 0,6 * LV	41	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>0,71223</b>	significant	Slope b	1,002	
Uncertainty of b	0,005		Uncertainty of b	0,008	
Intercept a	<b>1,95866</b>	significant	Intercept a	-0,046	
Uncertainty of a	0,260		Uncertainty of a	0,366	
r^2	0,989		r^2	0,989	
Slope b forced trough origin	<b>0,734</b>	significant			
Uncertainty of b (forced)	0,0053				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,38	µg/m³	Calibration	<b>1,404y -2,750</b>	µg/m³
Uncertainty of calibration (forced)	0,26	µg/m³	u(calibration)	<b>0,376</b>	µg/m³
Random term	2,96	µg/m³	Random term	4,29	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	-12,43	µg/m³	Bias at LV	0,05	µg/m³
Combined uncertainty	12,78	µg/m³	Combined uncertainty	4,29	µg/m³
Expanded relative uncertainty	<b>51,1%</b>	fail	Expanded relative uncertainty	<b>17,2%</b>	pass
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³

### A10. $PM_{10}$ equivalence tests for TEOM analyzers




Figure A10.1: Time series of 24 h concentration values of TEOM and the RMs during the  $PM_{10}$  equivalence campaigns in Kuopio.




Figure A10.2: Scatter plot of TEOM versus the RMs: Calibrated data.

Table A10.1. Equivalence test results for TEOM against the RMs for  $PM_{10}$ , all data.

RAW DATA		RESULTS AFTER CALIBRATING	
Regression	<b>0,868y + -2,068</b>	N (Spring)	69
Regression (i=0)	<b>0,848y</b>	N (Summer)	46
N	194	N (Fall)	11
Outliers	9	N (Winter)	68
Outliers	4,6%	Outliers	8
Mean CM	32,3	Outliers	4,1%
Mean RM	26,0	Mean CM	25,965
Number of RM > 0,6 * LV	44	Mean RM	25,965
Number of RM > LV	16	Number of CM > 0,6 * LV	43
		Number of CM > LV	16
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)	
Slope b	<b>1,15264</b>	Slope b	0,999
Uncertainty of b	0,007	Uncertainty of b	0,006
Intercept a	<b>2,38402</b>	Intercept a	0,014
Uncertainty of a	0,358	Uncertainty of a	0,311
r^2	0,992	r^2	0,992
Slope b forced trough origin	<b>1,179</b>		
Uncertainty of b (forced)	0,0069		
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)	
Uncertainty of calibration	0,51	Calibration	<b>0,868y -2,068</b>
Uncertainty of calibration (forced)	0,34	u(calibration)	<b>0,509</b>
Random term	4,13	Random term	3,59
Additional uncertainty (optional)	0,00	Additional uncertainty (optional)	0,00
Bias at LV	10,02	Bias at LV	-0,01
Combined uncertainty	10,84	Combined uncertainty	3,59
Expanded relative uncertainty	<b>43,3%</b>	Expanded relative uncertainty	<b>14,4%</b>
Ref sampler uncertainty	1,00	Ref sampler uncertainty	1,00
Limit value	50	Limit value	50

## A11: Calibration factors for CM for each PM<sub>10</sub> campaign




Figure A11.1: Slopes of the calibration equations for each CM at each campaign.




Figure A11.2: Intercepts for each CM for each campaign.

## A12. PM<sub>2.5</sub> equivalence tests for BAM analyzers




Figure A12.1: Time series of 24 h concentration values of BAM and the RMs during the  $\text{PM}_{2.5}$  equivalence campaigns in Kuopio.




Figure A12.2: Scatter plot of BAM versus the RMs: Calibrated data.

Table A12.1. Equivalence test results for BAM against the RMs for  $\text{PM}_{2.5}$ , all data.

RAW DATA		RESULTS AFTER CALIBRATING		
Regression	<b>1,1y + 0,733</b>		N (Spring)	50
Regression (i=0)	<b>1,215y</b>	n	N (Summer)	75
N	195		N (Fall)	69
Outliers	4	n	N (Winter)	1
Outliers	2,1%	%	Outliers	3
Mean CM	5,2	µg/m³	Mean CM	6,497
Mean RM	6,5	µg/m³	Mean RM	6,497
Number of RM > 0,6 * LV	0	n	Number of CM > 0,6 * LV	0
Number of RM > LV	0	n	Number of CM > LV	0
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>0,90941</b>	significant	Slope b	1,009
Uncertainty of b	0,027		Uncertainty of b	0,030
Intercept a	<b>-0,66676</b>	significant	Intercept a	-0,059
Uncertainty of a	0,192		Uncertainty of a	0,211
r^2	0,832		r^2	0,832
Slope b forced trough origin	<b>0,823</b>	significant		
Uncertainty of b (forced)	0,0109			
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,83	µg/m³	Calibration	<b>1,100y + 0,733</b>
Uncertainty of calibration (forced)	0,33	µg/m³	u(calibration)	<b>0,833</b>
Random term	0,47	µg/m³	Random term	1,09
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00
Bias at LV	-3,38	µg/m³	Bias at LV	0,21
Combined uncertainty	3,42	µg/m³	Combined uncertainty	1,11
Expanded relative uncertainty	<b>22,8%</b>	pass	Expanded relative uncertainty	<b>7,4%</b>
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00
Limit value	30	µg/m³	Limit value	30

### A13. PM<sub>2.5</sub> equivalence tests for DustTrak analyzers




Figure A13.1: Time series of 24 h concentration values of DustTrak and the RMs during the  $\text{PM}_{2.5}$  equivalence campaigns in Kuopio.




Figure A13.2: Scatter plot of DustTrak versus the RMs: Calibrated data.

Table A13.1. Equivalence test results for DustTrak against the RMs for  $\text{PM}_{2.5}$ , all data

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>0,602y + -1,002</b>		N (Spring)	83	n
Regression (i=0)	<b>0,55y</b>	n	N (Summer)	31	n
N	164	n	N (Fall)	0	n
Outliers	7	n	N (Winter)	50	n
Outliers	4,3%	%	Outliers	0	n
Mean CM	13,2	µg/m³	Mean CM	6,956	µg/m³
Mean RM	7,0	µg/m³	Mean RM	6,956	µg/m³
Number of RM > 0,6 * LV	7	n	Number of CM > 0,6 * LV	8	n
Number of RM > LV	4	n	Number of CM > LV	0	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>1,66217</b>	significant	Slope b	0,853	significant
Uncertainty of b	0,077		Uncertainty of b	0,046	
Intercept a	<b>1,66540</b>	significant	Intercept a	1,024	significant
Uncertainty of a	0,727		Uncertainty of a	0,437	
r^2	0,564		r^2	0,564	
Slope b forced trough origin	<b>1,820</b>	significant			
Uncertainty of b (forced)	0,0605				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	2,41	µg/m³	Calibration	<b>0,602y -1,002</b>	µg/m³
Uncertainty of calibration (forced)	1,82	µg/m³	u(calibration)	<b>2,408</b>	µg/m³
Random term	7,23	µg/m³	Random term	4,56	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	21,53	µg/m³	Bias at LV	-3,39	µg/m³
Combined uncertainty	22,71	µg/m³	Combined uncertainty	5,69	µg/m³
Expanded relative uncertainty	<b>151,4%</b>	fail	Expanded relative uncertainty	<b>37,9%</b>	fail
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	30	µg/m³	Limit value	30	µg/m³

### A14. PM<sub>2.5</sub> equivalence tests for FH 62 I-R analyzers




Figure A14.1: Time series of 24 h concentration values of FH 62 I-R and the RMs during the PM<sub>2.5</sub> equivalence campaigns in Kuopio.




Figure A14.2: Scatter plot of FH 62 I-R versus the RMs: Calibrated data.

Table A14.1. Equivalence test results for FH 62 I-R against the RMs for PM<sub>2.5</sub>, all data.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>0,85y + 1,709</b>		N (Spring)	51	n
Regression (i=0)	<b>1,097y</b>		N (Summer)	73	n
N	195	n	N (Fall)	70	n
Outliers	7	n	N (Winter)	1	n
Outliers	3,6%	%	Outliers	3	n
Mean CM	5,6	µg/m³	Mean CM	6,492	µg/m³
Mean RM	6,5	µg/m³	Mean RM	6,492	µg/m³
Number of RM > 0,6 * LV	0	n	Number of CM > 0,6 * LV	0	n
Number of RM > LV	0	n	Number of CM > LV	0	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>1,17657</b>	significant	Slope b	0,950	
Uncertainty of b	0,053		Uncertainty of b	0,045	
Intercept a	<b>-2,01057</b>	significant	Intercept a	0,322	
Uncertainty of a	0,376		Uncertainty of a	0,320	
r^2	0,578		r^2	0,578	
Slope b forced trough origin	<b>0,912</b>	significant			
Uncertainty of b (forced)	0,0216				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1,63	µg/m³	Calibration	<b>0,850y + 1,709</b>	
Uncertainty of calibration (forced)	0,65	µg/m³	u(calibration)	<b>1,632</b>	µg/m³
Random term	2,11	µg/m³	Random term	2,32	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	3,29	µg/m³	Bias at LV	-1,17	µg/m³
Combined uncertainty	3,91	µg/m³	Combined uncertainty	2,60	µg/m³
Expanded relative uncertainty	<b>26,0%</b>	fail	Expanded relative uncertainty	<b>17,3%</b>	pass
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	30	µg/m³	Limit value	30	µg/m³

### A15. PM<sub>2.5</sub> equivalence tests for Grimm analyzers




Figure A15.1: Time series of 24 h concentration values of Grimm and the RMs during the PM<sub>2.5</sub> equivalence campaigns in Kuopio.




Figure A15.2: Scatter plot of Grimm versus the RMs: Calibrated data.

Table A15.1. Equivalence test results for Grimm against the RMs for PM<sub>2.5</sub>, all data.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>0,747y + 0,532</b>		N (Spring)	73	n
Regression (i=0)	<b>0,78y</b>		N (Summer)	45	n
N	173	n	N (Fall)	5	n
Outliers	6	n	N (Winter)	50	n
Outliers	3,5%	%	Outliers	4	n
Mean CM	9,4	µg/m³	Mean CM	7,586	µg/m³
Mean RM	7,6	µg/m³	Mean RM	7,586	µg/m³
Number of RM > 0,6 * LV	6	n	Number of CM > 0,6 * LV	7	n
Number of RM > LV	3	n	Number of CM > LV	4	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>1,33912</b>	significant	Slope b	0,986	
Uncertainty of b	0,031		Uncertainty of b	0,023	
Intercept a	<b>-0,71199</b>	significant	Intercept a	0,108	
Uncertainty of a	0,301		Uncertainty of a	0,225	
r^2	0,907		r^2	0,907	
Slope b forced trough origin	<b>1,282</b>	significant			
Uncertainty of b (forced)	0,0193				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,97	µg/m³	Calibration	<b>0,747y + 0,532</b>	µg/m³
Uncertainty of calibration (forced)	0,58	µg/m³	u(calibration)	<b>0,974</b>	µg/m³
Random term	2,32	µg/m³	Random term	1,86	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	9,46	µg/m³	Bias at LV	-0,32	µg/m³
Combined uncertainty	9,74	µg/m³	Combined uncertainty	1,89	µg/m³
Expanded relative uncertainty	<b>65,0%</b>	fail	Expanded relative uncertainty	<b>12,6%</b>	pass
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	30	µg/m³	Limit value	30	µg/m³

### A16. *PM<sub>2.5</sub>* equivalence tests for Environnement MP101 analyzers




Figure A16.1: Time series of 24 h concentration values of MP101 and the RMs during the  $\text{PM}_{2.5}$  equivalence campaigns in Kuopio.




Figure A16.2: Scatter plot of MP101 versus the RMs: Calibrated data.

Table A16.1. Equivalence test results for MP101 against the RMs for  $\text{PM}_{2.5}$ , all data.

RAW DATA		RESULTS AFTER CALIBRATING	
Regression	<b>0,812y + -0,306</b>	N (Spring)	50
Regression (i=0)	<b>0,78y</b>	N (Summer)	73
N	196	N (Fall)	72
Outliers	4	N (Winter)	1
Outliers	2,0%	Outliers	1
Mean CM	8,2	Outliers	0,5%
Mean RM	6,4	Mean CM	6,379
Number of RM > 0,6 * LV	0	Mean RM	6,379
Number of RM > LV	0	Number of CM > 0,6 * LV	0
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)	
Slope b	<b>1,23166</b>	Slope b	0,980
Uncertainty of b	0,036	Uncertainty of b	0,029
Intercept a	<b>0,37721</b>	Intercept a	0,126
Uncertainty of a	0,249	Uncertainty of a	0,202
r^2	0,831	r^2	0,831
Slope b forced trough origin	<b>1,282</b>		
Uncertainty of b (forced)	0,0152		
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)	
Uncertainty of calibration	1,10	Calibration	<b>0,812y -0,306</b>
Uncertainty of calibration (forced)	0,46	u(calibration)	<b>1,099</b>
Random term	1,06	Random term	1,26
Additional uncertainty (optional)	0,00	Additional uncertainty (optional)	0,00
Bias at LV	7,33	Bias at LV	-0,47
Combined uncertainty	7,40	Combined uncertainty	1,34
Expanded relative uncertainty	<b>49,4%</b>	Expanded relative uncertainty	<b>8,9%</b>
Ref sampler uncertainty	1,00	Ref sampler uncertainty	1,00
Limit value	30	Limit value	30

### A17. PM<sub>2.5</sub> equivalence tests for Osiris analyzers




Figure A17.1: Time series of 24 h concentration values of Osiris and the RMs during the PM<sub>2.5</sub> equivalence campaigns in Kuopio.




Figure A17.2: Scatter plot of Osiris versus the RMs: Calibrated data.

Table A17.1. Equivalence test results for Osiris against the RMs for PM<sub>2.5</sub>, all data.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>3,324y + -1,073</b>		N (Spring)	51	n
Regression (i=0)	<b>2,92y</b>	n	N (Summer)	78	n
N	195	n	N (Fall)	65	n
Outliers	5	n	N (Winter)	1	n
Outliers	2,6%	%	Outliers	87	n
Mean CM	2,3	µg/m³	Mean CM	6,539	µg/m³
Mean RM	6,5	µg/m³	Mean RM	6,539	µg/m³
Number of RM > 0,6 * LV	0	n	Number of CM > 0,6 * LV	3	n
Number of RM > LV	0	n	Number of CM > LV	0	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>0,30086</b>	significant	Slope b	1,773	significant
Uncertainty of b	0,025		Uncertainty of b	0,082	
Intercept a	<b>0,32292</b>		Intercept a	-5,054	significant
Uncertainty of a	0,177		Uncertainty of a	0,587	
r^2	0,383		r^2	0,383	
Slope b forced trough origin	<b>0,343</b>	significant			
Uncertainty of b (forced)	0,0105				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,76	µg/m³	Calibration	<b>3,324y -1,073</b>	
Uncertainty of calibration (forced)	0,31	µg/m³	u(calibration)	<b>0,757</b>	µg/m³
Random term	0,27	µg/m³	Random term	4,29	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	-20,65	µg/m³	Bias at LV	18,13	µg/m³
Combined uncertainty	20,65	µg/m³	Combined uncertainty	18,63	µg/m³
Expanded relative uncertainty	<b>137,7%</b>	fail	Expanded relative uncertainty	<b>124,2%</b>	fail
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	30	µg/m³	Limit value	30	µg/m³

### A18. PM<sub>2.5</sub> equivalence tests for SHARP beta analyzers




Figure A18.1: Time series of 24 h concentration values of SHARP beta and the RMs during the PM<sub>2.5</sub> equivalence campaigns in Kuopio.




Figure A18.2: Scatter plot of SHARP beta versus the RMs: Calibrated data.

Table A18.1. Equivalence test results for SHARP beta against the RMs for PM<sub>2.5</sub>, all data.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>0,971y + -0,003</b>		N (Spring)	50	n
Regression (i=0)	<b>0,971y</b>	n	N (Summer)	73	n
N	196	n	N (Fall)	72	n
Outliers	6	n	N (Winter)	1	n
Outliers	3,1%	%	Outliers	4	n
Mean CM	6,6	µg/m³	Mean CM	6,365	µg/m³
Mean RM	6,4	µg/m³	Mean RM	6,365	µg/m³
Number of RM > 0,6 * LV	0	n	Number of CM > 0,6 * LV	0	n
Number of RM > LV	0	n	Number of CM > LV	0	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>1,02967</b>		Slope b	0,999	
Uncertainty of b	0,020		Uncertainty of b	0,019	
Intercept a	<b>0,00306</b>		Intercept a	0,007	
Uncertainty of a	0,137		Uncertainty of a	0,133	
r^2	0,929		r^2	0,929	
Slope b forced trough origin	<b>1,030</b>	significant			
Uncertainty of b (forced)	0,0080				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,61	µg/m³	Calibration	<b>0,971y -0,003</b>	
Uncertainty of calibration (forced)	0,24	µg/m³	u(calibration)	<b>0,607</b>	µg/m³
Random term	0,00	µg/m³	Random term	0,00	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	0,89	µg/m³	Bias at LV	-0,03	µg/m³
Combined uncertainty	0,89	µg/m³	Combined uncertainty	0,03	µg/m³
Expanded relative uncertainty	<b>6,0%</b>	pass	Expanded relative uncertainty	<b>0,2%</b>	pass
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	30	µg/m³	Limit value	30	µg/m³

### A19. PM<sub>2.5</sub> equivalence tests for SHARP c-dust analyzers




Figure A19.1: Time series of 24 h concentration values of SHARP c-dust and the RMs during the PM<sub>2.5</sub> equivalence campaigns in Kuopio.




Figure A19.2: Scatter plot of SHARP c-dust versus the RMs: Calibrated data.

Table A19.1. Equivalence test results for SHARP c-dust against the RMs for PM<sub>2.5</sub>, all data.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>0,854y + 1,187</b>		N (Spring)	50	n
Regression (i=0)	<b>1,009y</b>	n	N (Summer)	72	n
N	195	n	N (Fall)	72	n
Outliers	9	n	N (Winter)	1	n
Outliers	4,6%	%	Outliers	0	n
Mean CM	6,1	µg/m³	Mean CM	6,418	µg/m³
Mean RM	6,4	µg/m³	Mean RM	6,418	µg/m³
Number of RM > 0,6 * LV	0	n	Number of CM > 0,6 * LV	0	n
Number of RM > LV	0	n	Number of CM > LV	0	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>1,17085</b>	significant	Slope b	0,988	
Uncertainty of b	0,031		Uncertainty of b	0,027	
Intercept a	<b>-1,38994</b>	significant	Intercept a	0,078	
Uncertainty of a	0,220		Uncertainty of a	0,188	
r^2	0,860		r^2	0,860	
Slope b forced trough origin	<b>0,991</b>				
Uncertainty of b (forced)	0,0132				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,96	µg/m³	Calibration	<b>0,854y + 1,187</b>	µg/m³
Uncertainty of calibration (forced)	0,39	µg/m³	u(calibration)	<b>0,962</b>	µg/m³
Random term	0,81	µg/m³	Random term	1,06	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	3,74	µg/m³	Bias at LV	-0,29	µg/m³
Combined uncertainty	3,82	µg/m³	Combined uncertainty	1,10	µg/m³
Expanded relative uncertainty	<b>25,5%</b>	fail	Expanded relative uncertainty	<b>7,3%</b>	pass
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	30	µg/m³	Limit value	30	µg/m³

## A20. PM<sub>2.5</sub> equivalence tests for TEOM analyzers




Figure A20.1: Time series of 24 h concentration values of TEOM and the RMs during the PM<sub>2.5</sub> equivalence campaigns in Kuopio.




Figure A20.2: Scatter plot of TEOM versus the RMs: Calibrated data.

Table A20.1. Equivalence test results for TEOM against the RMs for PM<sub>2.5</sub>, all data.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	<b>1,009y + -1,681</b>		N (Spring)	50	n
Regression (i=0)	<b>0,821y</b>	n	N (Summer)	73	n
N	196	n	N (Fall)	72	n
Outliers	8	n	N (Winter)	1	n
Outliers	4,1%	%	Outliers	2	n
Mean CM	8,0	µg/m³	Mean CM	6,378	µg/m³
Mean RM	6,4	µg/m³	Mean RM	6,378	µg/m³
Number of RM > 0,6 * LV	0	n	Number of CM > 0,6 * LV	0	n
Number of RM > LV	0	n	Number of CM > LV	0	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	<b>0,99107</b>		Slope b	<b>1,001</b>	
Uncertainty of b	0,032		Uncertainty of b	0,033	
Intercept a	<b>1,66604</b>	significant	Intercept a	-0,007	
Uncertainty of a	0,226		Uncertainty of a	0,228	
r^2	0,794		r^2	0,794	
Slope b forced trough origin	<b>1,218</b>	significant			
Uncertainty of b (forced)	0,0163				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	<b>1,00</b>	µg/m³	Calibration	<b>1,009y -1,681</b>	µg/m³
Uncertainty of calibration (forced)	0,49	µg/m³	u(calibration)	<b>0,996</b>	µg/m³
Random term	0,85	µg/m³	Random term	<b>1,32</b>	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	1,40	µg/m³	Bias at LV	0,03	µg/m³
Combined uncertainty	1,64	µg/m³	Combined uncertainty	1,32	µg/m³
Expanded relative uncertainty	<b>10,9%</b>	pass	Expanded relative uncertainty	<b>8,8%</b>	pass
Ref sampler uncertainty	1,00	µg/m³	Ref sampler uncertainty	1,00	µg/m³
Limit value	30	µg/m³	Limit value	30	µg/m³

### A21: Calibration factors for CM for each PM<sub>2.5</sub> campaign




Figure A21.1: Slopes of the calibration equations for each CM at each campaign




Figure A31.2: Intercepts for each CM for each campaign.

## A22 Daily values of PM 10: Reference method, BAM, FH 62IR, Osiris

Date	Locket R1		Locket R2		BAM-A-21		BAM-B-25		FH62-A-29		FH62-B-33		Osiris-A-38		Osiris-B-42		Campaign		Date	Locket R1		Locket R2		BAM-A-21		BAM-B-25		FH62-A-29		FH62-B-33		Osiris-A-38		Osiris-B-42		Campaign	
	(09/08/01) PM10	(11/00/70) PM10	(09/08/01) conc. PM10	(11/00/70) PM10	(09/08/01) PM10	(11/00/70) PM10																															
20.3.2014	19.65	18.99	20.50	18.25	14.20	14.46	11.06	12.57	Tasa 1	5.12.2014	3.09	2.95	1.00	1.00	8.44	4.84	5.56	Sai 1																			
21.3.2014	7.89	7.90	7.57	5.94	7.22	6.03	5.96	3.85	Tasa 1	6.12.2014	3.43	3.24	1.76	0.38	1.64	2.99	3.24	Sai 1																			
22.3.2014	33.63	34.82	41.01	40.44	21.44	22.10	22.19	26.43	Tasa 1	7.12.2014	4.90	4.90	3.62	2.62	7.30	3.15	3.34	Sai 1																			
23.3.2014	16.92	17.34	17.92	17.30	13.85	12.08	10.75	11.49	Tasa 1	8.12.2014	5.24	5.16	3.63	3.25	12.52	4.56	4.45	Sai 1																			
25.3.2014	30.76	31.43	39.71	37.11	24.73	23.54	23.30	26.03	Tasa 1	9.12.2014	7.33	7.40	6.26	6.38	2.14	6.07	2.89	Sai 1																			
27.3.2014	31.04	29.63	35.41	33.25	23.80	23.43	23.57	23.43	Tasa 1	10.12.2014	10.14	9.65	11.12	10.49	5.64	6.88	3.30	Sai 1																			
28.3.2014	49.38	51.79	47.68	48.19	33.96	33.88	39.89	46.16	Tasa 1	11.12.2014	9.58	9.06	8.00	7.00	8.52	8.37	3.54	Sai 1																			
29.3.2014	23.84	23.96	30.25	27.68	18.21	17.46	16.35	18.14	Tasa 1	12.12.2014	10.36	10.25	9.00	8.87	9.32	8.91	3.97	Sai 1																			
1.4.2014	33.47	31.42	41.71	38.07	32.16	28.26	25.81	32.83	Tasa 1	13.12.2014	5.15	4.97	4.37	3.37	3.96	3.43	3.60	Sai 1																			
2.4.2014	36.66	37.94	50.49	48.31	34.39	29.46	23.76	31.71	Tasa 1	14.12.2014	5.87	6.21	5.25	3.75	7.03	6.44	3.69	Sai 1																			
3.4.2014	20.61	20.19	27.46	24.13	13.44	14.69	12.33	15.15	Tasa 1	15.12.2014	6.46	6.47	5.37	4.62	4.73	5.14	4.54	Sai 1																			
4.4.2014	37.80	37.49	43.79	41.21	25.96	28.77	28.31	34.08	Tasa 1	16.12.2014	6.33	6.38	4.65	4.15	6.93	6.93	4.36	Sai 1																			
5.4.2014	18.62	18.27	24.31	22.84	16.58	15.86	10.17	12.87	Tasa 1	17.12.2014	7.32	7.07	5.38	5.04	7.35	4.69	5.16	Sai 1																			
6.4.2014	29.25	28.80	34.51	32.51	23.71	21.84	22.14	26.14	Tasa 1	18.12.2014	10.63	11.43	6.62	7.79	15.71	11.01	1.65	Sai 1																			
7.4.2014	35.65	34.55	40.73	36.40	27.87	28.71	30.18	35.55	Tasa 1	19.12.2014	3.63	4.01	1.96	0.46	1.96	2.50	2.24	Sai 1																			
8.4.2014	41.30	39.79	51.10	47.15	30.67	30.54	32.79	37.05	Tasa 1	20.12.2014	10.20	9.75	8.50	7.97	1.46	2.38	2.38	Sai 1																			
10.4.2014	47.88	48.59	59.78	62.02	29.79	34.84	38.81	28.82	Tasa 1	22.12.2014	30.73	29.70	26.43	27.14	2.06	3.06	3.67	Sai 1																			
11.4.2014	7.79	7.46	7.66	7.04	9.57	8.80	3.72	4.26	Tasa 1	23.12.2014	30.85	29.57	27.04	28.15	18.76	23.07	13.27	Sai 1																			
12.4.2014	18.00	19.70	19.69	12.33	11.03	10.52	11.81	1.81	Tasa 1	24.12.2014	2.46	3.63	3.95	2.45	1.93	1.72	1.72	Sai 1																			
13.4.2014	8.45	9.12	6.52	5.79	7.34	7.10	6.15	6.83	Tasa 1	25.12.2014	4.00	3.97	2.54	1.37	4.63	5.58	2.80	Sai 1																			
14.4.2014	19.74	21.05	24.36	23.46	16.78	16.00	15.27	18.47	Tasa 1	1.1.2015	2.53	2.91	1.04	0.38	2.23	3.03	2.10	Sai 1																			
15.4.2014	15.22	14.46	27.03	20.52	20.60	15.17	15.10	20.83	Tasa 1	2.1.2015	1.55	1.70	-0.84	-0.71	1.65	1.65	1.65	Sai 1																			
16.4.2014	57.39	56.41	71.98	69.37	36.98	46.28	44.97	53.39	Tasa 1	3.1.2015	3.46	3.97	1.30	0.21	2.24	2.24	1.25	Sai 1																			
17.4.2014	40.45	40.42	46.75	49.97	28.02	33.89	38.81	38.81	Tasa 1	4.1.2015	9.35	9.25	9.30	8.14	2.97	2.11	5.97	Sai 1																			
18.4.2014	18.66	18.31	18.57	14.04	13.43	11.21	11.06	1.06	Tasa 1	5.1.2015	10.07	9.55	9.20	8.91	6.02	6.57	7.08	Sai 1																			
19.4.2014	25.29	25.09	27.31	27.93	20.07	19.84	16.22	20.31	Tasa 1	6.1.2015	49.06	40.07	32.29	32.62	27.90	25.65	24.24	Sai 1																			
20.4.2014	19.72	19.28	19.39	18.39	11.86	12.78	12.72	17.90	Tasa 1	7.1.2015	5.50	6.00	5.00	4.21	13.73	10.88	3.22	Sai 1																			
21.4.2014	29.68	30.03	35.77	34.03	28.19	26.64	28.11	29.38	Tasa 1	9.1.2015	6.91	5.67	4.51	2.91	3.39	4.60	2.87	Sai 1																			
22.4.2014	20.58	19.32	26.82	25.49	10.97	11.39	14.40	15.79	Tasa 1	10.1.2015	9.24	9.26	7.50	7.50	4.15	3.46	3.46	Sai 1																			
23.4.2014	31.63	31.04	42.03	39.66	25.46	26.29	25.35	31.87	Tasa 1	11.1.2015	13.29	16.10	13.49	15.46	13.85	15.77	16.67	Sai 1																			
24.4.2014	34.09	32.42	40.95	38.31	25.97	25.66	23.83	31.28	Tasa 1	12.1.2015	33.02	32.69	30.73	24.14	21.93	15.77	16.67	Sai 1																			
25.4.2014	24.49	24.47	25.89	24.52	17.32	18.85	19.33	21.60	Tasa 1	13.1.2015	26.97	26.38	24.25	24.28	22.72	13.24	12.92	Sai 1																			
26.4.2014	19.21	19.55	19.35	18.73	18.03	13.42	14.13	1.13	Tasa 1	14.1.2015	5.76	4.91	4.32	14.49	10.46	3.75	3.86	Sai 1																			
27.4.2014	27.18	27.24	33.38	31.01	22.76	25.08	24.19	28.71	Tasa 1	15.1.2015	5.27	3.54	2.58	3.22	3.57	4.52	4.22	Sai 1																			
28.4.2014	39.83	40.44	46.79	45.67	26.95	28.97	28.95	36.12	Tasa 1	16.1.2015	5.82	4.00	3.05	7.22	6.39	5.30	5.96	Sai 1																			
29.4.2014	17.76	17.88	23.88	21.47	13.01	13.04	13.07	17.52	Tasa 1	17.1.2015	8.80	7.12	5.99	7.50	7.85	8.06	8.84	Sai 1																			
30.4.2014	20.44	20.31	21.23	19.48	16.43	16.43	10.75	17.13	Tasa 1	18.1.2015	1.80	1.80	2.27	2.27	2.63	2.27	2.27	Sai 1																			
1.5.2014	10.88	11.40	10.91	10.54	10.79	10.75	10.78	8.39	Tasa 1	19.1.2015	29.37	28.67	10.09	10.77	0.78	4.74	5.16	Sai 1																			
2.5.2014	19.95	20.53	22.25	20.09	18.43	17.93	15.65	17.01	Tasa 1	20.1.2015	10.31	10.37	9.20	8.36	7.53	3.38	3.62	Sai 1																			
3.5.2014	4.27	4.01	2.99	1.87	1.66	1.91	2.72	3.07	Tasa 1	21.1.2015	6.93	5.75	5.08	5.71	6.08	3.41	1.42	Sai 1																			
4.5.2014	9.99	8.64	9.64	12.26	9.28	7.84	7.85	10.55	Tasa 1	22.1.2015	4.99	4.55	2.80	5.46	5.05	2.23	1.91	Sai 1																			
5.5.2014	16.01	15.79	15.37	15.73	13.12	9.71	13.76	13.73	Tasa 1	23.1.2015	9.27	10.25	15.20	16.28	15.56	5.09	5.33	Sai 1																			
6.5.2014	8.54	8.54	8.54	8.54	8.54	8.54	8.54	8.54	Tasa 1	24.1.2015	4.98	4.41	3.33	2.42	3.53	3.90	2.78	Sai 1																			
7.5.2014	6.15	5.14	5.64	4.52	3.81	4.33	5.89	5.89	Tasa 1	25.1.2015	4.01	2.41	2.08	4.14	3.23	3.56	3.56	Sai 1																			
8.5.2014	8.48	8.48	8.48	8.48	8.48	8.48	8.48	8.48	Tasa 1	26.1.2015	5.27	4.31	4.34	3.90	3.21	3.12	3.12	Sai 1																			
9.5.2014	33.15	35.46	35.46	33.71	26.91	27.66	35.45	35.45	Tasa 1	27.1.2015	9.14	7.31	5.14	4.34	3.90	3.55	3.77	Sai 1																			
7.6.2014	17.85	18.24	17.73	6.09	9.48	12.60	13.77	13.77	Tasa 1	28.1.2015	4.23	2.52	1.73	2.02	2.41	2.61	2.61	Sai 1																			
8.6.2014	12.67	12.75	14.50	11.63	11.87	11.23	7.59	9.04	Tasa 1	29.1.2015	8.65	7.01	5.48	7.66	5.98	3.91	4.02	S																			

## A23 Daily values of PM 10: Reference method, TEOM, SHARP, MP 101

Date	Leckel R1 (09/0081) PM10		Leckel R2 (11/0070) PM10		TEOM A-9 PM10		TEOM B-15 PM10		SHARP A-21 c- dust PM10		SHARP B-25 c- dust PM10		MP101 A-29 1 h mass conc. PM10		MP101 B-33 1 h mass conc. PM10		Campaign		Date	Leckel R1 (09/0081) PM10		Leckel R2 (11/0070) PM10		TEOM A-9 PM10		TEOM B-15 PM10		SHARP A-21 c- dust PM10		SHARP B-25 c- dust PM10		MP101 A-29 1 h mass conc. PM10		MP101 B-33 1 h mass conc. PM10		Campaign	
	Leckel R1 (09/0081) PM10	Leckel R2 (11/0070) PM10	Leckel R1 (09/0081) PM10	Leckel R2 (11/0070) PM10	TEOM A-9 PM10	TEOM B-15 PM10	SHARP A-21 c- dust PM10	SHARP B-25 c- dust PM10	MP101 A-29 1 h mass conc. PM10	MP101 B-33 1 h mass conc. PM10	MP101 A-29 1 h mass conc. PM10	MP101 B-33 1 h mass conc. PM10	Campaign	Leckel R1 (09/0081) PM10	Leckel R2 (11/0070) PM10	TEOM A-9 PM10	TEOM B-15 PM10	SHARP A-21 c- dust PM10	SHARP B-25 c- dust PM10	MP101 A-29 1 h mass conc. PM10	MP101 B-33 1 h mass conc. PM10	Campaign															
20.3.2014	19.65	18.99	21.01	21.12	15.39	14.73	18.79	Tasa 1	5.12.2014	3.09	2.95	4.64	4.30	3.95	4.05	Savi 1	21.3.2014	7.89	7.90	11.35	10.94	8.11	4.67	9.15	Tasa 1	6.12.2014	3.43	3.24	5.90	5.63	3.91	3.24	Savi 1				
21.3.2014	7.89	7.90	11.35	10.94	8.11	4.67	9.15	Tasa 1	6.12.2014	3.43	3.24	5.90	5.63	3.91	3.24	Savi 1	22.3.2014	33.63	34.82	50.92	49.35	24.84	24.95	33.96	36.27	Tasa 1	7.12.2014	4.90	4.90	7.61	7.20	4.06	3.99	Savi 1			
23.3.2014	16.92	17.34	25.69	24.97	13.63	11.08	15.86	Tasa 1	8.12.2014	5.24	5.16	8.87	8.40	5.18	7.03	Savi 1	24.3.2014	26.04	27.11	35.55	35.27	22.40	28.18	28.34	29.54	Tasa 1	9.12.2014	7.33	7.40	10.58	10.95	6.15	6.26	Savi 1			
25.3.2014	30.76	31.43	41.31	38.28	19.94	23.22	34.60	Tasa 1	10.12.2014	10.14	9.55	14.06	14.45	9.31	9.10	Savi 1	27.3.2014	31.04	29.63	39.87	38.04	21.00	23.07	32.94	35.83	Tasa 1	11.12.2014	9.58	9.06	11.90	12.11	10.21	9.93	Savi 1			
28.3.2014	49.38	51.79	61.51	57.91	30.40	36.60	48.67	Tasa 1	12.12.2014	10.36	10.25	11.52	11.80	10.26	10.80	Savi 1	29.3.2014	23.84	23.96	32.94	31.62	25.11	23.08	27.30	28.09	Tasa 1	13.12.2014	5.15	4.97	8.78	8.90	5.65	5.13	Savi 1			
30.3.2014	17.46	17.61	22.71	21.85	8.23	11.25	18.91	Tasa 1	14.12.2014	5.87	6.21	9.41	9.53	6.07	5.91	Savi 1	31.3.2014	32.40	31.19	41.89	39.59	21.00	36.03	37.57	37.57	Tasa 1	15.12.2014	6.46	6.47	9.14	9.14	7.07	6.29	Savi 1			
1.4.2014	33.47	31.42	45.68	44.01	35.19	37.86	Tasa 1	16.12.2014	6.33	6.38	9.41	9.57	7.52	6.09	Savi 1	2.4.2014	36.66	37.94	56.06	54.52	19.93	21.76	38.62	40.93	Tasa 1	17.12.2014	7.32	7.07	7.91	9.18	8.22	9.44	11.01	5.54	Savi 1		
3.4.2014	20.61	20.19	28.49	27.85	24.24	26.60	Tasa 1	30.12.2014	2.46	3.63	5.01	4.90	4.65	3.58	4.42	5.15	Savi 1	4.4.2014	37.80	37.49	47.48	45.97	33.54	30.59	44.91	45.13	Tasa 1	31.12.2014	4.00	3.97	6.43	6.37	5.22	4.58	3.81	Savi 1	
5.4.2014	18.62	18.27	25.49	26.42	18.98	16.27	20.14	Tasa 1	1.1.2015	2.53	2.91	3.99	3.96	3.39	3.68	1.89	Savi 1	6.4.2014	29.25	28.80	28.00	27.54	21.11	21.84	31.83	31.93	Tasa 1	2.1.2015	1.55	1.70	2.71	2.76	2.45	1.70	2.23	2.17	Savi 1
7.4.2014	35.65	34.55	46.69	43.07	30.36	29.30	41.12	Tasa 1	3.1.2015	3.46	3.97	4.27	3.52	2.74	3.00	8.00	Savi 1	8.4.2014	41.30	39.79	47.48	45.97	33.54	30.59	44.91	45.13	Tasa 1	4.1.2015	9.55	11.69	13.26	12.61	8.19	11.19	12.75	Savi 1	
9.4.2014	43.01	43.28	60.27	57.55	50.81	49.26	Tasa 1	5.1.2015	10.07	9.55	11.69	13.26	9.18	11.19	12.75	Savi 1	10.4.2014	47.88	48.59	77.75	75.71	39.27	33.46	50.98	50.45	Tasa 1	6.1.2015	43.06	40.07	35.13	35.07	23.52	20.70	42.57	39.04	Savi 1	
11.4.2014	7.79	7.46	11.81	11.22	13.49	10.49	9.52	Tasa 1	7.1.2015	5.50	6.00	7.36	7.29	10.67	8.66	8.66	Savi 1	12.4.2014	17.26	18.00	27.14	26.58	12.57	10.58	Tasa 1	9.1.2015	6.91	9.24	9.04	8.64	7.56	7.31	7.85	Savi 1			
13.4.2014	8.45	9.12	12.21	11.71	7.50	9.96	6.54	Tasa 1	10.1.2015	9.24	11.58	11.39	11.39	11.39	11.39	7.51	Savi 1	14.4.2014	19.74	21.05	31.61	31.81	15.74	13.74	23.76	21.64	Tasa 1	11.1.2015	13.29	20.13	19.57	11.04	12.62	17.59	18.73	Savi 1	
15.4.2014	15.22	14.46	17.59	20.98	10.21	11.21	21.93	Tasa 1	12.1.2015	33.02	35.46	34.89	21.87	21.50	34.45	35.08	Savi 1	16.4.2014	57.39	56.41	37.73	37.92	59.67	62.50	Tasa 1	13.1.2015	26.97	32.53	31.67	20.28	19.88	30.49	32.81	Savi 1			
17.4.2014	40.45	40.42	63.00	62.72	32.20	28.95	46.16	Tasa 1	14.1.2015	5.76	9.91	9.96	9.96	7.51	7.66	8.66	Savi 1	18.4.2014	19.22	18.66	24.64	23.52	15.45	15.27	20.52	18.47	Tasa 1	15.1.2015	5.27	7.45	7.28	6.00	5.29	6.43	4.87	Savi 1	
19.4.2014	25.29	25.09	35.61	35.23	18.89	18.69	28.46	Tasa 1	16.1.2015	5.82	8.24	8.24	8.24	8.24	8.24	8.24	Savi 1	20.4.2014	19.72	19.28	26.46	24.90	14.63	13.82	21.98	21.13	Tasa 1	17.1.2015	8.80	12.00	11.87	9.12	8.22	9.57	9.08	Savi 1	
21.4.2014	29.68	30.03	39.26	39.29	24.44	24.43	32.06	Tasa 1	18.1.2015	3.72	8.92	8.29	8.29	8.29	8.29	8.29	Savi 1	22.4.2014	20.58	20.97	28.96	14.83	14.18	27.30	25.64	Tasa 1	19.1.2015	29.37	33.12	30.56	19.43	21.21	26.34	23.87	Savi 1		
23.4.2014	31.63	31.04	44.66	41.71	22.32	23.82	38.85	Tasa 1	20.1.2015	5.77	5.07	5.07	5.07	5.07	5.07	5.07	Savi 1	24.4.2014	34.09	32.42	46.03	43.67	25.96	23.86	38.00	38.00	Tasa 1	21.1.2015	37.80	42.82	39.00	38.99	37.99	Savi 1			
25.4.2014	24.49	24.47	32.72	30.50	22.43	21.52	25.40	Tasa 1	22.1.2015	65.57	65.57	65.57	65.57	65.57	65.57	65.57	Savi 1	26.4.2014	19.21	19.65	24.37	23.66	20.79	20.68	18.50	20.48	Tasa 1	23.1.2015	23.68	28.64	26.81	21.79	23.26	25.01	Savi 1		
27.4.2014	27.18	27.24	34.95	34.24	18.87	19.64	29.28	Tasa 1	24.1.2015	9.14	12.31	12.90	12.90	12.90	12.90	12.90	Savi 1	28.4.2014	39.83	40.04	47.71	47.71	31.34	31.34	34.50	45.01	Tasa 1	25.1.2015	7.46	40.55	37.57	22.22	22.22	30.09	31.72	Savi 1	
29.4.2014	17.88	27.97	26.01	17.16	13.69	17.84	20.85	Tasa 2	2.2.2015	4.36	8.46	4.47	3.87	4.47	4.47	4.47	Savi 1	30.4.2014	20.44	20.31	25.79	25.48	17.98	17.34	22.60	21.97	Tasa 2	3.2.2015	13.04	13.04	12.72	12.04	14.16	15.37	Savi 1		
31.4.2014	18.99	18.51	28.54	26.75	14.43	14.67	19.78	Tasa 2	4.2.2015	12.83	12.83	12.83	12.83	12.83	12.83	12.83	Savi 1	32.4.2014	23.23	22.29	12.43	13.71	20.09	19.33	11.66	11.04	Tasa 2	5.2.2015	7.71	6.40	7.69	7.74	7.31	6.97	6.45	Savi 1	
3.5.2014	10.98	11.40	15.32	14.68	13.96	13.48	11.87	Tasa 2	6.2.2015	3.71	4.20	4.20	4.20	4.20	4.20	4.20	Savi 1	4.5.2014	15.04	14.96	20.46	19.51	12.46	12.00	15.32	13.55	Tasa 2	7.2.2015	3.63	4.40	4.96	3.25	3.43	4.56	4.88	Savi 1	
5.5.2014	14.96	14.76	17.74	17.04	11.52	10.15	16.75	Tasa 2	8.2.2015	10.31	10.31	10.31	10.31	10.31	10.31	10.31	Savi 1	6.5.2014	13.94	13.49	20.05	20.09	11.52	11.52	15.98	15.98	Tasa 2	9.2.2015	11.81	15.97	15.56	12.26	13.48	12.10	12.84	Savi 1	
7.5.2014	8.51	8.47	12.27	11.79	7.91	8.11	10.35	Tasa 2	10.2.2015	10.31	10.31	10.31	10.31	10.31	10.31	10.31	Savi 1	8.5.2014	13.24	13.04	22.78	22.5															

## A24 Daily values of PM 10: Reference method, Grimm 180

Date	Locket R1 (090081) PM10	Locket R2 (110070) PM10	Grimm 180 B-1 PM10	Grimm 180 C-1 PM10	Grimm 180 B-1 C average PM10	Tarea	Date	Locket R1 (090081) PM10	Locket R2 (110070) PM10	Grimm 180 B-1 PM10	Grimm 180 C-1 PM10	Grimm 180 B-1 C average PM10	Tarea
19.5.2014	48,11	46,54				Tarea 2	6.2.2015	3,62556666	1,9751125	2,185138889	2,08012595	Sav 2	
20.5.2014	37,66	35,46				Tarea 2	7.2.2015	3,62219993	2,476141667	2,85127778	2,663384723	Sav 2	
21.5.2014	31,58	30,14	38,62	25,50	32,06	Tarea 2	8.2.2015	38,05530372	17,23445831	31,49881944	24,36663264	Sav 2	
24.5.2014	37,02	35,13	31,46	28,30	29,88	Tarea 2	9.2.2015	15,20311933	9,112975	13,79951389	11,45124445	Sav 2	
25.5.2014	17,34	16,47	18,89	12,80	14,85	Tarea 2	10.2.2015	4,36343852	2,036470833	2,818686556	2,427575695	Sav 2	
26.5.2014	6,13		5,21	6,00	5,61	Tarea 2	11.2.2015	5,987011057	4,209991667	4,409513889	4,308252778	Sav 2	
27.5.2014	8,48		13,30	11,80	12,55	Tarea 2	12.2.2015	34,92384408	28,7231625	34,36201389	31,702582	Sav 2	
28.5.2014	8,03	9,11	11,04	9,80	10,42	Tarea 2	13.2.2015	6,17980411	87,29650411	90,05609444	89,10076881	Sav 2	
29.5.2014	7,34	8,02	9,90	8,70	9,30	Tarea 2	14.2.2015	4,233970525	3,000141667	3,191311	3,096633334	Sav 2	
30.5.2014	14,14	14,56	16,49	11,60	14,05	Tarea 2	15.2.2015	55,01178604	40,79784167	60,87515278	50,84254723	Sav 2	
31.5.2014	11,35	11,83	10,00	8,70	9,35	Tarea 2	16.2.2015	31,3218495	38,7169125	40,67986511	39,69239931	Sav 2	
1.6.2014	7,66	7,39	8,71	7,80	8,26	Tarea 2	17.2.2015	16,4606621	16,1079333	16,44511667	16,80116	Sav 2	
2.6.2014	9,28		11,03	10,10	10,57	Tarea 2	18.2.2015	6,151596243	4,25888333	4,899665	4,577254167	Sav 2	
3.6.2014	25,64		24,21	21,10	22,65	Tarea 2	19.2.2015	6,438358525	4,7644125	5,058661111	4,912136806	Sav 2	
4.6.2014	49,33	49,22	51,21	48,40	48,81	Tarea 2	20.2.2015	5,72703482	3,639491667	3,946486111	3,78548889	Sav 2	
5.6.2014	42,72	42,66	45,35	42,40	43,88	Tarea 2	21.2.2015	9,138556402	9,529166667	10,4311444	9,886180554	Sav 2	
6.6.2014	33,62	33,15	33,80	31,50	32,65	Tarea 2	22.2.2015	7,685433442	8,09572167	8,57708333	8,33790325	Sav 2	
7.6.2014	17,65	17,33	19,32	15,50	17,41	Tarea 2	23.2.2015	8,456266694	7,78252167	8,68958333	8,23991875	Sav 2	
8.6.2014	12,67	12,75	15,15	12,30	13,73	Tarea 2	24.2.2015	9,88304628	10,4248975	11,4273167	10,93119359	Sav 2	
9.6.2014	6,69	6,66	8,28	6,80	7,54	Tarea 2	25.2.2015	14,70982926	14,1487167	15,16666667	14,83227217	Sav 2	
10.6.2014	14,14		13,40	13,20	13,30	Tarea 2	5.3.2015	5,081515287	3,347970833	3,632669444	3,490270139	Sav 2	
11.6.2014	20,26	20,60				Tarea 2	6.3.2015	4,03995621	2,674866667	2,904513889	2,789600278	Sav 2	
12.6.2014	16,51	16,14				Tarea 2	7.3.2015	6,3320546	5,588991667	6,225446667	5,907664167	Sav 2	
13.6.2014	8,26	9,98				Tarea 2	8.3.2015	6,308437496	5,28935167	5,716666667	5,50310417	Sav 2	
14.6.2014	5,76					Tarea 2	9.3.2015	80,48413418	99,5643167	105,759306	102,7672861	Sav 2	
15.6.2014	8,81					Tarea 2	10.3.2015	13,82731725	14,38323333	16,13435556	15,25376745	Sav 2	
16.6.2014	4,65	4,71				Tarea 2	11.3.2015	89,5428986	34,03358333	61,06680556	47,55018195	Sav 2	
18.6.2014	5,74	5,54	5,49	4,60	5,04	Tarea 2	12.3.2015	196,1432467	189,1879875	193,0227778	191,105327	Sav 2	
19.6.2014	6,09	6,03	5,99	4,80	5,39	Tarea 2	13.3.2015	274,4406151	360,2203958	325,6	342,1010797	Sav 2	
20.6.2014	4,89	4,91	4,04	3,50	3,77	Tarea 2	14.3.2015	10,3889447	136,9015459	122,7932917	129,4469188	Sav 2	
21.6.2014	5,17	5,20	3,20	2,70	2,95	Tarea 2	15.3.2015	274,8295248	216,8323542	191,6322917	204,232323	Sav 2	
22.6.2014	7,58	7,62	6,17	5,10	5,64	Tarea 2	16.3.2015	290,8193326	292,7568208	237,0675694	264,9121951	Sav 2	
23.6.2014	8,47		8,69	8,59	7,20	Tarea 2	17.3.2015	24,826240404	249,0509833	199,7097222	224,380328	Sav 2	
24.6.2014	10,31		10,42	13,15	10,60	Tarea 2	18.3.2015	150,617484	192,4110262	148,1664167	170,5382396	Sav 2	
25.6.2014	15,76	15,89	17,52	13,90	16,71	Tarea 2	19.3.2015	21,44255165	30,1848167	32,1152778	31,1500073	Sav 2	
26.6.2014	15,00	15,30	16,67	13,60	15,13	Tarea 2	20.3.2015	26,8612874	34,7117375	34,84641667	34,78795209	Sav 2	
27.6.2014	14,87	14,46	11,54	9,70	10,62	Tarea 2	21.3.2015	27,72620127	33,775025	35,573761111	34,66619306	Sav 2	
28.6.2014	10,25	10,42	8,03	7,10	7,57	Tarea 2	22.3.2015	20,24202321	34,95684417	40,1972222	37,5882632	Sav 2	
29.6.2014	8,72	8,79	7,76	6,80	7,28	Tarea 2	26.3.2015	39,7815446	52,2332167	48,747945056	50,5540812	Sav 2	
30.6.2014	7,24	7,44	9,03	8,80	8,92	Tarea 2	27.3.2015	26,20590118	31,45097917	36,2950833	26,17797917	Sav 2	
1.7.2014	6,14	6,19	7,54	6,20	6,87	Tarea 2	28.3.2015	24,09758752	26,17197917	28,2770333	27,72284375	Sav 2	
2.7.2014	19,29	19,33	12,45	11,70	12,07	Tarea 2	29.3.2015	24,5827289	24,6819533	28,7451389	26,72326211	Sav 2	
3.7.2014	14,52	14,75	13,29	11,80	12,55	Tarea 2	30.3.2015	12,91530649	12,4150625	14,016527778	13,21579514	Sav 2	
4.7.2014	11,79	11,93	13,19	11,40	12,30	Tarea 2	31.3.2015	4,215352619	2,050483333	2,302222222	2,176352778	Sav 2	
5.7.2014	10,51	10,51	10,38	9,00	9,69	Tarea 2	1.4.2015	25,30173976	22,3232	24,1672222	23,24252811	Sav 2	
6.7.2014	15,85	16,74	13,95	11,50	12,72	Tarea 2	2.4.2015	6,422742111	6,002170833	6,570678899	6,286467361	Sav 2	
7.7.2014	21,03	21,33	18,30	15,90	17,10	Tarea 2	3.4.2015	14,90750816	19,401557	10,347075	13,70556556	Sav 2	
8.7.2014	16,66		16,25	14,30	15,68	Tarea 2	4.4.2015	49,63916691	49,49333333	49,4985167	49,2419375	Sav 2	
9.7.2014	11,76		11,53	10,10	10,82	Tarea 2	5.4.2015	59,25761421	65,56074583	65,68451389	65,71262996	Sav 2	
10.7.2014	12,80		10,70	9,30	10,00	Tarea 2	6.4.2015	63,73345432	58,1777333	63,83284722	60,36015268	Sav 2	
11.7.2014	11,28	11,43	8,78	7,40	8,09	Tarea 2	7.4.2015	86,38012323	105,391042	99,16997222	102,290032	Sav 2	
12.7.2014	12,92	13,13	9,52	7,90	8,71	Tarea 2	8.4.2015	56,748719934	78,719299568	88,4487167	83,33049366	Sav 2	
13.7.2014	17,43	17,90	15,79	13,60	14,70	Tarea 2	10.4.2015	20,3795796	24,1045125	26,15100025	25,15100025	Sav 3	
14.7.2014	26,26	25,31	24,64	21,80	23,22	Tarea 2	11.4.2015	9,02357208	81,2319667	84,05	82,64098334	Sav 3	
5.12.2014	3,09	2,95	3,59	3,73	3,66	Sav 1	12.4.2015	26,28176461	27,8338875	27,36	27,59949375	Sav 3	
6.12.2014	3,43	3,24	3,64	3,91	3,78	Sav 1	13.4.2015	16,83079434	18,0457833	18,0457833	19,51286417	Sav 3	
7.12.2014	4,90	4,90	4,41	4,67	4,54	Sav 1	14.4.2015	1,42113851	18,4899217	19,3	19,07444649	Sav 3	
8.12.2014	5,24	5,16	4,19	4,45	4,32	Sav 1	15.4.2015	3,990168692	2,082679167	2,34	2,11339584	Sav 3	
9.12.2014	7,33	7,40	6,24	5,43	5,33	Sav 1	16.4.2015	11,5132032	7,0879176	10,64	8,86408584	Sav 3	
10.12.2014	9,55	9,55	8,68	9,80	9,24	Sav 1	17.4.2015	49,64823209	40,9360833	48,87	49,0515147	Sav 3	
11.12.2014	9,68		9,93	10,40	10,16	Sav 1	18.4.2015	4,42020312	38,07782083	43,14	41,05891042	Sav 3	
12.12.2014	10,26	10,25	11,81	12,39	12,10	Sav 1	19.4.2015	25,19314472	26,191375	26,22	26,20566875	Sav 3	
13.12.2014	5,15	4,97	6,17	6,50	6,34	Sav 1	20.4.2015	68,7190417	67,69129167	84,98	75,33564584	Sav 3	
14.12.2014	5,87	6,21	5,74	5,96	5,85	Sav 1	21.4.2015	75,1361741	84,231967	81,49	82,864564	Sav 3	
15.12.2014	6,46	6,47	7,10	7,70	7,40	Sav 1	22.4.2015	15,72852483	11,205205417	11,54	11,7302709	Sav 3	
16.12.2014	6,33	6,38	6,27	6,67	6,47	Sav 1	23.4.2015	9,759567102	9,251883333	10,23	9,740941667	Sav 3	
17.12.2014	7,32	7,07	6,97	7,33	7,33	Sav 1	24.4.2015	1,85791041	1,36995333	1,56	1,464979167	Sav 3	
18.12.2014	10,63	11,43	4,14	4,51	4,33								

## A25 Daily values of PM<sub>2.5</sub>: Reference method, TEOM, SHARP c and MP 101

Date	Leckel R1 (000001 PM2.5 PPM)	Leckel R2 (110001 PM2.5 PPM)	TEOM B-15 (000001 PM2.5 PPM)	SHARP A-21 (000001 PM2.5 PPM)	SHARP B-25 c (000001 PM2.5 PPM)	MP101-A29 (000001 PM2.5 PPM)	MP101-B33 (000001 PM2.5 PPM)	Campaign	Date	Leckel R1 (000001 PM2.5 PPM)	Leckel R2 (110001 PM2.5 PPM)	TEOM A-9 (000001 PM2.5 PPM)	SHARP A-21 (000001 PM2.5 PPM)	SHARP B-25 c (000001 PM2.5 PPM)	MP101-A29 (000001 PM2.5 PPM)	MP101-B33 (000001 PM2.5 PPM)	Campaign		
17.7.2014	8.49	9.80	9.80	9.47	9.03	10.96	—	Tess 1	10.4.2015	—	—	5.73	4.05	1.86	4.98	—			
18.7.2014	9.03	9.54	9.93	9.79	11.01	11.32	10.70	16.16	11.42	11.54	12.04	10.65	12.93	16.61	19.25	Sess 1			
19.7.2014	8.30	8.46	8.85	8.57	9.73	9.85	8.18	12.75	Tess 1	12.4.2015	—	6.30	7.44	7.85	6.83	6.50	Tess 1		
20.7.2014	6.74	6.84	6.95	7.02	6.31	6.36	5.90	9.23	Tess 1	13.4.2015	—	2.55	4.53	4.46	2.90	0.53	Tess 1		
21.7.2014	6.3	6.77	7.71	7.73	7.53	7.38	8.26	10.75	Tess 1	14.4.2015	—	2.95	4.71	4.82	3.10	2.55	Tess 1		
22.7.2014	9.52	9.88	9.82	9.93	9.73	10.13	8.37	11.50	Tess 1	15.4.2015	—	1.76	3.51	3.55	2.67	1.06	Tess 1		
23.7.2014	10.89	11.02	11.76	11.45	11.04	12.80	14.80	Tess 1	16.4.2015	—	2.67	6.98	6.80	3.99	2.49	Sess 1			
24.7.2014	10.84	13.10	13.75	13.45	13.79	12.79	17.22	Tess 1	17.4.2015	—	1.54	5.69	7.90	2.30	23.11	14.65	Sess 1		
25.7.2014	10.86	10.84	10.61	10.42	11.47	10.80	11.85	15.79	Tess 1	18.4.2015	—	7.37	5.63	8.73	7.59	2.21	20.19	13.65	Sess 1
26.7.2014	13.20	—	11.29	11.41	13.08	14.71	—	—	Tess 1	19.4.2015	—	4.76	5.04	5.16	2.83	0.94	5.79	6.59	Sess 1
27.7.2014	12.70	—	—	—	—	—	—	—	Tess 1	20.4.2015	—	12.53	7.97	8.2	7.95	5.93	12.35	14.29	Sess 1
28.7.2014	14.25	—	—	—	—	—	—	—	Tess 1	21.4.2015	—	9.70	5.76	8.36	7.38	6.38	11.54	11.29	Sess 1
29.7.2014	6.5	8.51	7.73	7.61	23.95	11.50	10.85	12.20	Tess 1	23.4.2015	—	2.81	5.76	5.93	2.74	1.58	4.80	4.58	Sess 1
30.7.2014	9.93	9.95	10.68	10.42	9.03	8.94	11.94	13.30	Tess 1	24.4.2015	—	7.23	8.68	8.08	7.12	8.20	9.09	10.93	Sess 1
31.7.2014	15.87	20.22	20.55	14.25	14.26	6.55	13.96	Tess 1	25.4.2015	—	1.52	8.90	9.08	7.28	2.01	10.24	11.28	Sess 1	
1.8.2014	6.82	6.81	6.79	6.57	6.13	6.19	6.12	11.97	Tess 1	26.4.2015	—	2.99	5.57	5.63	3.13	3.11	3.19	3.11	Sess 1
2.8.2014	7.23	6.65	6.28	7.08	6.96	7.20	2.13	12.07	Tess 1	27.4.2015	—	4.91	6.80	7.13	4.73	6.06	3.37	4.92	Sess 1
3.8.2014	11.78	11.50	13.18	12.83	11.22	12.18	12.38	15.85	Tess 1	28.4.2015	—	10.02	7.04	7.09	6.96	5.37	10.64	12.02	Sess 1
4.8.2014	10.47	11.12	10.87	10.96	11.07	10.20	11.48	13.86	Tess 1	29.4.2015	—	5.27	6.63	6.50	4.82	5.28	4.57	6.21	Sess 1
5.8.2014	11.76	13.08	13.02	12.79	10.82	12.13	12.40	15.99	Tess 1	30.4.2015	—	6.26	6.93	7.01	4.45	4.88	6.38	6.30	Sess 1
6.8.2014	15.5	15.94	15.42	15.14	12.23	15.67	16.28	19.34	Tess 1	1.5.2015	—	3.68	5.01	5.11	2.53	3.87	3.25	3.15	Sess 1
7.8.2014	14.45	14.44	14.17	14.48	14.80	14.93	18.82	Tess 1	2.5.2015	—	1.39	3.73	3.72	2.61	0.89	3.03	2.53	Sess 1	
8.8.2014	9.63	9.74	9.45	9.23	10.72	10.91	10.06	13.05	Tess 1	3.5.2015	—	5.74	7.23	7.15	5.17	4.75	8.03	6.62	Sess 1
9.8.2014	7.27	7.43	7.90	7.33	8.29	8.18	8.66	12.15	Tess 1	4.5.2015	—	7.17	8.08	8.76	5.43	6.15	10.71	11.86	Sess 1
10.8.2014	15.28	15.39	16.95	15.96	8.65	9.42	9.82	12.87	Tess 1	5.5.2015	—	4.08	5.78	5.89	3.89	3.62	4.47	4.47	Sess 1
11.8.2014	14.04	14.99	15.49	15.86	13.89	12.46	16.91	21.98	Tess 1	6.5.2015	—	9.44	12.03	12.55	11.20	10.79	9.59	11.28	Sess 1
12.8.2014	6.95	—	—	—	—	—	—	—	Tess 1	7.5.2015	—	6.70	8.46	7.19	8.01	7.47	8.09	Sess 1	
13.8.2014	6.52	6.8	7.84	7.34	6.96	6.80	9.62	12.49	Tess 1	8.5.2015	—	4.56	5.24	5.37	3.89	4.09	5.30	6.08	Sess 1
14.8.2014	8.16	9.50	9.20	8.80	7.07	8.42	9.33	Tess 1	9.5.2015	—	2.23	4.36	4.47	0.97	1.13	3.57	5.29	Sess 1	
15.8.2014	5.81	6.21	6.40	6.06	6.14	7.39	9.16	Tess 1	10.5.2015	—	3.91	6.47	6.46	1.30	4.22	3.58	5.85	Sess 1	
16.8.2014	3.41	3.43	4.70	4.64	3.48	4.11	4.59	6.25	Tess 1	11.5.2015	—	3.60	4.34	4.55	2.88	3.89	—	—	Sess 1
17.8.2014	4.02	4.6	5.24	5.26	3.56	4.45	5.58	6.72	Tess 1	12.5.2015	—	5.38	6.59	6.61	6.04	6.90	5.69	5.59	Sess 1
18.8.2014	6.79	6.88	6.13	6.51	6.51	7.22	9.50	Tess 1	13.5.2015	—	2.30	3.27	3.30	3.06	3.61	3.61	3.61	Sess 1	
19.8.2014	4.42	4.58	4.86	4.48	5.42	4.13	5.31	7.77	Tess 1	14.5.2015	—	3.07	4.08	3.93	1.28	3.04	3.67	4.21	Sess 1
20.8.2014	6.19	6.35	6.98	6.72	6.50	5.74	7.41	10.11	Tess 1	15.5.2015	—	3.77	6.07	8.10	3.03	4.09	6.51	5.71	Sess 1
21.8.2014	5.33	5.42	6.10	7.85	5.53	4.95	7.38	7.69	Tess 1	16.5.2015	—	2.90	4.50	4.59	1.76	2.35	3.66	3.66	Sess 1
22.8.2014	5.71	8.47	6.23	6.17	5.57	7.25	7.89	Tess 1	17.5.2015	—	2.38	4.01	4.06	0.98	2.56	3.42	2.51	Sess 1	
23.8.2014	6.44	6.64	9.04	8.72	6.88	8.49	7.70	8.33	Tess 1	18.5.2015	—	4.66	7.62	7.56	5.57	5.82	7.27	7.27	Sess 1
24.8.2014	5.38	5.49	7.54	7.36	5.85	5.80	6.83	8.17	Tess 1	19.5.2015	—	4.90	7.22	7.2	4.93	5.05	7.89	8.47	Sess 1
25.8.2014	6.84	6.99	7.46	7.22	7.75	7.73	9.29	11.11	Tess 1	20.5.2015	—	5.54	9.03	9.12	5.09	7.23	9.01	10.87	Sess 1
26.8.2014	6.14	6.9	7.11	7.31	4.93	4.96	3.82	4.08	Tess 1	21.5.2015	—	3.65	4.74	4.92	0.47	4.54	5.85	6.40	Sess 1
27.8.2014	3.19	3.39	5.69	5.83	5.18	5.67	5.51	6.77	Tess 1	22.5.2015	—	5.71	5.49	5.47	1.77	3.29	5.49	5.49	Sess 1
28.8.2014	4.03	4.16	6.18	6.16	4.94	4.90	5.05	7.80	Tess 1	23.5.2015	—	2.70	3.72	3.76	0.70	1.92	4.37	4.23	Sess 1
29.8.2014	4.82	5.05	7.27	7.27	6.99	4.96	7.70	8.37	Tess 1	24.5.2015	—	3.54	5.52	5.60	0.66	3.68	4.61	5.20	Sess 1
30.8.2014	4.39	4.39	7.18	7.18	6.90	4.59	4.98	5.62	Tess 1	25.5.2015	—	6.49	7.93	8.23	2.22	3.23	7.12	8.00	Sess 1
31.8.2014	5.57	5.68	7.84	7.43	5.04	5.08	6.74	7.84	Tess 1	26.5.2015	—	6.14	8.14	8.23	-0.35	6.14	5.19	7.75	Sess 1
1.9.2014	7.27	7.52	8.89	9.92	7.10	7.43	8.48	10.23	Tess 1	27.5.2015	—	5.39	6.92	6.53	0.15	5.16	6.74	6.44	Sess 1
2.9.2014	5.25	8.45	8.13	8.13	5.93	5.20	6.59	7.72	Tess 1	28.5.2015	—	4.71	4.78	4.97	-0.56	5.30	2.84	3.05	Sess 2
3.9.2014	4.16	4.21	4.89	4.94	5.04	5.21	5.03	5.61	Tess 1	29.5.2015	—	5.39	5.71	5.86	1.16	3.49	5.65	6.88	Sess 2
4.9.2014	5.27	5.62	6.29	6.44	5.42	5.42	5.04	6.04	Tess 1	30.5.2015	—	5.09	5.47	5.47	0.05	4.44	5.23	5.23	Sess 2
5.9.2014	7.49	8.42	9.08	—	8.41	11.08	—	—	Tess 1	31.5.2015	—	4.49	5.17	5.23	0.38	4.54	5.29	5.29	Sess 2
6.9.2014	7.44	7.77	10.05	9.59	10.46	11.50	11.03	Tess 1	1.6.2015	—	7.03	7.63	7.22	0.05	6.64	6.95	8.65	Sess 2	
7.9.2014	7.94	8.41	10.01	9.78	9.47	9.89	10.20	Tess 1	2.6.2015	—	5.82	6.21	6.21	0.51	5.95	7.08	7.08	Sess 2	
8.9.2014	11.28	12.04	13.45	13.35	12.86	13.03	12.98	Tess 1	3.6.2015	—	2.79	5.08	4.98	2.02	3.06	5.29	4.03	Sess 2	
9.9.2014	6.04	6.33	6.45	6.55	6.23	6.23	7.02	Tess 1	4.6.2015	—	3.16	4.50	4.57	0.08	1.85	6.62	5.67	Sess 2	
10.9.2014	1																		

## A26 Daily values of PM<sub>2.5</sub>: Reference method, Grimm 180

Date	Lockel R1 (09/0081)		Lockel R2 (11/0070)		Grimm 180 B-4		Grimm 180 C-2		Campaign	Date	Lockel R1 (09/0081)		Lockel R2 (11/0070)		Grimm 180 B-4		Grimm 180 C-2		Campaign
	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>			PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>	PM <sub>2.5</sub>		
17.7.2014	8,49	8,46	9,18	8,23	Tasa 1	22.2.2015	6,21	7,48	7,95	Sav1 2-2									
18.7.2014	9,03	9,54	11,41	10,25	Tasa 1	23.2.2015	6,14	7,10	7,69	Sav1 2-2									
19.7.2014	8,30	8,46	9,25	8,19	Tasa 1	24.2.2015	7,03	8,93	9,69	Sav1 2-2									
20.7.2014	6,74	6,84	7,44	6,69	Tasa 1	25.2.2015	8,55	12,37	13,26	Sav1 2-2									
21.7.2014	6,32	6,77	7,88	7,15	Tasa 1	26.2.2015	8,19	10,51	11,26	Sav1 2-2									
22.7.2014	9,52	9,88	10,87	9,82	Tasa 1	27.2.2015	9,15	12,32	13,10	Sav1 2-2									
23.7.2014	10,66	11,02	13,45	12,33	Tasa 1	28.2.2015	9,92	13,35	14,14	Sav1 2-2									
24.7.2014	12,64	13,10	14,87	12,92	Tasa 1	1.3.2015	10,90	13,00	13,82	Sav1 2-2									
25.7.2014	10,56	10,84	12,1	10,95	Tasa 1	2.3.2015	9,74	11,64	12,32	Sav1 2-2									
26.7.2014	13,20	15,13	13,68	Tasa 1	3.3.2015	4,04	5,83	6,25	Sav1 2-2										
27.7.2014	12,70	12,48	11,64	Tasa 1	4.3.2015	2,04	2,41	2,60	Sav1 2-2										
28.7.2014	14,25	16,27	14,89	Tasa 1	5.3.2015	2,56	2,81	3,03	Sav1 2-2										
29.7.2014	8,38	8,51	10,31	9,33	Tasa 1	6.3.2015	1,97	2,21	2,39	Sav1 2-2									
30.7.2014	9,93	9,95	12,35	11,21	Tasa 1	7.3.2015	3,75	5,35	5,89	Sav1 2-2									
31.7.2014	15,07				Tasa 1	8.3.2015	3,22	3,60	3,87	Sav1 2-2									
1.8.2014	8,49	8,02	8,09	7,37	Tasa 1	9.3.2015	6,66	12,13	13,07	Sav1 2-2									
2.8.2014	7,22	6,65	6,83	6,38	Tasa 1	10.3.2015	3,33	4,38	4,82	Sav1 2-2									
3.8.2014	11,78	11,50	12,72	11,72	Tasa 1	11.3.2015	8,67	16,63	14,07	Sav1 2-2									
4.8.2014	10,47	11,12	9,89	9,09	Tasa 1	12.3.2015	19,01	32,93	34,78	Sav1 2-2									
5.8.2014	11,76	13,08	12,78	11,95	Tasa 1	13.3.2015	34,83	49,33	51,88	Sav1 2-2									
6.8.2014	15,26	15,94	15,83	14,66	Tasa 1	14.3.2015	15,10	24,02	25,16	Sav1 2-2									
7.8.2014	14,45	15,17	14,28	13,01	Tasa 1	15.3.2015	32,20	54,98	56,64	Sav1 2-2									
8.8.2014	9,63	9,74			Tasa 1	16.3.2015	46,59	69,62	60,00	Sav1 2-2									
9.8.2014	7,27	7,43			Tasa 1	17.3.2015	46,93	53,31	53,72	Sav1 2-2									
10.8.2014	10,28	10,59			Tasa 1	18.3.2015	23,41	39,54	40,51	Sav1 2-2									
11.8.2014	14,04				Tasa 1	19.3.2015	9,50	9,93	11,39	Sav1 2-2									
12.8.2014	6,95		7,22	6,72	Tasa 1	20.3.2015	5,77	4,42	4,68	Sav1 2-2									
13.8.2014	6,52	6,65	5,62	5,33	Tasa 1	21.3.2015	8,36	6,10	6,34	Sav1 2-2									
14.8.2014	8,18	8,16	7,6	7,27	Tasa 1	22.3.2015	15,39	8,43	10,28	Sav1 2-2									
15.8.2014	5,82	5,71	5,39	4,72	Tasa 1	26.3.2015	12,99	9,98	9,95	Sav1 2-2									
16.8.2014	3,41	3,43	3,74	3,36	Tasa 1	27.3.2015	19,98	13,33	14,72	Sav1 2-2									
17.8.2014	4,02	4,03	4,59	4,29	Tasa 1	28.3.2015	18,09	24,52	26,27	Sav1 2-2									
18.8.2014	6,79	6,88	7,01	6,49	Tasa 1	29.3.2015	8,55	22,15	24,12	Sav1 2-2									
19.8.2014	4,92	4,93	5,13	4,72	Tasa 1	30.3.2015	2,39	10,41	11,61	Sav1 2-2									
20.8.2014	6,19	6,35	5,94	5,61	Tasa 1	31.3.2015	0,24	1,74	1,89	Sav1 2-2									
21.8.2014	5,33	5,42	5,35	5,06	Tasa 1	1.4.2015	7,02	10,65	11,69	Sav1 2-2									
22.8.2014	5,71	6	5,58	Tasa 1	2.4.2015	4,38	5,32	5,79	Sav1 2-2										
23.8.2014	6,44	6,64	7,96	7,33	Tasa 1	3.4.2015	4,69	7,79	7,83	Sav1 2-2									
24.8.2014	5,38	5,49	6,71	6,14	Tasa 1	4.4.2015	9,88	14,14	15,28	Sav1 2-2									
25.8.2014	6,84	6,99	9,95	8,85	Tasa 1	5.4.2015	8,84	15,91	16,73	Sav1 2-2									
26.8.2014	4,17	7,38	6,63	Tasa 1	6.4.2015	7,46	12,80	13,82	Sav1 2-2										
27.8.2014	3,10	3,20	5,14	4,64	Tasa 1	7.4.2015	9,50	14,84	16,02	Sav1 2-2									
28.8.2014	4,03	4,16	6,39	5,56	Tasa 1	8.4.2015	12,89	11,25	13,41	Sav1 2-2									
29.8.2014	4,82	5,06	8,43	7,52	Tasa 1	10.4.2015	3,23	4,03	4,54	Sav1 1-3									
30.8.2014	4,39	4,62	7,26	6,52	Tasa 1	11.4.2015	11,42	19,54	20,67	Sav1 1-3									
31.8.2014	5,57	5,56	8,08	7,42	Tasa 1	12.4.2015	6,30	9,62	10,32	Sav1 1-3									
1.9.2014	7,28	7,57	9,42	8,35	Tasa 1	13.4.2015	2,55	3,45	3,79	Sav1 1-3									
2.9.2014	7,95	8,11	10,02	9,05	Tasa 1	14.4.2015	2,95	4,93	5,34	Sav1 1-3									
3.9.2014	6,65	7,15		7,04	Tasa 1	15.4.2015	1,76	1,93	2,11	Sav1 1-3									
4.9.2014	8,55	9,01		8,25	Tasa 1	16.4.2015	2,87	3,74	4,02	Sav1 1-3									
5.9.2014	7,59	7,84		8,96	Tasa 1	17.4.2015	9,04	13,14	13,68	Sav1 1-3									
6.9.2014	7,44	7,77			Tasa 1	18.4.2015	2,81	3,05	3,16	Sav1 1-3									
7.9.2014	7,94	8,41			Tasa 1	19.4.2015	4,76	4,78	5,01	Sav1 1-3									
8.9.2014	11,28	12,04			Tasa 1	20.4.2015	12,53	12,70	13,91	Sav1 1-3									
24.9.2014	5,68	5,73			Tasa 1	21.4.2015	9,70	14,53	14,76	Sav1 1-3									
25.9.2014	9,23	9,35			Tasa 1	23.4.2015	2,81	3,05	3,16	Sav1 1-3									
26.9.2014	7,72	7,52			Tasa 1	24.4.2015	7,23	8,52	9,03	Sav1 1-3									
27.9.2014	5,23	5,25			Tasa 1	25.4.2015	7,52	10,81	11,40	Sav1 1-3									
28.9.2014	4,29	4,28			Tasa 1	26.4.2015	2,59	2,96	3,28	Sav1 1-3									
29.9.2014	5,20	5,08			Tasa 1	27.4.2015	4,91	7,29	7,80	Sav1 1-3									
30.9.2014	4,12	4,17			Tasa 1	28.4.2015	10,02	9,15	9,85	Sav1 1-3									
1.10.2014	6,04	6,26			Tasa 1	29.4.2015	5,27	4,85	5,57	Sav1 1-3									
9.1.2015	5,82	6,54	7,06	Sav1 1-1	30.4.2015	6,26	9,36	9,94	Sav1 1-3										
10.1.2015	5,29	6,77	7,18	Sav1 1-1	1.5.2015	3,68	4,25	4,72	Sav1 1-3										
12.1.2015	12,58	14,28	14,99	Sav1 1-1	2.5.2015	1,39	1,26	1,41	Sav1 1-3										
13.1.2015	9,50	12,23	12,97	Sav1 1-1	3.5.2015	5,74	6,17	6,76	Sav1 1-3										
14.1.2015	4,13	4,74	5,09	Sav1 1-1	4.5.2015	7,12	8,56	9,24	Sav1 1-3										
15.1.2015	3,19	3,34	3,62	Sav1 1-1	5.5.2015	4,08	3,95	4,48	Sav1 1-3										
16.1.2015	2,89	3,87	4,12	Sav1 1-1	6.5.2015	9,44	10,35	11,64	Sav1 1-3										
17.1.2015	4,37	6,39	6,60	Sav1 1-1	7.5.2015	6,70	8,79	9,78	Sav1 1-3										
18.1.2015	2,26	2,87	3,05	Sav1 1-1	8.5.2015	4,56	5,65	6,29	Sav1 1-3										
19.1.2015	8,86	13,63	11,45	Sav1 1-1	9.5.2015	2,23	3,26	2,57	Sav1 1-3										
20.1.2015	17,63	23,14	17,30	Sav1 1-1	10.5.2015	3,91	3,69	4,25	Sav1 1-3										
21.1.2015	8,65	12,25	12,23	Sav1 1-1	11.5.2015	3,60	3,36	3,79	Sav1 1-3										
22.1.2015	12,47	16,68	15,22	Sav1 1-1	12.5.2015	5,38	6,08	6,78	Sav1 1-3										
23.1.2015	9,01	10,23	10,83	Sav1 1-1	13.5.2015	2,20	2,21	2,44	Sav1 1-3										
24.1.2015	5,73	7,11	7,71	Sav1 1-1	14.5.2015	2,07	1,99	2,22	Sav1 1-3										
25.1.2015	6,59	8,25	8,84	Sav1 1-1	15.5.2015	3,77	3,64	4,05	Sav1 1-3										
26.1.2015	3,98	4,33	4,72	Sav1 1-1	16.5.2015	2,90	2,33	2,65	Sav1 1-3										
27.1.2015	2,83	3,35	3,65	Sav1 1-1	17.5.2015	2,38	2,15	2,42	Sav1 1-3										
28.1.2015	1,73	2,50	2,72	Sav1 1-1	18.5.2015	4,66	6,48	6,93	Sav1 1-3										
29.1.2015	10,08	13,81	14,63	Sav1 1-1	19.5.2015	4,90	7,24	7,99	Sav1 1-3										
30.1.2015	7,96	12,41	13,27	Sav1 1-1	20.5.2015	6,54	7,89	8,72	Sav1 1-3										
31.1.2015	7,18	7,22	7,87	Sav1 1-1	21.5.2015	3,85	3,32	3,67	Sav1 1-3										
1.2.2015	4,98	4,94	5,32	Sav1 1-1															
2.2.2015	6,39	7,96	8,59	Sav1 1-1															
3.2.2015	11,11	15,41	16,29	Sav1 1-1															
4.2.2015	8,51	11,34	12,17	Sav1 1-1															
5.2.2015	7,48	6,92	7,47	Sav1 1-1															
6.2.2015	2,34	1,81	1,97	Sav1 1-1															
7.2.2015	1,27	2,03	2,27	Sav1 1-1															
8.2.2015	5,30	7,90	7,74	Sav1 1-1															
9.2.2015	3,18	4,33	4,75	Sav1 1-1															
10.2.2015	1,53	1,42	1,54	Sav1 1-1															

## A27 Daily values of PM<sub>2.5</sub>: Reference method, BAM 1020 and FH-62IR

Locket R1 (09/0081) PM <sub>2.5</sub>	Locket R2 (11/0070) PM <sub>2.5</sub>	BAM A-21 conc. PM <sub>2.5</sub>	BAM B-25 conc. PM <sub>2.5</sub>	FH-62 A-29 conc. PM <sub>2.5</sub>	FH-62 B-33 conc. PM <sub>2.5</sub>	Campaign	Date	Locket R1 (09/0081) PM <sub>2.5</sub>	Locket R2 (11/0070) PM <sub>2.5</sub>	BAM A-21 conc. PM <sub>2.5</sub>	BAM B-25 conc. PM <sub>2.5</sub>	FH-62 A-29 conc. PM <sub>2.5</sub>	FH-62 B-33 conc. PM <sub>2.5</sub>	Campaign	
8.49	8.46		7.89	7.72		Tess 1	10.4.2015	5.23	0.67	0.13					Sav 1
9.03	9.54		10.45	8.37		Tess 1	11.4.2015	11.42	7.05	6.05					Sav 1
8.30	8.46		7.67	6.51		Tess 1	12.4.2015	6.39	4.79	4.57					Sav 1
6.74	6.84		5.3	5.37		Tess 1	13.4.2015	2.55	-0.09	-0.71					Sav 1
6.32	6.77		7.1	6.31		Tess 1	14.4.2015	2.95	1.17	-0.64	2.65	1.82			Sav 1
9.52	9.88		10.36	9.77		Tess 1	15.4.2015	1.76	0.41	-0.68	1.47	1.56			Sav 1
10.66	11.02		10.1	8.84		Tess 1	16.4.2015	2.37	0.92	0.71	3.99	2.48			Sav 1
12.64	13.10	11.14	10.62	12.34		Tess 1	17.4.2015	6.04	4.42	3.72	8.66	5.17			Sav 1
10.56	10.84	9.3	7.67	8.1		Tess 1	18.4.2015	7.37	5.49	4.20	10.38	8.11			Sav 1
15.58	15.93	10.75	10.41	11.95	11.9	Tess 1	19.4.2015	1.34	-0.29	-0.71	2.13	2.13			Sav 1
12.70		9.75	8.96	8.09	8.25	Tess 1	20.4.2015	12.53	2.25	2.30	7.81	3.43			Sav 1
14.05		11.38	10.21	13.5	10.84	Tess 1	21.4.2015	9.70	4.12	3.04	10.72	6.95			Sav 1
8.88	8.51	6.9	4.76	13.51	9.87	Tess 1	23.4.2015	2.81	-0.12	-0.68	3.73	3.79			Sav 1
9.93	9.95		7.76	6.14	-0.78	Tess 1	24.4.2015	7.29	2.55	1.55	7.90	4.48			Sav 1
15.07						Tess 1	25.4.2015	7.53	5.08	4.20	6.49	4.54			Sav 1
8.49	8.02	5.08	5.16	3.78	3.87	Tess 1	26.4.2015	2.59	1.09	0.00	4.04	3.89			Sav 1
7.22	6.65	5.16	4.71	3.5	4.06	Tess 1	27.4.2015	4.91	2.38	1.13	11.68	8.34			Sav 1
11.78	11.50	10.51	10.14	11.11	10.27	Tess 1	28.4.2015	10.03	3.00	2.16	2.46	1.13			Sav 1
10.47	11.12	9.33	7.6	10.82	10.61	Tess 1	29.4.2015	5.27	1.46	1.09	-0.17	-0.51			Sav 1
11.76	13.08	11.42	9.22			Tess 1	30.4.2015	6.26	3.34	2.12	11.29	7.49			Sav 1
15.26	15.94	12.75	12.8			Tess 1	1.5.2015	3.68	1.95	1.20	4.11	3.65			Sav 1
14.45	15.17	11.54	12.62			Tess 1	2.5.2015	1.39	-0.42	-1.34	-1.84	-1.34			Sav 1
9.63	9.74	7.2	7.12	1.41	4.73	Tess 1	3.5.2015	5.74	1.42	1.47	6.78	4.25			Sav 1
7.27	7.43	5.37	4.5	6.07	5.56	Tess 1	4.5.2015	7.12	3.50	2.58	8.28	5.23			Sav 1
10.28	10.59	7.97	6.76	9.51	8.83	Tess 1	5.5.2015	4.08	2.17	0.96	7.13	4.87			Sav 1
14.04		12.29	11.29	13.74	11.27	Tess 1	6.5.2015	9.44	7.57	7.85	4.11	5.71			Sav 1
6.95		5.58	4.59	5.03	3.99	Tess 1	7.5.2015	6.70	5.83	4.50	3.61	3.94			Sav 1
6.52	6.65	3.6	3.54	2.92	3.31	Tess 1	8.5.2015	4.58	3.74	2.62	1.70	3.20			Sav 1
8.18	8.16	6.21	4.92	5.75	5.97	Tess 1	9.5.2015	2.23	1.75	0.75	4.27	4.10			Sav 1
5.82	5.71	3.7	2.99	4.09	4.68	Tess 1	10.5.2015	3.91	2.25	1.37	4.63	3.58			Sav 1
3.41	3.43	2.12	1.17	-0.38	1.56	Tess 1	11.5.2015	3.80	2.38	1.50	1.67	2.24			Sav 1
4.02	4.03	2.59	1.46	4.72	4.08	Tess 1	12.5.2015	5.38	4.88	3.87	5.87	5.50			Sav 1
6.79	6.88	4.75	4.16	4.73	5.33	Tess 1	13.5.2015	2.20	1.37	0.37	0.59	1.07			Sav 1
4.92	4.93	3.13	2.42	5.95	2.18	Tess 1	14.5.2015	2.07	1.00	-0.87	2.92	0.78			Sav 1
6.19	6.35	4.54	4.08	2.48	3.66	Tess 1	15.5.2015	3.77	2.00	0.38	3.04	2.61			Sav 1
5.33	5.42	4.3	3.38	6.11	5.25	Tess 1	16.5.2015	2.90	1.75	0.62	3.58	3.14			Sav 1
5.71		4.99	3.66	1.53	2.95	Tess 1	17.5.2015	2.38	1.00	0.12	0.02	0.20			Sav 1
6.44	6.64	5.71	4.63	7.57	5.91	Tess 1	18.5.2015	4.66	4.14	3.13	8.34	5.52			Sav 1
5.38	5.49	4.12	3.92	3.74	3.22	Tess 1	19.5.2015	4.90	3.87	2.87	6.69	6.30			Sav 1
6.84	6.99	5.62	5.74	6.96	6.43	Tess 1	20.5.2015	6.54	5.75	4.25	9.53	8.67			Sav 1
4.17		3.75	2.54	4.77	4.84	Tess 1	21.5.2015	3.85	2.87	2.00	7.73	7.11			Sav 1
3.10	3.20	1.92	1.34	3.59	3.5	Tess 1	22.5.2015	5.71	2.87	1.00	11.92	9.84			Sav 2
4.03	4.16	3.29	2.08	-0.76	1.18	Tess 1	23.5.2015	2.70	0.50	-0.50	-1.76	-0.07			Sav 2
4.82	5.06	5.04	3.46	4.02	3.63	Tess 1	24.5.2015	3.54	1.13	0.75	4.85	3.41			Sav 2
4.39	4.62	3.79	3.33	3.46	3.73	Tess 1	25.5.2015	6.49	4.00	3.50	12.02	11.15			Sav 2
5.57	5.56	4.72	4.05	4.64	4.61	Tess 1	26.5.2015	8.14	5.25	4.50	6.56	6.34			Sav 2
7.28	7.57	6.33	5.79	5.4	5.28	Tess 1	27.5.2015	5.39	3.12	2.12	6.73	6.96			Sav 2
7.95	8.11	6.37	5.67	8.15	6.86	Tess 1	28.5.2015	5.40	2.75	1.75	5.43	6.30			Sav 2
6.65	7.15	5.48	4.58	9.01	6.35	Tess 1	29.5.2015	5.53	3.38	3.00	3.90	3.33			Sav 2
8.55	9.01	6.17	5.09	2.97	6.9	Tess 1	30.5.2015	6.05	2.88	1.88	7.13	5.86			Sav 2
7.59	7.84	6.25	4.83	4.16	5.14	Tess 1	31.5.2015	4.49	2.05	1.37	3.27	3.06			Sav 2
7.44	7.77			4.68	6.19	Tess 1	1.6.2015	7.63	3.62	3.26	9.89	7.91			Sav 2
7.94	8.41			2.92		Tess 1	2.6.2015	5.82	2.29	2.57	-6.82	-2.79			Sav 2
11.28	12.04	10.57	9	16.96	11.87	Tess 1	3.6.2015	2.78			1.81	1.94			Sav 2
5.68	5.73	3.48	3.21	7.64	6.38	Tess 1	4.6.2015	3.16	2.25	-0.37	4.73	2.22			Sav 2
9.23	9.35	7.67	6.88	8.6	8.46	Tess 1	5.6.2015	2.29			3.21	3.58			Sav 2
7.72	7.52	6.03	5.11	6.46	6.27	Tess 1	6.6.2015	4.71			2.75	5.92			Sav 2
5.23	5.25	3.29	2.96	4.71	4.77	Tess 1	7.6.2015	4.59			2.63	1.84			Sav 2
4.29	4.28	2.67	2.29	4.08	4.37	Tess 1	8.6.2015	3.59			2.50	5.39	3.89		Sav 2
5.20	5.08	4.33	3.75	3.16	3.89	Tess 1	9.6.2015	5.84			2.88	6.43	3.40		Sav 2
4.12	4.17	3.34	2.29	2.14	2.4	Tess 1	10.6.2015	5.57			3.75	4.95	3.55		Sav 2
6.04	6.26	5.84	5.34	5.9	5.28	Tess 1	11.6.2015	4.56			2.25	6.40	3.98		Sav 2
9.14	9.03	7.87	7.54	10.04	9.25	Tess 2	12.6.2015	7.25			2.97	6.70	4.33		Sav 2
7.34	7.76	8.53	7.41	5.63	5.93	Tess 2	13.6.2015	6.17			2.75	5.92	4.24		Sav 2
4.78	4.85	2.79	3.33	4.79	3.91	Tess 2	14.6.2015	4.30			2.38	2.59	1.13		Sav 2
10.60	10.98	9.59	9.3	9.62	9.03	Tess 2	16.6.2015	3.36	1.21	-0.37	6.07	4.29	4.29		Sav 2
11.03	11.39	10.62	9.7	10.93	10.97	Tess 2	18.6.2015	4.73	2.50	2.26	9.44	2.68			Sav 2
10.40	10.97	10.62	9.7	8.03	9.98	Tess 2	19.6.2015	11.64	11.05	10.83	13.94	13.09			Sav 2
14.57	15.80	13.83	14.4	16.37	14.85	Tess 2	21.6.2015	9.69	8.79	8.12	9.41	9.04			Sav 2
9.50	10.03	10.11	8.11	8.26	8.26	Tess 2	22.6.2015	7.74	6.64	5.75	9.13	6.35			Sav 2
2.11	2.30	1.92		2.31	2.85	Tess 2	23.6.2015	5.57	4.21	3.13	4.38	4.97			Sav 2
4.01	4.53	3.79	3.06	4.00	3.89	Tess 2	24.6.2015	4.71	3.91	2.74	13.36	11.74			Sav 2
4.87	5.64	5.22	3.63	5.56	5.07	Tess 2	25.6.2015	4.88	3.33	2.19	1.93	2.42			Sav 2
5.33	4.46	3.63	2.83	4.79	3.91	Tess 2	26.6.2015	5.79	4.04	3.12	7.32	6.50			Sav 2
3.14	3.44	2.46	1.79	1.28	1.11	Tess 2	28.6.2015	4.53	2.04	0.63	6.45	5.80			Sav 2
3.61	3.99	2.33	2.00	3.03	3.04	Tess 2	29.6.2015								

## **A28 Findings from the audit surveillance study**

The findings and the responses from the audit report (in Finnish) are reported here.

- The logbook at the site covers the description of the both measurement cabins and the installation of the instruments as well as their connection to the data acquisition system. Visits and activities have been recorded accordingly. However, some of the error messages have not been recorded.
  - The normal procedure is to remark the information from the instruments during the visits. The error messages from the instruments are normally collected by the data software.
- The inside temperature recording was not functional in one of the measurement cabin.
  - The cause was a lack of temperature sensor that could be connected to the data acquisition system. The manual temperature probe exist. The inside temperature was controlled by the air conditioning system.
- The sampling flow rate of the size selective instruments were checked regularly. The sampling flowrate of the MP 101 was lower than required at the beginning of the campaign. The missing flowrates of SHARP A was also observed.
  - The problem with the flow rate was the difficulty in adjusting the flowrate with MP 101 A based on the calibration of the flowrate. The technician from the representative company finally solved the problem by changing the sampling pump. The influence on the results (Tasa 1 and Tasa 2) is not significant.
- The temperature, pressure and humidity sensors of the CMs were checked between the same type of CMs to see if there is any differences. The

maximum deviation of pressure sensors was 3 kPa, which was slightly more than requirement (= 2 kPa).

- Calibration of the sensors of CMs were not calibrated within a year.
  - Calibration of the sensors were on the responsibility of the manufacturer/representative.
- Leak checks of the instruments were not performed
  - This was not considered necessary since the flow rates were measured at the sampling tube and the instrument itself recorded the flow rate continuously which was stored by the software. The leak test was performed by the manufacturer/representative.



ILMATIETEEN LAITOS  
METEOROLOGISKA INSTITUTET  
FINNISH METEOROLOGICAL INSTITUTE

FINNISH METEOROLOGICAL INSTITUTE

Erik Palménin aukio 1  
FIN-00560 Helsinki  
tel. +358 29 539 1000  
[WWW.FMI.FI](http://WWW.FMI.FI)

FINNISH METEOROLGRAPORTTEJA 2017:1  
ISBN 978-952-336-010-5 (pdf)  
ISSN 0782-6079  
Helsinki 2017

